

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 2.00PM)	Quotation Fee	Requesting Department	Focal Person
2	BME/A2K/(981)/2022/HTD (A12)	<p style="text-align: center;"><b>THE PROVISION OF MAINTENANCE SERVICES FOR MAQUET CARDIOHELP ECMO MACHINE IN MINISTRY OF HEALTH A PERIOD OF ONE (1) YEAR</b></p> <p style="text-align: center;"><u>PLACE OF SUBMISSION:</u> QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM</p>	07/01/2023	21/01/2023	\$5.00	JABATAN TEKNOLOGI PENJAGAAN KESIHATAN, KEMENTERIAN KESIHATAN.	HAJAH SYAZWANI AZAH BINTI HAJI ABU BAKAR JURUTERA PERUBATAN HAYAT KANAN TEL: 2381740 ext. 7561 (HTD)

# LAMPIRAN - A1

RUJ: BME/A2K/(981)/2023/HTD (A12)

5/1/2023

## THE PROVISION OF MAINTENANCE SERVICES FOR MAQUET CARDIOHELP ECMO MACHINE IN MINISTRY OF HEALTH FOR A PERIOD OF ONE (1) YEAR

### SCOPE OF WORK

Equipment details:

- i) Brand: Maquet
- ii) Model: Cardiohelp
- iii) BME24456, S/N: 90414581

Scope of work on ECMO system to follow Maquet manual/recommendation specific for model Cardiohelp which include the following where necessary:

- 1) Perform preventive maintenance on the equipment and its sub-systems twice (2x) a year.
- 2) Perform all necessary visual inspection.
- 3) Perform cleaning of interior and exterior of the system.
- 4) Perform any necessary calibration on the system.
- 5) Perform electrical safety test.
- 6) Perform test run on the equipment.
- 7) Apply preventive maintenance sticker/tag after successful service.
- 8) Disposal of any used spare parts/pm kits, unless otherwise instructed.

Others:

- 1) Work on the ECMO systems to be carried out by authorised engineer(s) authorised in providing maintenance to Maquet Cardiohelp system.
- 2) Vendor to submit maintenance checklist specific for Maquet Cardiohelp along with other necessary documents to join this tender.