

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 2.00PM)	Quotation Fee	Requesting Department	Focal Person
2	DSS/FS/PROC/B50K/2022-2023/14(PEV)	<p>"TO SUPPLY AND DELIVER FOUR (4) UNITS OF DEHUMIDIFIERS FOR PHYSICAL EVIDENCE SECTION, DEPARTMENT OF SCIENTIFIC SERVICES, MINISTRY OF HEALTH"</p> <p>PLACE OF SUBMISSION: QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM</p>	17/10/2022	05/11/2022	\$5.00	JABATAN PERKHIDMATAN SAINTIFIK, KEMENTERIAN KESIHATAN.	POLYYATI HAJI MUHD YUNUS PHYSICAL EVIDENCE SECTION TEL: 2381946

EQUIPMENT PURCHASE SPECIFICATION

TITLE OF TENDER/QUOTATION	TO SUPPLY, DELIVER, INSTALL AND COMMISSION FOUR (4) UNITS OF DEHUMIDIFIERS FOR PHYSICAL EVIDENCE SECTION, DEPARTMENT OF SCIENTIFIC SERVICES, MINISTRY OF HEALTH.
----------------------------------	--

TENDER REFERENCE NO.	
-----------------------------	--

NAME OF ITEM	FOUR (4) UNITS OF DEHUMIDIFIERS
---------------------	---------------------------------

NO	ITEM DESCRIPTIONS AND SPECIFICATIONS	VENDOR'S OFFER			
		COMPLY (Please tick ✓) (Provide evidence for compliance(s))		ITEM DESCRIPTIONS AND SPECIFICATIONS	PRICE (\$)
		YES	NO		
1	GENERAL				
1.1	Four (4) units of dehumidifiers suitable for regulating moisture within a certain area.				
2	Dehumidifiers with complete accessories				
2.1	Tank Capacity: at least 8 Litres				
2.2	Able to extract approximately at least 50 litres per day				
2.3	Running Temperature: allow 20°C ± 10°C				

Procurement Ref No : DSS/FS/PROC/B50K/2022-2023/14 (PEV)
 Section Ref No : DSS/PEV/PR/2022/007

2.4	Coolant: R-410A				
2.5	Coverage: at least up to 900 sq. ft.				
2.6	Fan speed (for air flow rate): at least 2 variable settings				
2.7	The unit should automatically stop during the following conditions: - Water tank is full Water tank is removed				
2.8	Weight when empty: no more than 20 kg				
2.9	Dehumidifiers must be equipped with:				
2.10	Auto humidity sensor				
2.10.1	Preferred humidity settings				
2.10.2	Full tank sensor/indicator				
2.10.3	Continuous air circulation				
2.10.4	Programmable timer				
2.10.5	Pre-filter screen				
2.10.6	Continuous drainage option				
2.10.7	Humidity display (%)				

2.10.8	Temperature display				
3	<u>ACCESSORIES</u>				
3.1	The tenderer shall provide all the necessary mandatory accessories needed to operate the dehumidifiers successfully.				
3.2	The accessories shall include hose for continuous drainage.				
4	<u>POWER SUPPLY</u>				
4.1	All electrical and electronic components in the system must be able to utilise mains power supply of 220 – 240 V AC, 50 – 60 Hz. Otherwise, vendor to include additional items or works required to ensure successful operation.				
5	<u>DOCUMENTATION</u>				
5.1	Operators manual for users and BME.				
6	<u>WARRANTY</u>				
6.1	A minimum of one (1) year warranty covering manufacturer's defects on hardware and software where all cost of repairs or replacements should be included.				
7	<u>PRICE VALIDITY</u> Price validity MUST NOT BE LESS THAN 90 days or three (3) months.				
					TOTAL PRICE (\$)

Procurement Ref No : DSS/FS/PROC/B50K/2022-2023/14 (PEV)
Section Ref No : DSS/PEV/PR/2022/007

No.	GENERAL SPECIFICATIONS	VENDOR'S OFFER
A	Total Price: B\$	
B	Delivery Period:	
C	Model & Brand:	
D	Country of Origin:	(leave blank)
E	Where marketed:	
F	Year of Manufacture:	
G	Warranty:	
H	Power Requirement:	
I	Battery Back-up:	
J	International Safety Standard:	
K	Technical Support:	
L	Equipment Whole Life Support:	
M	Dimensions (WxHxD) cm:	
N	User Manuals:	

Procurement Ref No : DSS/FS/PROC/B50K/2022-2023/14 (PEV)
Section Ref No : DSS/PEV/PR/2022/007

No.	GENERAL SPECIFICATIONS	VENDOR'S OFFER
O	Service Manuals:	
P	Spare-parts & Consumables Listing:	
Q	Technical Training On-Site:	
R	Site Requirements:	

*To all participating companies, please fill in the table above along with your other documents during submission of tender.

Bahagian/Unit	Physical Evidence	
Bil. Rujukan Bahagian/Unit:	DSS/PEV/PR/2022/007	
Pegawai di rujuk	Nama :	Polyyati Haji Muhd Yunus
	E-mail :	polyyati.yunus@moh.gov.bn
	Tel.No.:	2382424 ext 6057
	Fax No. :	2381946