

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 2.00PM)	Quotation Fee	Requesting Department	Focal Person
4	40/QTN/JPK/2022	<p>SUPPLY, DELIVERY INSTALLATION, TESTING AND COMMISSIONING, TWELVE (12) OF EAR THERMOSCAN FOR NEW BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, DEPARTMENT OF HEALTH SERVICES, MINISTRY OF HEALTH</p> <p><u>PLACE OF SUBMISSION:</u> QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM</p>	16/11/2022	03/12/2022	\$5.00	JABATAN PERKHIDMATAN KESIHATAN, KEMENTERIAN KESIHATAN.	HAJI AMIR HISMA BIN MASRI JURUTERA PERUBATAN HAYAT TEL: 2381640

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING,
TWELVE (12) UNITS OF EAR THERMOSCAN FOR NEW BLOCK SILVER JUBILEE
SENGKURONG HEALTH CENTRE, HEALTH SERVICES DEPARTMENT, MINISTRY OF HEALTH**

	SECTION I - USER REQUIREMENT
STANDARD FEATURES:	
	Ear thermoscan should be compact design, easy to use, gentle and fast. Takes readings in one second only and can helps to interpret the temperature with audio fever guidance.
	<ul style="list-style-type: none"> • Displayed temperature range. (please state range of temperature)
	<ul style="list-style-type: none"> • Uses coin type disposable battery
	<ul style="list-style-type: none"> • With LCD display and audio fever guidance
	<ul style="list-style-type: none"> • Suitable also for newborns.
	<ul style="list-style-type: none"> • Battery door design with safety lock to prevent choke hazard
	<ul style="list-style-type: none"> • Last reading recall to track temperature
	<ul style="list-style-type: none"> • Hygienic disposable lens filters.
	<ul style="list-style-type: none"> • Supply with additional disposable lens filter
	<ul style="list-style-type: none"> • Supplied with carrying casing

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING,
TWELVE (12) UNITS OF EAR THERMOSCAN FOR NEW BLOCK SILVER JUBILEE
SENGKURONG HEALTH CENTRE, HEALTH SERVICES DEPARTMENT, MINISTRY OF HEALTH**

	SECTION II - PROCUREMENT REQUIREMENT
MODEL AND BRAND:	<ul style="list-style-type: none"> • State Model and Brand of Equipment.
COUNTRY OF ORIGIN:	<ul style="list-style-type: none"> • State Country of Origin.
PRICE (CIF) B\$:	<ul style="list-style-type: none"> • State Unit and Total Price Offered.
WHERE MARKETED:	<ul style="list-style-type: none"> • Specify in which other countries the equipment marketed.
YEAR MANUFACTURED:	<ul style="list-style-type: none"> • State year of equipment manufactured.
WARRANTY:	<ul style="list-style-type: none"> • State number of years.
DELIVERY TIME:	<ul style="list-style-type: none"> • State number of weeks.
PRICE VALIDITY:	<ul style="list-style-type: none"> • State number of weeks.

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING,
 TWELVE (12) UNITS OF EAR THERMOSCAN FOR NEW BLOCK SILVER JUBILEE
 SENGKURONG HEALTH CENTRE, HEALTH SERVICES DEPARTMENT, MINISTRY OF HEALTH**

	SECTION III - TECHNICAL REQUIREMENT
MAINS POWER SUPPLY:	<ul style="list-style-type: none"> • Mains Supply 240VAC 50Hz, Battery Backup.
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:	<ul style="list-style-type: none"> • State equipment ambient operating temperature range.
INTERNATIONAL SAFETY STANDARD:	<ul style="list-style-type: none"> • The equipment must comply to either one of the five international safety standard, namely; US FDA Standard, European Union CE MARK, Australian TGA Standard, Canadian CSA Standard or Japanese Standard.
TECHNICAL SUPPORT:	<ul style="list-style-type: none"> • State number of locally stationed engineers/technicians and the nearest overseas support.
DIMENSIONS:	<ul style="list-style-type: none"> • State overall dimensions of equipment.
WEIGHT:	<ul style="list-style-type: none"> • State nett weight of the equipment.
EQUIPMENT WHOLE LIFE TIME SUPPORT:	<ul style="list-style-type: none"> • Number of years the spare parts are available after the installation of the equipment.
BROUCHER:	<ul style="list-style-type: none"> • Tenderers must submit with their offer, a detailed broucher which supports the information the tenderer has entered in the "form-to-be-used (FORM B)" for reference.
USER AND SERVICE MANUALS:	<ul style="list-style-type: none"> • Two hard copies and one soft copy [PDF FORMAT] to be submitted to BME office during commissioning.
SPARE-PARTS AND CONSUMABLES LISTINGS:	<ul style="list-style-type: none"> • Comprehensive spare-parts and consumables listings in the form of softcopy must be provided. Listing must have the following details; Part Number and Part Description.

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, TWELVE (12) UNITS OF EAR THERMOSCAN FOR NEW BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, HEALTH SERVICES DEPARTMENT, MINISTRY OF HEALTH

	SECTION I - USER REQUIREMENT	YES	NO	(IF 'NO') Please State Alternative Offer
STANDARD FEATURES:				
	Ear thermoscan should be compact design, easy to use, gentle and fast. Takes readings in one second only and can helps to interpret the temperature with audio fever guidance.			
	• Displayed temperature range. (please state range of temperature)			
	• Uses coin type disposable battery			
	• With LCD display and audio fever guidance			
	• Suitable also for newborns.			
	• Battery door design with safety lock to prevent choke hazard			
	• Last reading recall to track temperature			
	• Hygienic disposable lens filters.			
	• Supply with additional disposable lens filter			
	• Supplied with carrying casing			

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, TWELVE (12) UNITS OF EAR THERMOSCAN FOR NEW BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, HEALTH SERVICES DEPARTMENT, MINISTRY OF HEALTH

SECTION II – PROCUREMENT REQUIREMENT			
MODEL AND BRAND			
COUNTRY OF ORIGIN:			
PRICE (CIF) B\$	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 60%;">UNIT PRICE (B\$ CIF):</td> <td>TOTAL PRICE (B\$ CIF):</td> </tr> </table>	UNIT PRICE (B\$ CIF):	TOTAL PRICE (B\$ CIF):
UNIT PRICE (B\$ CIF):	TOTAL PRICE (B\$ CIF):		
WHERE MARKETED:			
YEAR MANUFACTURED:			
WARRANTY:			
DELIVERY TIME:			
PRICE VALIDITY:			

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, TWELVE (12) UNITS OF EAR THERMOSCAN FOR NEW BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, HEALTH SERVICES DEPARTMENT, MINISTRY OF HEALTH

SECTION III - TECHNICAL REQUIREMENTS	
MAINS POWER SUPPLY:	
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:	
BATTERY BACKUP:	
INTERNATIONAL SAFETY STANDARD:	
TECHNICAL SUPPORT:	
DIMENSIONS(mm):	
WEIGHT(in Kg):	
EQUIPMENT WHOLE LIFE TIME SUPPORT:	Number of years, spare parts are available after the installation of the equipment: _____ years.
	YES
	NO

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, TWELVE (12) UNITS OF EAR THERMOSCAN FOR NEW BLOCK SILVER JUBILEE SENGKURONG HEALTH CENTRE, HEALTH SERVICES DEPARTMENT, MINISTRY OF HEALTH

BROUCHER	Detailed broucher submitted?		
USER MANUALS	Tenderers must supply during commissioning, three sets of user manuals, one set must be in the form of soft copy. (PDF Format)		
SERVICE MANUALS	Tenderers must supply during commissioning, three sets of service manuals, one set must be in the form of soft copy. (PDF Format).		
SPARE-PARTS AND CONSUMABLES LISTINGS	<p>Tenderers must supply during commissioning, a comprehensive list of equipment spare-parts and consumables in the form of soft copy (e.g: Microsoft EXCEL).</p> <p>The listing must have the following details;</p> <ul style="list-style-type: none"> • Part Number • Part Description 		