

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 2.00PM)	Quotation Fee	Requesting Department	Focal Person
1	BME/A2K/(965)/2022/HTD(B4)	<p><b>TO SERVICE ONE (1) OF HAAG STREIT LASER BIOMETER (BME 18817) AT EYE CLINIC, RIPAS HOSPITAL, MINISTRY OF HEALTH</b></p> <p>PLACE OF SUBMISSION:            QUOTATION BOX (GROUND FLOOR)            MINISTRY OF HEALTH COMMONWEALTH DRIVE            BANDAR SERI BEGAWAN, BB 3910            NEGARA BRUNEI DARUSSALAM</p>	20/12/2022	07/01/2023	\$5.00	JABATAN TEKNOLOGI PENJAGAAN KESIHATAN, KEMENTERIAN KESIHATAN.	MUTABILLAH BIN HAJI ABDULLAH  PENOLONG PEGAWAI PENYELENGGARA PERKAKAS PERUBATAN  TEL: 2242424 ext: 8693 (RIPASH)

# LAMPIRAN - A1

RUJ: BME/A2K/(965)/2022/HTD (B4)

7/12/2022

TO SERVICE ONE(1) UNIT OF HAAG STREIT LASER BIOMETER (BME18817) AT EYE CLINIC, RIPAS HOSPITAL,  
MINISTRY OF HEALTH

## SCOPE OF WORK

### EQUIPMENT DETAILS:

EQUIPMENT: LENSTAR

MODEL: LS900

BRAND: HAAG STREIT

BME 18817

SN: 03879

### SCOPE OF WORK:

1. One off preventive maintenance (PM)
2. Physical inspection
3. Function and performance testing after PM
4. Perform cleaning of system
5. Include any necessary calibration
6. Conduct electrical safety testing
7. All test performed must be documented and available for user and BME

**LAMPIRAN - A2**

**RUJ: BME/A2K/(965)/2022/HTD (B4)**

**TO SERVICE ONE(1) UNIT OF HAAG STREIT LASER BIOMETER (BME18817) AT EYE CLINIC, RIPAS HOSPITAL, MINISTRY OF HEALTH**

NO	SPAREPART	PART NUMBER	QUANTITY
1	ONE OFF PM(TROUBLESHOOT, CALIBRATION, TESTING AND COMMISSIONING		1