

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 2.00PM)	Quotation Fee	Requesting Department	Focal Person
3	(08) PSD/QTN/2023 – (SSBH)	<p>“TO SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (02) UNITS OF PHARMACEUTICAL REFRIGERATOR, SURI SERI BEGAWAN (SSB) HOSPITAL, MINISTRY OF HEALTH (CLUSTERING)”</p> <p><u>PLACE OF SUBMISSION:</u> QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM</p>	02/02/2023	25/02/2023	\$5.00	HOSPITAL SURI SERI BEGAWAN, KUALA BELAIT.	<p>MOHAMAD ALIKHAN BIN HAJI MOHD YAKOP</p> <p>PMK KETUA PEGAWAI EKSEKUTIF TINGKAT 1</p> <p>TEL: 3335331 ext. 4043</p>

ITEM(S) SPECIFICATIONS FOR ADVERTISEMENT (ABOVE \$2000)

QUOTATION REFERENCE NO:	
QUOTATION NAME:	TO SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (02) UNITS OF PHARMACEUTICAL REFRIGERATOR, SURI SERI BEGAWAN (SSB) HOSPITAL, MINISTRY OF HEALTH (CLUSTERING).

USER'S REQUIREMENTS				VENDOR'S OFFER					
NO	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY OFFERED	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS (B\$)
1	PHARMACEUTICAL REFRIGERATOR	-	2 UNITS						
								TOTAL PRICE (B\$)	

NO	TERMS AND CONDITIONS	VENDOR'S OFFER (PLEASE STATE)
1	Tenderer must be registered with the Ministry of Health.	
2	QUOTATION/TENDER FORM should be filled completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form MAY cause DISQUALIFICATION OF QUOTATION/TENDER .	
4	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF QUOTATION/TENDER .	
5	All consumables supplied throughout this tender shall have a minimum expiry date of twelve (12) months / on delivery . Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made.	
6	Brochures / catalogues should be submitted / attached with quotation/tender document.	
7	Samples should be submitted together with quotation/ tender or within fourteen (14 days) of the quotation/tender closing date (if applicable).	
8	DELIVERY PERIOD: Not later than 8 weeks Staggered delivery upon request	(Yes / No) (If No, please specify)
9	PRICE VALIDITY: The quotation shall remain valid for 12 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).	

FOR QUOTATION ONLY

TERMS AND CONDITIONS

a.	Tenderer must be registered with the Ministry of Health.	<i>Acknowledgement:</i>	Company's Official Stamp
b.	Please fill in the QUOTATION FORM completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form MAY cause DISQUALIFICATION OF QUOTATION .	<i>Company Ref. No.:</i> I hereby certify the above quote to be correct. Signature:	
c.	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF QUOTATION .	Name:	
d.	Please do not use TIPPEX for amendment.	Designation: Date:	

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (2) UNITS OF
PHARMACEUTICAL REFRIGERATOR FOR SURI SERI BEGAWAN HOSPITAL**

SECTION I - USER REQUIREMENTS	
STANDARD FEATURES	<ul style="list-style-type: none"> • Must be suitable for the storage of vaccines and pharmaceuticals • Must come with an integrated temperature sensor with digital display
TECHNICAL SPECIFICATIONS	<ul style="list-style-type: none"> • Undercounter type • Single door • Lockable door • Door must close properly and have secure sealing • Control panel accessible via key (physical or password) • Adjustable temperature range must include 2 – 8 °C. • Microprocessor controlled temperature • Digital temperature display • Net capacity: between 200 – 300 litres • Drawer and shelf combination which are height adjustable • Have minimum and maximum temperature memory • Visual and audible alarm • Automatic defrosting • Melt water evaporation • All electrical and electronic components in the system must be able to utilize mains power supply of 220-240 V AC, 50-60 Hz. Otherwise, vendor to include additional items or work required to ensure successful operation
ACCESSORY: DIGITAL NONMERCURY	<ul style="list-style-type: none"> • Must include two (2) units of suitable digital nonmercury thermometer (one for each refrigerator) to verify internal temperature of refrigerator with traceability to NIST • The digital nonmercury thermometer must be calibrated and include a certificate informing the user of the level of accuracy compared to a recognized standard • Temperature range must include 2 – 8 °C. • The digital display should be able to be placed outside the storage unit

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (2) UNITS OF
PHARMACEUTICAL REFRIGERATOR FOR SURI SERI BEGAWAN HOSPITAL**

THERMOMETER	• The probe should be placed in buffered material
	• Has visual and audible alarms
	• Displays and records current, minimum and maximum temperatures
	• Accuracy: approximately +/- 0.5°C
	• Battery operated

SECTION II - PROCUREMENT REQUIREMENT	
MODEL AND BRAND	• State Model and Brand of Equipment.
COUNTRY OF ORIGIN	• State Country of Origin.
PRICE (CIF) B\$	• State Unit and Total Price Offered.
WHERE MARKETED	• Specify in which other countries the equipment marketed.
YEAR MANUFACTURED	• State year of equipment manufactured.
WARRANTY	• State number of years.
DELIVERY TIME	• State number of weeks.
PRICE VALIDITY	• State number of weeks.

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (2) UNITS OF
PHARMACEUTICAL REFRIGERATOR FOR SURI SERI BEGAWAN HOSPITAL**

SECTION III - TECHNICAL REQUIREMENT	
MAINS POWER SUPPLY	<ul style="list-style-type: none"> Mains Supply 220-240VAC 50-60Hz
INTERNATIONAL SAFETY STANDARD	<ul style="list-style-type: none"> The equipment must comply to either one of the five international safety standards, namely; US Standard (FDA), European Union CE MARK, Australian Standard TGA), Canadian Standard (CSA) or Japanese Standard (JIS).
TECHNICAL SUPPORT	<ul style="list-style-type: none"> State number of locally stationed engineers/technicians and the nearest overseas support.
DIMENSIONS	<ul style="list-style-type: none"> State overall dimensions of equipment.
WEIGHT	<ul style="list-style-type: none"> State net weight of the equipment.
EQUIPMENT WHOLE LIFE TIME SUPPORT	<ul style="list-style-type: none"> Number of years the spare parts are available after the installation of the equipment.
SITE REQUIREMENT	<ul style="list-style-type: none"> N/A
BROCHURE	<ul style="list-style-type: none"> Tenderers must submit with their offer, a detailed brochure which supports the information the tenderer has entered in the "form-to-be-used (FORM B)" for reference.
USER AND SERVICE MANUALS:	<ul style="list-style-type: none"> Two hard copies and one soft copy [PDF FORMAT] to be submitted to BME office during commissioning.
SPARE PARTS AND CONSUMABLES LISTINGS:	<ul style="list-style-type: none"> Comprehensive spare-parts and consumables listings in the form of softcopy must be provided. Listing must have the following details; Part Number and Part Description.

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (2) UNITS OF
PHARMACEUTICAL REFRIGERATOR FOR SURI SERI BEGAWAN HOSPITAL**

WARRANTY PLANNED PREVENTIVE MAINTENANCE:	<ul style="list-style-type: none"> • Minimum of two times PPM per year during the equipment warranty, six months after the date of commissioning and at the end of warranty period.
TRAINING: ON-SITE:	<ul style="list-style-type: none"> • Tenderers are required to conduct technical training to Biomedical Engineers and Technicians. • Training to be conducted locally, tenderers are required to: <ul style="list-style-type: none"> • Provide training materials, test equipment, demo equipments, etc. • Provide training to two groups of technical staffs. • Provide 2 days (minimum) of training for each group. • Training to be conducted on the premises of the installed equipment or other locations as determined by the user
MANUFACTURER'S PREMISES (IF REQUIRED):	<ul style="list-style-type: none"> • Training to be conducted at the manufacturer's training premises/ facilities. Tenderers are required to; <ul style="list-style-type: none"> • Provide air tickets, accommodation, transportation, etc. for 2 persons. • Provide training for 2 persons.