

Rujukan Kami: (81) MOH/HQ/P/IKLAN-SH/2024

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 02.00PM)	Quotation Fee	Requesting Department	Focal Person
2	(23) PPN/QTN/2024 (BME)	<p style="text-align: center;">SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (2) UNITS OF HYDRAULIC BENCH PRESS FOR DENTAL LABORATORY, DEPARTEMENT OF DENTAL SERVICES, MINISTRY OF HEALTH</p> <p style="text-align: center;"><u>PLACE OF SUBMISSION:</u> QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM</p>	24/06/2024	06/07/2024	\$5.00	JABATAN PERKHIDMATAN PERGIGIAN.	<p style="text-align: center;">NURAMALIAH BINTI HAJI JAMALUDIN</p> <p style="text-align: center;">JURUTERA PERUBATAN HAYAT</p> <p style="text-align: center;">TEL: 2380433</p>

**UNIT SEBUTHARGA
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (2) UNITS OF
HYDRAULIC BENCH PRESS FOR DENTAL LABORATORY,
DEPARTMENT OF DENTAL SERVICES**

NO.	TERMS AND CONDITIONS	VENDOR'S OFFER (PLEASE STATE)
1	Tenderer must be registered with the Ministry of Health.	
2	TENDER FORM should be filled completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form MAY cause DISQUALIFICATION OF TENDER.	
3	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF TENDER.	
4	All consumables supplied throughout this tender <u>shall</u> have a minimum expiry date of twelve (12) months / on delivery. Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made.	
5	Brochures / catalogues should be submitted / attached with tender document.	
6	Any room renovation which may be required, it is mandatory to conduct site visit (if applicable)	
7	Samples should be submitted together with tender or within fourteen (14 days) of the tender closing date (if applicable).	
8	DELIVERY PERIOD: (Please state) Not More Than 90 days upon confirmation	(Yes / No) (If No, please specify)
9	PRICE VALIDITY: The quotation shall remain valid for 12 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).	

**UNIT SEBUTHARGA
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (2) UNITS OF
HYDRAULIC BENCH PRESS FOR DENTAL LABORATORY,
DEPARTMENT OF DENTAL SERVICES**

Location for installation	1. National Dental Centre 2. PMMPMHAMB Hospital
----------------------------------	--

SECTION I - USER REQUIREMENT		YES	NO	PLEASE SPECIFY THE SPECIFICATION/ SCOPE OFFERED <small>FAILURE TO FILL THE OFFER DETAIL SHALL BE IDENTIFIED AS NON-COMPLIANCE TO THE OFFER</small>
Hydraulic Bench Press				
Standard features	For dental laboratory purpose.			
	Suitable for pressing 1-3 or 2-4 flasks			
	Piston made of break-proof and rust-proof material			
	Adjustable working pressure via pump lever			
	High stability			
	Equipped with valve which can be opened to allow the piston to return to its starting position automatically			
	Pressure gauge available for precision display recessed in housing			
	Hydraulic oil level display integrated to the unit			
Distance between columns	Approximately 152mm (in order to allow use of larger flasks and flask clamps)			
Working pressure	Approximately up to 6 tonnes			
Accessories	For pressing 1-3 flasks			

**UNIT SEBUTHARGA
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (2) UNITS OF
HYDRAULIC BENCH PRESS FOR DENTAL LABORATORY,
DEPARTMENT OF DENTAL SERVICES**

SECTION II - PROCUREMENT REQUIREMENT

Model and Brand of Equipment.	
Country of Origin.	
Unit (B\$)	
Total Price Offered (CIF)	
Country of Origin	
Year manufactured.	
Warranty period (at least 1 year)	_____ years
Delivery period	_____ weeks
Price validity (not less than 12 months)	_____ years



**UNIT SEBUTHARGA
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (2) UNITS OF
HYDRAULIC BENCH PRESS FOR DENTAL LABORATORY,
DEPARTMENT OF DENTAL SERVICES**

TECHNICAL SPECIFICATIONS/REQUIREMENTS	
Mains Supply 240VAC 50 Hz.	
State equipment ambient operating temperature range.	
The equipment must comply to either one of the five international safety standards, namely; US FDA Standard, European Union CE MARK, Australian TGA Standard, Canadian CSA Standard or Japanese JIS Standard.	
State number of locally stationed engineers/technicians and the nearest overseas support.	
State overall dimensions of equipment.	
State net weight of the equipment (kg)	
Equipment whole life time support Supplier must guarantee spare parts are available at least for 10 years after the installation of the equipment.	_____years
Brochure Tenderers must submit with their offer, a detailed brochure which supports the information the tenderer has entered in the "form-to-be-used (FORM B)" for reference.	
User and Service Manuals Two hard copies and one soft copy [PDF FORMAT] to be submitted to BME office during commissioning.	
Spare-parts and consumables Comprehensive spare-parts and consumables listings in the form of softcopy must be provided. Listing must have the following details; Part Number and Part Description.	
Warranty planned preventive maintenance. Minimum of two times PPM per year during the equipment warranty, six months after the date of commissioning and at the end of warranty period.	
Training Tenderers are required to conduct technical	



**UNIT SEBUTHARGA
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TWO (2) UNITS OF
HYDRAULIC BENCH PRESS FOR DENTAL LABORATORY,
DEPARTMENT OF DENTAL SERVICES**

training to Biomedical Engineers and Technicians.

- Training to be conducted at the manufacturer's training premises / facilities. Tenderers are required to;
- Provide air tickets, accommodation, transportation, etc. for 2 persons.
- Provide training for 2 persons.
- Minimum no of days for training – 3 days.

OR

- Training to be conducted locally, tenderers are required to:
- Provide training materials, test equipment, demo equipments, etc.
- Provide training to two groups of technical staffs.
- Provide 2 days (minimum) of training for each group.
- Training to be conducted at BME office/ Hospital.

