

Rujukan Kami: (81) MOH/HQ/P/IKLAN-SH/2024

BIL.	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 02.00PM)	Quotation Fee	Requesting Department	Focal Person
5	(26) PPN/QTN/2024 (BME)	<p><b>SUPPLY AND DELIVERY OF DENTAL SLOW SPEED HIGHSPEED HANDPIECE FOR DEPARTMENT OF DENTAL SERVICES, MINISTRY OF HEALTH</b></p> <p><u>PLACE OF SUBMISSION:</u>            QUOTATION BOX (GROUND FLOOR)            MINISTRY OF HEALTH            COMMONWEALTH DRIVE            BANDAR SERI BEGAWAN, BB 3910            NEGARA BRUNEI DARUSSALAM</p>	24/06/2024	06/07/2024	\$5.00	JABATAN PERKHIDMATAN PERGIGIAN.	NURAMALIAH BINTI HAJI JAMALUDIN JURUTERA PERUBATAN HAYAT TEL: 2380433

**UNIT SEBUTHARGA  
KEMENTERIAN KESIHATAN  
NEGARA BRUNEI DARUSSALAM**

**SUPPLY AND DELIVERY DENTAL SLOW SPEED HANDPIECE, DEPARTMENT OF  
DENTAL SERVICES**

NO.	TERMS AND CONDITIONS	VENDOR'S OFFER (PLEASE STATE)
1	Tenderer must be registered with the Ministry of Health.	
2	<b>TENDER FORM should be filled</b> completely including the <b>USER REQUIREMENT FORM</b> (if available). Submission of incomplete form <b>MAY</b> cause <b>DISQUALIFICATION OF TENDER.</b>	
3	Each tenderer is allowed to quote <b>ONE BRAND WITH ONE PRICE ONLY</b> for each item. Submission of more than one brand and price will cause <b>DISQUALIFICATION OF TENDER.</b>	
4	All consumables supplied throughout this tender <u>shall</u> have a minimum expiry date of <b>twelve (12) months / on delivery.</b> Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made.	
5	<b>Brochures / catalogues should be submitted / attached</b> with tender document.	
6	Any <b>room renovation</b> which may be required, <b>it is mandatory to conduct site visit</b> (if applicable)	
7	<b>Samples should be submitted together with tender or within fourteen (14 days)</b> of the tender closing date (if applicable).	
8	<b>DELIVERY PERIOD:</b> (Please state) Not More Than <b>90 days</b> upon confirmation	<b>(Yes / No)</b> <b>(If No, please specify)</b>
9	<b>PRICE VALIDITY:</b> The quotation shall remain valid for <b>12 MONTHS</b> from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).	

**UNIT SEBUTHARGA  
KEMENTERIAN KESIHATAN  
NEGARA BRUNEI DARUSSALAM**

**SUPPLY AND DELIVERY DENTAL SLOW SPEED HANDPIECE, DEPARTMENT OF  
DENTAL SERVICES**

SECTION I - USER REQUIREMENT		YES	NO	PLEASE SPECIFY THE SPECIFICATION/ SCOPE OFFERED <small>FAILURE TO FILL THE OFFER DETAIL SHALL BE IDENTIFIED AS NON-COMPLIANCE TO THE OFFER</small>
<b>Item 1: Slow speed handpiece (Contra angle)</b>				
<b>Standard features</b>	Quantity: Sixty five (65) units			
	Dental Contra-angle handpiece			
	For contra angle burs			
	Transmission 1:1			
	Push-button chuck			
	Cellular optic			
	Anti-Heat system			
	Equipped with mechanism that prevents the aerosol particles to enter the turbine head from the oral cavity.			
<b>Head size (Diameter in mm) :</b>	Approximately 8.9 mm			
<b>Maximum speed:</b>	Approximately 40,000 rpm			
<b>Water spray:</b>	1-port			
<b>Connection:</b>	Universal "E" type connection			
<b>Compatibility:</b>	Compatible with the current system of dental chair			

**UNIT SEBUTHARGA  
KEMENTERIAN KESIHATAN  
NEGARA BRUNEI DARUSSALAM**

**SUPPLY AND DELIVERY DENTAL SLOW SPEED HANDPIECE, DEPARTMENT OF  
DENTAL SERVICES**

<b>Item 2: Slow speed handpiece (Straight)</b>				
<b>Standard features</b>	Quantity: Fifteen (15) units			
	Dental straight handpiece			
	For straight handpiece burs			
	Transmission 1:1			
	Anti-Heat system			
	Equipped with mechanism that prevents the aerosol particles to enter the turbine head from the oral cavity.			
<b>Maximum speed:</b>	Approximately 40,000 rpm			
<b>Water spray:</b>	1-port			
<b>Connection:</b>	Universal "E" type connection			
<b>Compatibility:</b>	Compatible with the current system of dental chair			

**UNIT SEBUTHARGA  
KEMENTERIAN KESIHATAN  
NEGARA BRUNEI DARUSSALAM**

**SUPPLY AND DELIVERY DENTAL SLOW SPEED HANDPIECE, DEPARTMENT OF  
DENTAL SERVICES**

SECTION II - PROCUREMENT REQUIREMENT	
<b>Model and Brand of Equipment.</b>	
<b>Country of Origin.</b>	
<b>Unit (B\$)</b>	
<b>Total Price Offered (CIF)</b>	
<b>Country of Origin</b>	
<b>Year manufactured.</b>	
<b>Warranty period (at least 1 year)</b>	_____ years
<b>Delivery period</b>	_____ weeks
<b>Price validity (not less than 12 months)</b>	_____ years



**UNIT SEBUTHARGA  
KEMENTERIAN KESIHATAN  
NEGARA BRUNEI DARUSSALAM**

**SUPPLY AND DELIVERY DENTAL SLOW SPEED HANDPIECE, DEPARTMENT OF  
DENTAL SERVICES**

TECHNICAL SPECIFICATIONS/REQUIREMENTS	
Mains Supply 240VAC 50 Hz.	
State equipment ambient operating temperature range.	
The equipment must comply to either one of the five international safety standards, namely; US FDA Standard, European Union CE MARK, Australian TGA Standard, Canadian CSA Standard or Japanese JIS Standard.	
State number of locally stationed engineers/technicians and the nearest overseas support.	
State overall dimensions of equipment.	
State net weight of the equipment (kg)	
<b>Equipment whole life time support</b> Supplier must guarantee spare parts are available at least for <b>10 years</b> after the installation of the equipment.	_____years
<b>Brochure</b> Tenderers must submit with their offer, a detailed brochure which supports the information the tenderer has entered in the "form-to-be-used (FORM B)" for reference.	
<b>User and Service Manuals</b> Two hard copies and one soft copy [PDF FORMAT] to be submitted to BME office during commissioning.	
<b>Spare-parts and consumables</b> Comprehensive spare-parts and consumables listings in the form of softcopy must be provided. Listing must have the following details; Part Number and Part Description.	
<b>Warranty planned preventive maintenance.</b> Minimum of two times PPM per year during the equipment warranty, six months after the date of commissioning and at the end of warranty period.	
<b>Training</b> Tenderers are required to conduct technical training to Biomedical Engineers and Technicians.  ● Training to be conducted at the manufacturer's training premises / facilities. Tenderers are required to;	

**UNIT SEBUTHARGA  
KEMENTERIAN KESIHATAN  
NEGARA BRUNEI DARUSSALAM**

**SUPPLY AND DELIVERY DENTAL SLOW SPEED HANDPIECE, DEPARTMENT OF  
DENTAL SERVICES**

<ul style="list-style-type: none"><li>● Provide air tickets, accommodation, transportation, etc. for 2 persons.</li><li>● Provide training for 2 persons.</li><li>● Minimum no of days for training – 3 days.</li></ul> <p><b>OR</b></p> <ul style="list-style-type: none"><li>● Training to be conducted locally, tenderers are required to:</li><li>● Provide training materials, test equipment, demo equipments, etc.</li><li>● Provide training to two groups of technical staffs.</li><li>● Provide 2 days (minimum) of training for each group.</li><li>● Training to be conducted at BME office/ Hospital.</li></ul>	
--	--