## Rujukan Kami: (82) MOH/HQ/P/IKLAN-SH/2024

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 02.00PM)	Quotation Fee	Requesting Department	Focal Person
2	19/QTN/JPK/2024	SUPPLY AND DELIVERY OF TWO HUNDRED (200) PIECES OF ADULT PEAK FLOW METER FOR ALL HEALTH CENTRE, DEPARTMENT OF HEALTH  PLACE OF SUBMISSION: QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM	25/06/2024	13/07/2024	\$5.00	JABATAN PERKHIDMATAN KESIHATAN, KEMENTERIAN KESIHATAN.	HAJAH NORHALIZAWATY BINTI HJ ABD RAZAK PEGAWAT JURURAWAT TEL: 2382652

### TAWARAN SEBUTHARGA JABATAN PERKHIDMATAN KESIHATAN

Rujukan Tawaran: ,	
--------------------	--

Nama Tawaran:

# SUPPLY AND DELIVERY OF TWO HUNDRED (200) PIECES OF ADULT PEAK FLOW METER FOR ALL HEALTH CENTRES, DEPARTMENT OF HEALTH SERVICES

	SPECIFICATIONS		
	200 PCS ADULT PEAK FLOW METER		
	TO PROVIDE ACCURATE PEAK EXPIRATORY FLOW (PEF) MEASUREMENTS AS PART OF ASTHMA MANAGEMENT PLAN		
	AVAILABLE WITH ATS AND EU SCALE		
	MEASUREMENT RANGE: 60-880 L/MIN (EU SCALE), 60 -880 L/MIN (ATS SCALE)		
	ACCURACY: ±10% OR 10 L / MIN (EU SCALE), ±10% OR 20 L/ MIN (ATS SCALE)		
	EACH DEVICE IS INDIVIDUALLY CALIBRATED FOR ABSOLUTE ACCURANCY		
STANDARD FEATURES/ USER REQUIREMENTS	EASY-TO-READ SCALE (TEXT ON A DIFFERENT COLOUBACKGROUND) DISPLAY IN SCALE INCREMENTS OF 10 L/MIN		
USER REQUIREMENTS	BUILT-IN ONE-WAY VALVE PREVENTS INHALING THROUGH THE DEVICE		
	COMPLETE WITH DETACHABLE, REUSABLE MOUTHPIECE AND ALSO ALLOWS FOR MULTI-PATIENT USE WITH DISPOSABLE MOUTHPIECES		
	PORTABLE		
	LIGHTWEIGHT - ±10% 76G WITH MOUTHPIECE		
	DIMENSIONS NOMINAL: 157 x 44MM W/O MOUTHPIECE, 210 X 44MM WITH MOUTHPIECE		
	EASY TO CLEAN		
	MATERIALS: ALL LATEX FREE		

SPECIFICATIONS FORM A

#### PROCUREMENT REQUIREMENT

- Brand/ Model
- Accessories
- Country of origin
- International Safety Standard
- Year Manufactured
- Shelf life
- Price per unit
- Total price
- Delivery period
- Price Validity
- Sample & Brochure
- Warranty

Authorized by:

HAJAH NORHALIZAWATY BINTI HJ ABD RAZAK PEGAWAI JURURAWAT FORM TO BE USED FORM B

## TAWARAN SEBUTHARGA JABATAN PERKHIDMATAN KESIHATAN

Rujukan Tawaran:				
Nama Tawaran:				

## SUPPLY AND DELIVERY OF TWO HUNDRED (200) PIECES OF ADULT PEAK FLOW METER FOR ALL HEALTH CENTRES DEPARTMENT OF HEALTH SERVICES

	ITEMS	YES	NO	IF 'NO' PLEASE STATE ALTERNATIVE OFFER
	200 Pcs Adult Peak Flow Meter			
	To Provide Accurate Peak Expiratory Flow (PEF) Measurements as Part of Asthma Management Plan Available With ATS and EU Scale Measurement Range: 60-880 L/Min (Eu Scale), 60-880 L/Min			
	(Ats Scale) Accuracy: ±10% Or 10 L / Min (Eu Scale), ±10% Or 20 L/ Min (ATS Scale)			
SPECIFICATIONS/ USER	Each Device Is Individually Calibrated for Absolute Accuracy			
REQUIRMENTS	Easy-To-Read Scale (Text on A Different Colour Background) Display in Scale Increments Of 10 L/Min			
	Built-In One-Way Valve Prevents Inhaling Through the Device			
	Complete With Detachable, Reusable Mouthpiece and Also Allows for Multi-Patient Use with Disposable Mouthpieces			
	Portable			
	Lightweight - ±10% 76g With Mouthpiece			
	Dimensions Nominal: 157 X 44mm W/O Mouthpiece, 210 X 44mm With Mouthpiece			
	Easy To Clean			
	Materials: All Latex Free			

FORM TO BE USED FORM B

PROCUREMENT REQUIREMENT			
Brand/ Model			
Country of origin			
International Safety Standard			
Year Manufactured			
Shelf life (minimum 1.5 years)			
Price per unit			
Total price			
Delivery period			
Price Validity			
Sample & Brochure			
Warranty			