

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 02.00PM)	Quotation Fee	Requesting Department	Focal Person
2	259/CPC/2025/IKLAN/MOH	<p>SUPPLY, DELIVERY AND COMMISSIONING OF SINGLE DOOR REFRIGERATOR FOR WELLNESS CENTRE MINISTRY OF HEALTH</p> <p>PLACE OF SUBMISSION: QUOTATION BOX (GROUND FLOOR) MINISTRY OF HEALTH COMMONWEALTH DRIVE BANDAR SERI BEGAWAN, BB 3910 NEGARA BRUNEI DARUSSALAM</p>	30/01/2025	15/02/2025	\$5.00	JABATAN TEKNOLOGI PENJAGAAN KESIHATAN, KEMENTERIAN KESIHATAN.	<p>NUR ATIQAH BINTI HAJI ABD LATIP</p> <p>JURUTERA PERUBATAN HAYAT KANAN</p> <p>TEL: 2381740 ext. 7561 (HTD)</p>

**KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING, AND COMMISSIONING OF SINGLE DOOR REFRIGERATOR FOR WELLNESS CENTRE
MINISTRY OF HEALTH**

NO.	SECTION 1: TERMS AND CONDITIONS	VENDOR'S OFFER (PLEASE STATE)
1	Tenderer must be registered with the Ministry of Health.	
2	TENDER FORM should be filled completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form MAY cause DISQUALIFICATION OF TENDER .	
3	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF TENDER .	
4	All consumables supplied throughout this tender <u>shall</u> have a minimum expiry date of twelve (12) months / on delivery . Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made.	
5	Brochures / catalogues should be submitted / attached with tender documents.	
6	Any room renovation which may be required, it is mandatory to conduct site visit (if applicable)	
7	Samples should be submitted together with tender or within fourteen (14 days) of the tender closing dates (if applicable).	
8	DELIVERY PERIOD:(Please state) Not More Than 90-120 days upon confirmation	
9	PRICE VALIDITY: The quotation shall remain valid for at least 12 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).	

**KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING, AND COMMISSIONING OF SINGLE DOOR REFRIGERATOR FOR WELLNESS CENTRE
MINISTRY OF HEALTH**

	One (1) unit Single door refrigerator	Yes	No	Remarks
		(☑ Tick where appropriate)		
1.	<p>Standards Compliance: The refrigerator must conform to one or more of the following standards (tick where appropriate):</p> <ul style="list-style-type: none"> <input type="checkbox"/> DIN 13277/58345 <input type="checkbox"/> European Standard EN 60068-3 for environmental testing and equipment durability. <input type="checkbox"/> US FDA 21 CFR Part 820 (Quality System Regulations) for medical device manufacturing and storage. <input type="checkbox"/> USP <1079> Good Storage and Shipping Practices for temperature-sensitive pharmaceutical storage. <input type="checkbox"/> Others: Please specify 			
2	General features			
2.1	Single glass door			
2.2	Insulating lockable-dual panel glass door			
2.3	Capacity: 500-600 Litre			
2.4	Adjustable temperature range: 2–8°C			
2.5	CFC free refrigerant			
2.6	Microprocessor controlled temperature.			
2.7	Digital temperature LED display.			
2.8	Automatic defrost			
3	Storage system			
3.1	The refrigerator must include at least 6 adjustable drawers or shelves .			
3.2	Retaining mechanism: Each drawer or shelf must be equipped with a secure mechanism to ensure items stay in place during use			
3.3	Dividers: Each drawer/shelf must include adjustable dividers or a comparable system to allow flexible and customizable compartmentalization of pharmaceutical products. Dividers must be made of durable, non-reactive, and easy-to-clean material.			
4	Equipped with 4 castors with locks .			

**KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING, AND COMMISSIONING OF SINGLE DOOR REFRIGERATOR FOR WELLNESS CENTRE
MINISTRY OF HEALTH**

	One (1) unit Single door refrigerator	Yes	No	Remarks
		(☑ Tick where appropriate)		
5	Interior fans for forced-air circulation to ensure uniform temperature distribution.			
6	Alarm; <ul style="list-style-type: none"> • Power failure alarm (visual and audible). • Glass door ajar alarm 			
7	Low-noise compressor.			
8	Interior lighting for clear visibility			
9	Equipped with a porthole to allow insertion of temperature sensor			
10	ACCESSORIES			
10.1	One (1) Uninterrupted Power Supply (UPS) with a minimum capacity of 2100 Watts and backup time of at least 1 hour.			
10.2	One (1) non-mercury verification thermometer , with traceability to NIST , temperature range: -5°C to 15°C.			
10.3	The thermometer must include calibration certification and traceability to ensure compliance.			
11	ELECTRICAL REQUIREMENTS			
11.1	All electrical and electronic components in the system must be able to utilise mains power supply of 220 – 240 V AC, 50 – 60 Hz. Otherwise, vendors include additional items or works required to ensure successful operation.			
12	If additional components are required to ensure compatibility, the vendor must include these in the proposal.			

2	END-USER TRAINING			Remarks
	Yes	No	Remarks	
2.1	Conduct user training to the all-end users by an application specialist or competent local engineer including but not limited to: <ul style="list-style-type: none"> • Basic user operation, user troubleshooting and user maintenance • CPACS and/or RPACS guide through (if applicable) • Provide Operating manual (Hardcopy and/or Softcopy) 			

**KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING, AND COMMISSIONING OF SINGLE DOOR REFRIGERATOR FOR WELLNESS CENTRE
MINISTRY OF HEALTH**

3	TECHNICAL TRAINING			
3.1	<p>Introductory Technical Training to Biomedical Engineers and Technicians at BME RIPASH Office by competent Tenderer's Engineer/Technicians that includes but not limited to:</p> <ul style="list-style-type: none"> • Troubleshooting and basic corrective maintenance • Handling and basic inspection maintenance <p>*(Two sessions/groups if required)</p>			
4	WARRANTY			
		Yes	No	Remarks
4.1	<p>Tenderer to include warranty period of at least two (2) years</p> <p>Tenderers to INCLUDE a Warranty Undertaking Letter stating the terms of warranty provided for the equipment in the tender. This includes but not limited to:</p> <ul style="list-style-type: none"> • Duration and warranty coverage • Excluded from Warranty • 2x Warranty Planned Preventive Maintenance (one of which includes, if applicable, PM kit) on the sixth month after date of commissioning and end of warranty period. • Scope of Planned Preventive Maintenance 			

**KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING, AND COMMISSIONING OF SINGLE
DOOR REFRIGERATOR FOR WELLNESS CENTRE MINISTRY OF HEALTH**

PRICE PROPOSAL	
PRICE PER UNIT:	
TOTAL PRICE:	
DELIVERY TIME:	(Please state) Not More Than 90-120 days upon confirmation
PRICE VALIDITY:	[AT LEAST ONE (1) YEAR PRICE VALIDTY]

PROCUMENT AND TECHNICAL SPECIFICATION				
BRAND:				
MODEL:				
COUNTRY OF ORIGIN:				
WHERE MARKETED:				
YEAR INTRODUCED TO MARKET:				
DETAILED BROCHURE INCLUDED?	YES		NO	<input type="checkbox"/> Tick where appropriate
USER AND SERVICE MANUALS:	YES		NO	Tenderers to acknowledge that they must provide at least TWO sets of USER AND SERVICE manuals when applying commissioning form. One Set for End User, One Set for BME. (Please provide hardcopy or softcopy)
MAINS POWER SUPPLY:				
POWER RATINGS:				
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:				
BATTERY BACKUP:				
INTERNATIONAL SAFETY STANDARD:	<input type="checkbox"/> Tick where appropriate Must comply at least one of the following; <ul style="list-style-type: none"> <input type="checkbox"/> US FDA Standard, <input type="checkbox"/> European Union CE MARK, <input type="checkbox"/> Australian TGA Standard, <input type="checkbox"/> Canadian CSA Standard or 			

**KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING, AND COMMISSIONING OF SINGLE
DOOR REFRIGERATOR FOR WELLNESS CENTRE MINISTRY OF HEALTH**

	<input type="checkbox"/> Japanese JIS Standard. <input type="checkbox"/> Others (Please specify): _____	
TECHNICAL SUPPORT:	NO OF LOCAL ENGINEER/TECHNICIAN: _____ NEAREST OVERSEA SUPPORT: _____	
DIMENSIONS OF MAIN UNIT:		MM / CM / INCH
WEIGHT OF MAIN UNIT:		KG / G / LBS
EQUIPMENT WHOLE LIFE TIME SUPPORT:	Number of years, spare parts are available after the installation of the equipment: _____ years	