REFERENCE OF TENDER	DESCRIPTION OF TENDER	TIME PERIOD OF TENDER	DEPARTMENT/ DIVISION/UNIT REQUESTING TENDER	FEES	CLOSING DATE NOT LATER THAN 2.00PM	FOCAL PERSON
KK/13/2025/UPP(TC)	SUPPLY AND DELIVERY MEDICAL ITEMS (CARDIAC IMPLANTABLE ELECTRONIC DEVICE (CIED) FOR DEPARTMENT OF CARDIOLOGY RAJA ISTER PENGIRAN ANAK SALEHA (RIPAS (HOSPITAL) FOR A PERIOD OF FIVE (5) YEARS	5 YEARS	RIPAS HOSPITAL	\$500.00	18 <sup>TH</sup> FEB 2025	Dr. Sofian bin Dato Paduka Dr Haji Johar Head of Cardiology Department Raja Isteri Pengiran Anak Saleha Hospital Ministry of Health Negara Brunei Darussalam Contact No.: 2242424 ext. 4808 e-mail: sofian.johar@moh.gov.bn

# **TENDER REFERENCE NO.: KK/13/2025/UPP(TC)**

# MINISTRY OF HEALTH NEGARA BRUNEI DARUSSALAM

SUPPLY AND DELIVERY MEDICAL ITEMS (CARDIAC IMPLANTABLE ELECTRONIC DEVICE (CIED) FOR DEPARTMENT OF CARDIOLOGY RAJA ISTER PENGIRAN ANAK SALEHA (RIPAS (HOSPITAL) FOR A PERIOD OF FIVE (5) YEARS

TENDER FEES : \$500.00

RECEIPT NO. :

**CLOSING DATE**: ON TUESDAY, 18TH FEBRUARY 2025

TIME : 2.00 PM

FOA :

THE CHAIRMAN
MINI TENDER BOARD, TENDER BOX
GROUND FLOOR, MINISTRY OF HEALTH
COMMONWEALTH DRIVE
BANDAR SERI BEGAWAN BB3910
NEGARA BRUNEI DARUSSALAM

(NON CLUSTERING)

### **SECTION 2**

## **SPECIFICATIONS**

TENDER REFERENCE NO: KK/13/2025/UPP(TC)

### **INVITATION TO TENDER**

SUPPLY AND DELIVERY OF MEDICAL ITEMS (CARDIAC IMPLANTABLE ELECTRONIC DEVICE (CIED) – SUPPLEMENTAL) FOR DEPARTMENT OF CARDIOLOGY, RAJA ISTERI PENGIRAN ANAK SALEHA HOSPITALFOR A PERIOD OF FIVE (5) YEARS

DELIVERY PERIOD	NOT LATER THAN 90 DAYS
DELIVERY PERIOD	STAGGERED DELIVERY UPON REQUEST

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR
	SINGLE CHAMBER PACEMAKER WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE		0.111170
1	<ul> <li>SPECIFICATIONS         <ul> <li>MRI compatible - preferably up to 3T</li> <li>To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change</li> </ul> </li> </ul>	UNIT	3 UNITS
	SINGLE CHAMBER PACEMAKER WITHOUT REMOTE MONITORING - FOR CIED GENERATOR CHANGE		
2	SPECIFICATIONS  - MRI compatible - preferably up to 3T  - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change	UNIT	2 UNITS
	DUAL CHAMBER PERMANENT PACEMAKER WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE		
3	<ul> <li>SPECIFICATIONS         <ul> <li>MRI compatible - preferably up to 3T</li> <li>To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change</li> <li>Atrial arrhythmia therapy</li> <li>Mode-based physiologic pacing</li> </ul> </li> </ul>	UNIT	3 UNITS
	DUAL CHAMBER PERMANENT PACEMAKER WITHOUT REMOTE MONITORING - FOR CIED GENERATOR CHANGE		
4	SPECIFICATIONS     MRI compatible - preferably up to 3T     To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change     Atrial arrhythmia therapy     Mode-based physiologic pacing	UNIT	2 UNITS
5	DUAL CHAMBER PERMANENT PACEMAKER WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE	UNIT	3 UNITS

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR
	<ul> <li>SPECIFICATIONS         <ul> <li>MRI compatible - preferably up to 3T</li> <li>To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change</li> <li>Mode-based physiologic pacing</li> </ul> </li> </ul>		
	DUAL CHAMBER PERMANENT PACEMAKER WITHOUT REMOTE MONITORING - FOR CIED GENERATOR CHANGE		
6	<ul> <li>SPECIFICATIONS</li> <li>MRI compatible - preferably up to 3T</li> <li>To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change</li> <li>Mode-based physiologic pacing</li> </ul>	UNIT	2 UNITS
	SINGLE CHAMBER IMPLANTABLE CARDIOVERTER DEFIBRILLATOR WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE		
7	SPECIFICATIONS  - MRI compatible - preferably up to 3T  - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change	UNIT	2 UNITS
	Dual chamber implantable cardioverter defibrillator with remote monitoring - for CIED generator change		
8	MRI compatible - preferably up to 3T     To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change     Heart failure management algorithms	UNIT	3 UNITS
	DUAL CHAMBER IMPLANTABLE CARDIOVERTER DEFIBRILLATOR WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE		
9	SPECIFICATIONS  - MRI compatible - preferably up to 3T  - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change	UNIT	3 UNITS
	CARDIAC RESYNCHRONISATION THERAPY - PACEMAKER (CRT-P) WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE		
10	SPECIFICATIONS  - MRI compatible - preferably up to 3T  - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change	UNIT	2 UNITS
	CARDIAC RESYNCHRONISATION THERAPY - DEFIBRILLATOR (CRT-D) WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE		
11	SPECIFICATIONS  - MRI compatible - preferably up to 3T  - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change	UNIT	3 UNITS

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR
	LEADLESS PACEMAKER WITH REMOTE MONITORING		
12	SPECIFICATIONS  - MRI compatible - preferably up to 3 T  - Single or dual chamber system	UNIT	3 UNITS
13	SUBCUTANEOUS IMPLANTABLE CARDIOVERTER-DEFIBRILLATOR WITH REMOTE MONITORING  SPECIFICATIONS - MRI compatible - preferable up to 3T	UNIT	5 UNITS

NO.	TERMS AND CONDITIONS
1	Tenderer must be registered with the Ministry of Health.
2	TENDER FORM should be filled completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form <u>MAY</u> cause <b>DISQUALIFICATION OF TENDER</b> .
3	Each tenderer is allowed to quote <b>ONE BRAND WITH ONE PRICE ONLY</b> for each item. Submission of more than one brand and price will cause <b>DISQUALIFICATION OF TENDER.</b>
4	All consumables supplied throughout this tender shall have a minimum expiry date of twelve (12) months / on delivery. Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made.
5	Brochures / catalogues should be submitted / attached with tender document.
6	Samples should be submitted together with tender or within fourteen (14 days) of the tender closing date (if applicable).
7	DELIVERY PERIOD: Not later than 90 DAYS Staggered delivery upon request
8	PRICE VALIDITY: The quotation shall remain valid for 12 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).

# **SECTION 3**

# FORMS TO BE USED

# **CONTENTS**

SCHEDULE 1 – TENDER FORM	1
SCHEDULE 2 – INFORMATION SUMMARY	4
SCHEDULE 3 –SUB-CONTRACTS	5
SCHEDULE 4 –COMPANY BACKGROUND	6
SCHEDULE 5 –REFERENCES	7
SCHEDULE 6 -SUBMISSION OF SAMPLE	8
SCHEDULE 7 - LETTER OF DECLARATION	11

#### **TENDER FORM**

To:

## TENDER REFERENCE NO: KK/13/2025/UPP(TC)

### **INVITATION TO TENDER**

SUPPLY AND DELIVERY OF MEDICAL ITEMS (CARDIAC IMPLANTABLE ELECTRONIC DEVICE (CIED) – SUPPLEMENTAL) FOR DEPARTMENT OF CARDIOLOGY, RAJA ISTERI PENGIRAN ANAK SALEHA HOSPITALFOR A PERIOD OF FIVE (5) YEARS

DELIVERY PERIOD	

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS PER YEAR (B\$)	TOTAL COSTS FOR FIVE (5) YEARS (B\$)
1	SINGLE CHAMBER PACEMAKER WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change	UNIT	3 UNITS							
2	SINGLE CHAMBER PACEMAKER WITHOUT REMOTE MONITORING -	UNIT	2 UNITS							

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS PER YEAR (B\$)	TOTAL COSTS FOR FIVE (5) YEARS (B\$)
	FOR CIED GENERATOR CHANGE									
	SPECIFICATIONS  - MRI compatible - preferably up to 3T  - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change									
3	DUAL CHAMBER PERMANENT PACEMAKER WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI	UNIT	3 UNITS							
	compatible with existing Medtronic CIED leads during CIED generator change - Atrial arrhythmia therapy - Mode-based physiologic pacing									

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS PER YEAR (B\$)	TOTAL COSTS FOR FIVE (5) YEARS (B\$)
4	DUAL CHAMBER PERMANENT PACEMAKER WITHOUT REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change - Atrial arrhythmia therapy - Mode-based physiologic pacing	UNIT	2 UNITS							
5	PERMANENT PACEMAKER WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with	UNIT	3 UNITS							

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS PER YEAR (B\$)	TOTAL COSTS FOR FIVE (5) YEARS (B\$)
	existing Medtronic CIED leads during CIED generator change - Mode-based physiologic pacing									
6	DUAL CHAMBER PERMANENT PACEMAKER WITHOUT REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change - Mode-based physiologic pacing	UNIT	2 UNITS							
7	SINGLE CHAMBER IMPLANTABLE CARDIOVERTER DEFIBRILLATOR WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE	UNIT	2 UNITS							

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS PER YEAR (B\$)	TOTAL COSTS FOR FIVE (5) YEARS (B\$)
	SPECIFICATIONS  - MRI compatible - preferably up to 3T  - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change									
8	Dual chamber implantable cardioverter defibrillator with remote monitoring - for CIED generator change  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change - Heart failure management algorithms	UNIT	3 UNITS							
9	DUAL CHAMBER IMPLANTABLE CARDIOVERTER	UNIT	3 UNITS							

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS PER YEAR (B\$)	TOTAL COSTS FOR FIVE (5) YEARS (B\$)
	DEFIBRILLATOR WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE SPECIFICATIONS									
	- MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change									
	CARDIAC RESYNCHRONISATION THERAPY - PACEMAKER (CRT-P) WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE									
10	SPECIFICATIONS  - MRI compatible - preferably up to 3T  - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change	UNIT	2 UNITS							

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS PER YEAR (B\$)	TOTAL COSTS FOR FIVE (5) YEARS (B\$)
	CARDIAC RESYNCHRONISATION THERAPY - DEFIBRILLATOR (CRT- D) WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE									
11	SPECIFICATIONS  - MRI compatible - preferably up to 3T  - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change	UNIT	3 UNITS							
12	LEADLESS PACEMAKER WITH REMOTE MONITORING  SPECIFICATIONS - MRI compatible - preferably up to 3 T - Single or dual chamber system	UNIT	3 UNITS							
13	SUBCUTANEOUS IMPLANTABLE CARDIOVERTER- DEFIBRILLATOR WITH	UNIT	5 UNITS							

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY USAGE PER YEAR	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS PER YEAR (B\$)	TOTAL COSTS FOR FIVE (5) YEARS (B\$)
	REMOTE MONITORING									
	SPECIFICATIONS - MRI compatible - preferable up to 3T									
	TOTAL PRICE (B\$)									

NO.	TERMS AND CONDITIONS	VENDOR'S OFFER (PLEASE STATE)
1	Tenderer must be registered with the Ministry of Health.	
2	TENDER FORM should be filled completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form MAY cause DISQUALIFICATION OF TENDER.	
3	Each tenderer is allowed to quote <b>ONE BRAND WITH ONE PRICE ONLY</b> for each item. Submission of more than one brand and price will cause <b>DISQUALIFICATION OF TENDER</b> .	
4	All consumables supplied throughout this tender shall have a minimum expiry date of twelve (12) months / on delivery. Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made.	
5	Brochures / catalogues should be submitted / attached with tender document.	
6	Samples should be submitted together with tender or within fourteen (14 days) of the tender closing date (if applicable).	
7	DELIVERY PERIOD: Not later than 90 DAYS Staggered delivery upon request	(YES/NO) (IF NO, PLEASE SPECIFY)
8	PRICE VALIDITY: The quotation shall remain valid for 12 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).	

- 1. We offer and undertake on your acceptance of our Tender to supply and deliver the above mentioned goods in accordance with your Invitation To Tender.
- 2. Our Tender is fully consistent with and does not contradict or derogate from anything in your Invitation To Tender. We have not qualified or changed any of the provisions of your Invitation To Tender.
- 3. We shall execute a formal agreement in the appropriate form set out in Section 4 - Contract of the Invitation to Tender together with such further terms and conditions, if any, agreed between the Government and us.
- OUR OFFER IS VALID FOR TWELVE (12) CALENDER MONTHS FROM THE TENDER 4. CLOSING DATE.
- When requested by you, we shall extend the validity of this offer. 5.
- 6. We further undertake to give you any further information which you may require.

Dated thisday of	
[Signature of authorised officer of Tenderer] Name:	Tenderer's official stamp:

Designation:

## **INFORMATION SUMMARY**

- 2.1 Tenderers shall provide in this Schedule the following information:
  - a. Management summary
  - b. Company profile (including Contractor and sub-contractor(s), if any)
  - c. Years of experience (as of the Tender Closing Date) of the Contractor and subcontractor(s) in the:
    - Supply and Delivery of Medical Consumables
  - d. Other information which is considered relevant

#### **SUB-CONTRACTS**

- 3.1 Tenderers shall complete Table 3.1 with information about all the companies involved in the provision of the services and items specified in this tender. This shall include details about the Contractor and each sub-contractor involved, as well as their respective responsibilities.
- 3.2 Tenderers shall also indicate in Table 3.1 any alliance relationship established with each subcontractor. An alliance is defined as a formal and binding business relationship between the allied parties.

Table 3.1 Responsibility Table

		Alliance Relationship between Contractor and Sub-contractor(s)		
Company Name	Responsibility Description	Alliance Exists? (Y/N)	Date Established	Alliance Description
Contractor				
		Not Applicable	Not Applicable	Not Applicable
Sub-contractor(s)				

### **COMPANY'S BACKGROUND**

4.1 Each of the companies involved in this tender, including Contractor and sub-contractor(s) (if any), shall provide information on the company's background, scope of operations, financial standing and certified copy of its Certificate of Incorporation or Certificate of Registration (as the case may be).

#### **REFERENCES**

5.1 Tenderers shall submit a list of customers in Table 5.1 to whom the Contractor has provided similar services and items as specified in this tender in the recent 5 years as of the Tender Closing Date.

Table 5.1 References of previous customers

Customer Name and Address	Customer Type (Govt or Quasi Govt)*	Contact Person	Title	Contact Number, Fax Number and E- mail Address

\*Note: Tenderers shall indicate whether the customer is a Government or Quasi Government organisation. A Quasi Government is defined as an organisation which (1) is managed and controlled by the Government; or (2) has at least 50% shares being held by the Government. Please leave the column blank if the customer is neither a Government or Quasi Government organisation.

- 5.2 The Ministry of Health shall treat all the information submitted under this schedule in strict confidence.
- 5.3 The Ministry of Health reserves the right to contact the references for tender assessment purposes.

## **SUBMISSION OF SAMPLE**

- 6.1 Tenderers shall submit the Submission of Sample form below in respect of the items specified in this tender.
- 6.2 Samples of the items to be submitted shall be:
  - a) identical in packing and manufacture to the items to be offered by the Tenderer; and
  - b) marked with the corresponding item number of the tender.

## SUBMISSION OF SAMPLE FORM

To,

## TENDER REFERENCE NO: KK/13/2025/UPP(TC)

#### **INVITATION TO TENDER**

SUPPLY AND DELIVERY OF MEDICAL ITEMS (CARDIAC IMPLANTABLE ELECTRONIC DEVICE (CIED) – SUPPLEMENTAL) FOR DEPARTMENT OF CARDIOLOGY, RAJA ISTERI PENGIRAN ANAK SALEHA HOSPITALFOR A PERIOD OF FIVE (5) YEARS

## **SUBMISSION OF SAMPLE FORM OF (NAME OF TENDERER)**

\_\_\_\_\_

ITEM NO.	DESCRIPTION	SAMPLE SUBMITTED (indicate with )	SAMPLE NOT SUBMITTED (indicate with X)	OFFERED/NOT OFFERED (indicate as appropriate)
	SINGLE CHAMBER PACEMAKER WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE			
1	<ul> <li>SPECIFICATIONS</li> <li>MRI compatible - preferably up to 3T</li> <li>To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change</li> </ul>			
	SINGLE CHAMBER PACEMAKER WITHOUT REMOTE MONITORING - FOR CIED GENERATOR CHANGE			
2	<ul> <li>SPECIFICATIONS</li> <li>MRI compatible - preferably up to 3T</li> <li>To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change</li> </ul>			
	DUAL CHAMBER PERMANENT PACEMAKER WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE			
3	<ul> <li>SPECIFICATIONS</li> <li>MRI compatible - preferably up to 3T</li> <li>To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change</li> <li>Atrial arrhythmia therapy</li> <li>Mode-based physiologic pacing</li> </ul>			
4	DUAL CHAMBER PERMANENT PACEMAKER WITHOUT REMOTE MONITORING - FOR CIED GENERATOR CHANGE			
4	SPECIFICATIONS  - MRI compatible - preferably up to 3T  - To be fully MRI compatible with existing Medtronic CIED leads during CIED			

		SAMPLE	
DESCRIPTION	SAMPLE SUBMITTED (indicate with ✓)	NOT SUBMITTED (indicate	OFFERED/NOT OFFERED (indicate as appropriate)
generator change		, , ,	
- Atrial arrhythmia therapy			
, , , , , ,			
PACEMAKER WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE			
<u>SPECIFICATIONS</u>			
- MRI compatible - preferably up to 3T			
<ul> <li>Mode-based physiologic pacing</li> </ul>			
CHANGE			
Medtronic CIED leads during CIED			
CARDIOVERTER DEFIBRILLATOR WITH			
REMOTE MONITORING - FOR CIED			
GENERATOR CHANGE			
SPECIFICATIONS			
- MRI compatible - preferably up to 3T			
Dual chamber implantable cardioverter			
defibrillator with remote monitoring - for			
CIED generator change			
SPECIFICATIONS			
- MRI compatible - preferably up to 3T			
Heart failure management algorithms			
DUAL CHAMBER IMPLANTABLE			
GENERATOR CHANGE			
SPECIFICATIONS  ADDITIONS			
generator change			
	generator change Atrial arrhythmia therapy Mode-based physiologic pacing  DUAL CHAMBER PERMANENT PACEMAKER WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS MRI compatible - preferably up to 3T To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change Mode-based physiologic pacing  DUAL CHAMBER PERMANENT PACEMAKER WITHOUT REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS MRI compatible - preferably up to 3T To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change Mode-based physiologic pacing  SINGLE CHAMBER IMPLANTABLE CARDIOVERTER DEFIBRILLATOR WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS MRI compatible - preferably up to 3T To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change  Dual chamber implantable cardioverter defibrillator with remote monitoring - for CIED generator change  SPECIFICATIONS MRI compatible - preferably up to 3T To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change  Dual chamber implantable cardioverter defibrillator with remote monitoring - for CIED generator change  SPECIFICATIONS MRI compatible - preferably up to 3T To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change  Heart failure management algorithms  DUAL CHAMBER IMPLANTABLE CARDIOVERTER DEFIBRILLATOR WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS MRI compatible - preferably up to 3T To be fully MRI compatible with existing Medtronic CIED leads during CIED GENERATOR CHANGE  SPECIFICATIONS MRI compatible - preferably up to 3T To be fully MRI compatible with existing Medtronic CIED leads during CIED GENERATOR CHANGE  SPECIFICATIONS  MRI compatible - preferably up to 3T To be fully MRI compatible with existing Medtronic CIED leads during CIED	generator change - Atrial arrhythmia therapy - Mode-based physiologic pacing  DUAL CHAMBER PERMANENT PACEMAKER WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change - Mode-based physiologic pacing  DUAL CHAMBER PERMANENT PACEMAKER WITHOUT REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change - Mode-based physiologic pacing  SINGLE CHAMBER IMPLANTABLE CARDIOVERTER DEFIBRILLATOR WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change  - Heart failure management algorithms  DUAL CHAMBER IMPLANTABLE CARDIOVERTER DEFIBRILLATOR WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change  - Heart failure management algorithms  DUAL CHAMBER IMPLANTABLE CARDIOVERTER DEFIBRILLATOR WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED	generator change - Atrial arrhythmia therapy - Mode-based physiologic pacing  DUAL CHAMBER PERMANENT PACEMAKER WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic OIED leads during CIED generator change - MOde-based physiologic pacing  DUAL CHAMBER PERMANENT PACEMAKER WITHOUT REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change - Mode-based physiologic pacing  SNGLE CHAMBER IMPLANTABLE CARDIOVERTER DEFIBRILLATOR WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change  Dual chamber implantable cardioverter defibrillator with remote monitoring - for CIED generator change  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change - Heart failure management algorithms  DUAL CHAMBER IMPLANTABLE CARDIOVERTER DEFIBRILLATOR WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED generator CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED GENERATOR CHANGE  SPECIFICATIONS - MRI compatible - preferably up to 3T - To be fully MRI compatible with existing Medtronic CIED leads during CIED

ITEM NO.	DESCRIPTION	SAMPLE SUBMITTED (indicate with <	SAMPLE NOT SUBMITTED (indicate with X)	OFFERED/NOT OFFERED (indicate as appropriate)
40	CARDIAC RESYNCHRONISATION THERAPY - PACEMAKER (CRT-P) WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE			
10	<ul> <li>SPECIFICATIONS</li> <li>MRI compatible - preferably up to 3T</li> <li>To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change</li> </ul>			
	CARDIAC RESYNCHRONISATION THERAPY - DEFIBRILLATOR (CRT-D) WITH REMOTE MONITORING - FOR CIED GENERATOR CHANGE			
11	<ul> <li>SPECIFICATIONS</li> <li>MRI compatible - preferably up to 3T</li> <li>To be fully MRI compatible with existing Medtronic CIED leads during CIED generator change</li> </ul>			
12	LEADLESS PACEMAKER WITH REMOTE MONITORING			
	SPECIFICATIONS  - MRI compatible - preferably up to 3 T  - Single or dual chamber system			
13	SUBCUTANEOUS IMPLANTABLE CARDIOVERTER-DEFIBRILLATOR WITH REMOTE MONITORING			
	SPECIFICATIONS - MRI compatible - preferable up to 3T			

We understand as stated in the Instructions to Tenderers that Tenders without samples shall not be considered.

[Signature of authorized officer of Tenderer]	Tenderer's official stamp:
Name:	
Designation:	
Date:	
FOR OFFIC	E USE
Date of receipt:	
Receiving Officer:	