

Rujukan Kami: (67) MOH/HQ/P/IKLAN-SH/2025

LAMPIRAN 3

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 09.00AM)	Quotation Fee	Requesting Department	Focal Person
3	BME/A2K/(1411)/2025/HTD (A20)	ONE-OFF PREVENTIVE MAINTENANCE FOR TWO (2) UNITS OF BIOTEK ABSORBANCE MICROPLATE READERS AT FOOD CHEMISTRY LABORATORY, DEPARTMENT OF SCIENTIFIC SERVICES.	05/05/2025	24/05/2025	\$5.00	JABATAN TEKNOLOGI PENJAGAAN KESIHATAN, KEMENTERIAN KESIHATAN	HAJI ZAENUL 'ARIFIEN BIN DATO PADUKA HAJI MOHD ROSLI JURUTERA PERUBATAN HAYAT TEL: 2382424 ext. 282424 ext. 6122 / 6232

LAMPIRAN - A1

RUJ: BME/A2K/(1411)/2025/HTD (A20)

29/4/2025

ONE-OFF PREVENTIVE MAINTENANCE FOR TWO (2) UNITS OF BIOTEK ABSORBANCE MICROPLATE READERS AT FOOD CHEMISTRY LABORATORY, DEPARTMENT OF SCIENTIFIC SERVICES.

SCOPE OF WORK

One-off preventive maintenance for two (2) units of Biotek absorbance motorplate readers at Food Chemistry Laboratory, Department of Scientific Services, Serasa branch as follows:

- 1) Model: ELx808 S/N: 210286, BME6654(B)
- 2) Model: ELx800 S/N: 206961, BME119921

-Scope of work to follow equipment manual/recommendation specific for the model which include the following where necessary:

- (i) inspection
- (ii) cleaning
- (iii) installation of PM kits, consumables and spare parts listed in spareparts list (Item 1 for BME119921 in Lampiran A2)
- (iv) alignment
- (v) calibration
- (vi) any other related preventive maintenance works required.

- Functional check post-maintenance.

- Disposal of any used spare parts, unless otherwise instructed.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.

- Vendor(s) to submit maintenance checklist/details/scope specific for the equipment along with other necessary documents to join this tender.

- Upon approval, the earliest the preventive maintenance work can be carried out on the equipment is one month before the next PM due date, and not anytime earlier.

LAMPIRAN - A2

RUJ:

BME/A2K/(1411)/2025/HTD (A20)

ONE-OFF PREVENTIVE MAINTENANCE FOR TWO (2) UNITS OF BIOTEK ABSORBANCE MICROPLATE READERS AT FOOD CHEMISTRY LABORATORY, DEPARTMENT OF SCIENTIFIC SERVICES.

	SPAREPART	PART NUMBER	QUANTITY
1	630nm, 490nm, 450nm, 405nm and 340nm		