

REFERENCE OF TENDER	DESCRIPTION OF TENDER	TIME PERIOD OF TENDER	DEPARTMENT/ DIVISION/UNIT REQUESTING TENDER	FEES	CLOSING DATE NOT LATER THEN 2.00PM	FOCAL PERSON
KK/152/2025/SSBH	SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING PATIENT TROLLEY FOR SURI SERI BEGAWAN (SSB) HOSPITAL, MINISTRY OF HEALTH	-	Suri Seri Begawan (SSB) Hospital, Kuala Belait	\$10.00	1 st July 2025	Koh Eng Son Nursing Officer Hospital Suri Seri Begawan, Kuala Belait Ministry of Health Negara Brunei Darussalam Contact No.: 3335331 ext. 8717

TENDER REF. NO.: KK/152/2025/SSBH

**MINISTRY OF HEALTH
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND
COMMISSIONING PATIENT TROLLEY FOR SURI SERI
BEGAWAN (SSB) HOSPITAL, MINISTRY OF HEALTH**

TENDER FEES : \$10.00

RECEIPT NO. :

CLOSING DATE : ON TUESDAY, 01ST JULY 2025

TIME : 2.00 PM

FOA :

**THE CHAIRMAN
MINI TENDER BOARD, TENDER BOX
GROUND FLOOR, MINISTRY OF HEALTH
COMMONWEALTH DRIVE
BANDAR SERI BEGAWAN BB3910
NEGARA BRUNEI DARUSSALAM**

(CLUSTERING)

SECTION 2

SPECIFICATIONS AND REQUIREMENTS

TENDER REFERENCE NO: KK/152/2025/SSBH

INVITATION TO TENDER
SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING PATIENT TROLLEY
FOR SURI SERI BEGAWAN (SSB) HOSPITAL, MINISTRY OF HEALTH

REF. NO.	DESCRIPTION
1	THIRTY-FIVE (35) UNITS OF PATIENT TROLLEY
2	MATTRESS
3	TRAINING AND WARRANTY
4	DISTRIBUTIONS
1	THIRTY-FIVE (35) UNITS OF PATIENT TROLLEY
1.1	Hydraulic height adjustable patient trolley
1.2	Height adjustment pedal on both sides
1.3	3 sections mattress surface
1.4	Can do Trendelenburg and reverse Trendelenburg positions
1.5	Adjustable backrest and leg rest
1.6	Radio-translucent mattress platform for x-ray filming
1.8	Foldable and lockable side rails
1.9	Comes with oxygen tank holder
1.10	Foldable pushing handles for easy manoeuvrability
1.11	Main frame is electrostatic powder (epoxy) coated
1.12	Dual functional pedal for central and directional lock at foot and head sides
1.13	Height adjustable and foldable stainless-steel IV pole
1.14	Urine bag hanger hooks
1.15	Anti-static 360° swivel castors
1.16	200mm castors with central braking mechanism and steer control
1.17	Comes with monitor shelf
1.18	Safe working load: 200kg or more

REF. NO.	DESCRIPTION
1.19	Trolley Length: not more than 210cm to ensure it fits within the facility's elevator dimensions.
1.20	Trolley Width: not more than 90cm to ensure it fits within the facility's elevator dimensions.
2	MATTRESS
2.1	Removeable foam mattress
2.2	Anti-static and fire-retardant
2.3	Waterproof and compatible with the size of the stretcher
2.4	Thickness: 8cm or thicker
3	TRAINING AND WARRANTY
3.1	Tenderers to conduct training to end users by either an application specialist or competent local engineer. Tenderer must prepare a training attendance or proof of training done to end user during commissioning . Training must include but not limited to basic user operation, user troubleshooting, user daily maintenance.
3.3	Tenderer to include warranty period of at least two (2) years
3.4	Tenderers to INCLUDE a Warranty Undertaking Letter stating the terms of warranty provided for the equipment in the tender. This includes but not limited to Warranty coverage and what is excluded from Warranty
4	DISTRIBUTIONS
4.1	Twenty (24) units for Emergency Department
4.2	Four (4) units for Ward 3
4.3	Three (3) units for Ward 5
4.4	Two (2) units for Surgical Out Patient Department (SOPD) Clinic
4.5	One (1) unit for Physician Clinic
4.6	One (1) unit for ObsGynae Clinic

SECTION 2 – PROCUMENT AND TECHNICAL SPECIFICATION

BRAND:	MODEL:
COUNTRY OF ORIGIN:	UNIT PRICE (B\$):
WARRANTY PERIOD:	TOTAL PRICE (B\$):
YEAR INTRODUCED TO MARKET:	PRICE VALIDITY: [AT LEAST ONE (1) YEAR PRICE VALIDTY]
LAST COUNTRY SOLD TO:	DELIVERY TIME:
WARRANTY UNDERTAKING LETTER?	
DETAILED BROCHURE INCLUDED?	
USER AND SERVICE MANUALS:	
MAINS POWER SUPPLY:	
BATTERY:	
POWER RATINGS:	
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:	
INTERNATIONAL SAFETY STANDARD List of Safety Standards and certification (Please attached the copy of stated standards and certifications)	
NO OF TECHNICAL SUPPORT (ENGINEER/TECHNICIAN)	
DIMENSIONS AND WEIGHT OF MAIN UNIT:	
EQUIPMENT WHOLE LIFE TIME SUPPORT:	

SECTION 3

TENDER FORM

To:

TENDER REFERENCE NO: KK/152/2025/SSBH

**INVITATION TO TENDER
SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING PATIENT TROLLEY
FOR SURI SERI BEGAWAN (SSB) HOSPITAL, MINISTRY OF HEALTH**

TENDER OF (*name of tenderer*) : _____

Company/Business Registration No. : _____

Tender Closing Date : _____

REF. NO.	DESCRIPTION	Tick (✓)		STATE OR SPECIFY REMARKS/ BROCHURE PAGE
		YES	NO	
1	THIRTY-FIVE (35) UNITS OF PATIENT TROLLEY			
2	MATTRESS			
3	TRAINING AND WARRANTY			
4	DISTRIBUTIONS			
1	THIRTY-FIVE (35) UNITS OF PATIENT TROLLEY			
1.1	Hydraulic height adjustable patient trolley			
1.2	Height adjustment pedal on both sides			
1.3	3 sections mattress surface			
1.4	Can do Trendelenburg and reverse Trendelenburg positions			
1.5	Adjustable backrest and leg rest			
1.6	Radio-translucent mattress platform for x-ray filming			
1.8	Foldable and lockable side rails			
1.9	Comes with oxygen tank holder			
1.10	Foldable pushing handles for easy manoeuvrability			
1.11	Main frame is electrostatic powder (epoxy) coated			
1.12	Dual functional pedal for central and directional lock at foot and head sides			

REF. NO.	DESCRIPTION	Tick (✓)		STATE OR SPECIFY REMARKS/ BROCHURE PAGE
		YES	NO	
1.13	Height adjustable and foldable stainless-steel IV pole			
1.14	Urine bag hanger hooks			
1.15	Anti-static 360° swivel castors			
1.16	200mm castors with central braking mechanism and steer control			
1.17	Comes with monitor shelf			
1.18	Safe working load: 200kg or more			
1.19	Trolley Length: not more than 210cm to ensure it fits within the facility's elevator dimensions.			
1.20	Trolley Width: not more than 90cm to ensure it fits within the facility's elevator dimensions.			
2	MATTRESS			
2.1	Removeable foam mattress			
2.2	Anti-static and fire-retardant			
2.3	Waterproof and compatible with the size of the stretcher			
2.4	Thickness: 8cm or thicker			
3	TRAINING AND WARRANTY			
3.1	Tenderers to conduct training to end users by either an application specialist or competent local engineer. Tenderer must prepare a training attendance or proof of training done to end user during commissioning . Training must include but not limited to basic user operation, user troubleshooting, user daily maintenance.			
3.3	Tenderer to include warranty period of at least two (2) years			
3.4	Tenderers to INCLUDE a Warranty Undertaking Letter stating the terms of warranty provided for the equipment in the tender. This includes but not limited to Warranty coverage and what is excluded from Warranty			
4	DISTRIBUTIONS			
4.1	Twenty (24) units for Emergency Department			
4.2	Four (4) units for Ward 3			

REF. NO.	DESCRIPTION	Tick (✓)		STATE OR SPECIFY REMARKS/ BROCHURE PAGE
		YES	NO	
4.3	Three (3) units for Ward 5			
4.4	Two (2) units for Surgical Out Patient Department (SOPD) Clinic			
4.5	One (1) unit for Physician Clinic			
4.6	One (1) unit for ObsGynae Clinic			

SECTION 2 – PROCUREMENT AND TECHNICAL SPECIFICATION									
BRAND:				MODEL:					
COUNTRY OF ORIGIN:				UNIT PRICE (B\$):					
WARRANTY PERIOD:				TOTAL PRICE (B\$):					
YEAR INTRODUCED TO MARKET:				PRICE VALIDITY: [AT LEAST ONE (1) YEAR PRICE VALIDITY]					
LAST COUNTRY SOLD TO:				DELIVERY TIME:					
WARRANTY UNDERTAKING LETTER?		YES		NO	<input checked="" type="checkbox"/> Where appropriate				
DETAILED BROCHURE INCLUDED?		YES		NO	<input checked="" type="checkbox"/> Where appropriate				
USER AND SERVICE MANUALS:		YES		NO	Tenderers to acknowledge that they must provide at least TWO sets of USER AND SERVICE manuals when applying commissioning form. One Set for End User, One Set for BME. (Please provide hardcopy or softcopy)				
MAINS POWER SUPPLY:		220V-240V			OTHERS:				
		50-60HZ			OTHERS:				
BATTERY		RECHARGEABLE				SINGLE-USE			REPLACEABLE
		OTHERS:							
	TYPE OF BATTERY:								
POWER RATINGS:	POWER ADAPTER/CHARGER:								
	BATTERY:								
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:									
INTERNATIONAL SAFETY STANDARD List of Safety Standards and certification (Please attached the copy of stated standards and certifications)									
NO OF TECHNICAL SUPPORT (ENGINEER/TECHNICIAN)	LOCAL								
	OVERSEA (SPECIFY LOCATION)				LOCATION:				
DIMENSIONS AND WEIGHT OF MAIN UNIT:					MM / CM / INCH				
					KG / G / LBS				
EQUIPMENT WHOLE LIFE TIME SUPPORT:	Number of years, spare parts are available after the installation of the equipment: _____ years								

1. We offer and undertake on your acceptance of our Tender to provide the above-mentioned services in accordance with your Invitation To Tender.
2. Our Tender is fully consistent with and does not contradict or derogate from anything in your Invitation To Tender. We have not qualified or changed any of the provisions of your Invitation To Tender.
3. OUR OFFER IS VALID FOR **TWELVE (12)** CALENDAR MONTHS FROM THE TENDER CLOSING DATE.
4. When requested by you, we shall extend the validity of this offer.
5. We further undertake to give you any further information which you may require.

Dated this _____ day of _____, 2025.

Signature of authorised officer of Tenderer

Name:

Designation:

Tenderer's official stamp