

Rujukan Kami: (94) MOH/HQ/P/IKLAN-SH/2025

LAMPIRAN 4

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 09.00AM)	Quotation Fee	Requesting Department
4	BME/A2K/(1470)/2025/HTD (B8)	TO REPAIR ONE (1) UNIT OF RECLINING CHAIR, BRAND: DIGITERM, MODEL: COMFORT 4 AT LABOUR ROOM WCC RIPAS HOSPITAL.	03/07/2025	19/07/2025	\$5.00	JABATAN TEKNOLOGI PENJAGAAN KESIHATAN, KEMENTERIAN KESIHATAN.

LAMPIRAN - A1

RUJ: BME/A2K/(1470)/2025/HTD (B8)

24/6/2025

TO REPAIR ONE (1) UNIT OF RECLINING CHAIR, BRAND: DIGITERM, MODEL: COMFORT 4 AT LABOUR ROOM WCC RIPAS HOSPITAL.

SCOPE OF WORK

To repair one (1) unit of Reclining chair, Brand: Digiterm, Model: Comfort 4 at Labour Room WCC Ripas Hospital. Detail are as follow:

1. BME17636, Brand: Digiterm, Model: Comfort 4, SN: C4E4C20141179, at Labour Room, W/order 2025-6719

- Scope of work to follow equipment manual/recommendation specific for this system which include the following where if necessary:
 - (i) installation of items listed in spareparts list (Lampiran A2)
 - (ii) alignment
 - (iii) calibration
 - (iv) optimisation
 - (vi) any other related maintenance works required after installation of items listed in spareparts list (Lampiran A2)
- Perform functional test, verification and performance check post-maintenance.
- Disposal of any used spare parts, unless otherwise instructed.
- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.

LAMPIRAN - A2

RUJ: BME/A2K/(1470)/2025/HTD (B8)

TO REPAIR ONE (1) UNIT OF RECLINING CHAIR, BRAND: DIGITERM, MODEL: COMFORT 4 AT LABOUR ROOM WCC RIPAS HOSPITAL.

	SPAREPART	PART NUMBER	QUANTITY
1	UPHOLSTERY 80MM (PALMA)	34545306	1
2	ARMREST MECHANISM FOR TILT-UP VERSION	34545212	1
3	HANDSET, PLAST, GREY, IP66	34545149	1