

REFERENCE OF TENDER	DESCRIPTION OF TENDER	TIME PERIOD OF TENDER	DEPARTMENT/ DIVISION/UNIT REQUESTING TENDER	FEES	CLOSING DATE NOT LATER THAN 2.00PM	FOCAL PERSON
KK/195/2025/SMS(TC)	SUPPLY AND DELIVERY OF MEDICAL CONSUMABLES ITEM (CSSD CONSUMABLES) FOR THE STATE MEDICAL STORE FOR A PERIOD OF THREE (3) YEARS	3 years	Unit Pembelian Gunahabis Perubatan	\$50.00	05 AUGUST 2025	Siti Mahirah Liyana binti Haji Emran Supply Officer Ministry of Health Negara Brunei Darussalam Contact No: 2381979 e-mail: mahirah.emran@moh.gov.bn

**TENDER REFERENCE NO.: KK/195/2025/SMS(TC)**

**MINISTRY OF HEALTH  
NEGARA BRUNEI DARUSSALAM**

**SUPPLY AND DELIVERY OF MEDICAL CONSUMABLES  
ITEM (CSSD CONSUMABLES) FOR THE STATE MEDICAL  
STORE FOR A PERIOD OF THREE (3) YEARS**

**TENDER FEES : \$50.00**

**RECEIPT NO. :**

**CLOSING DATE : ON TUESDAY, 05TH AUGUST 2025**

**TIME : 2.00 PM**

**FOA :**

**THE CHAIRMAN  
MINI TENDER BOARD, TENDER BOX  
GROUND FLOOR, MINISTRY OF HEALTH  
COMMONWEALTH DRIVE  
BANDAR SERI BEGAWAN BB3910  
NEGARA BRUNEI DARUSSALAM**

**(NON CLUSTERING)**

## SECTION 2

### SPECIFICATIONS

TENDER REFERENCE NO: KK/195/2025/SMS(TC)

INVITATION TO TENDER  
SUPPLY AND DELIVERY OF MEDICAL CONSUMABLES ITEM (CSSD CONSUMABLES) FOR  
THE STATE MEDICAL STORE FOR A PERIOD OF THREE (3) YEARS

DELIVERY PERIOD	First delivery within 2 weeks, then staggered deliveries will occur continuously throughout the contract. Please ensure that you maintain a buffer stock of at least 2 months supply at all times to prevent any disruptions in supply.
-----------------	--

NO.	ITEM	ESTIMATED ONE YEAR USAGE	BUFFER STOCK
1	Sterilization Pouches In Fleet Reel Size:75mm x 200m (R1) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	250 ROLLS	40 ROLLS
2	Sterilization Pouches In Fleet Reel Size: 100mm x 200m (R2) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	250 ROLLS	40 ROLLS
3	Sterilization Pouches in Fleet Reel: Size: 150mm x 200m (R4) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	300 ROLLS	50 ROLLS
4	Sterilization Pouches in Fleet Reel: Size: 250mm x 200m (R5) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	250 ROLLS	40 ROLLS
5	Sterilization Pouches in Fleet Reel: Size: 300mm x 200m (R6) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	300 ROLLS	50 ROLLS
6	Heat Seal Sterilization Bag: Code C (Size: 110mm x 30mm x 190mm) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	220,000 PCS	36,000 PCS
7	Heat Seal Sterilization Bag: Code D (Size: 140mm x 30mm x 250mm) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	140,000 PCS	23,000 PCS
8	Heat Seal Sterilization Bag: Code L (Size: 250mm x 100mm x 380mm) - Code M34 <b>SPECIFICATIONS: REFER ATTACHMENT</b>	40,000 PCS	6,500 PCS
9	Heat Seal Sterilization Bag: Code T (Size: 380mm x 125mm x 510mm) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	15,000 PCS	2,500 PCS
10	Sterilization Pouches with Gussetted in Fleet Reel: Size: 300mm x 50mm / 65mm x 200m (RG12) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	50 ROLLS	8 ROLLS

NO.	ITEM	ESTIMATED ONE YEAR USAGE	BUFFER STOCK
11	Sterilization Pouches with Gussetted in Fleet Reel: Size: 380mm x 125mm / 150mm x 200m (RG15)  <b>SPECIFICATIONS: REFER ATTACHMENT</b>	50 ROLLS	8 ROLLS

COMPULSORY REQUIREMENTS	
▪ FOR MEDICAL PURPOSES	
▪ PRICE QUOTED SHOULD NOT BE MORE THAN 2 DECIMAL PLACES	
▪ PRICE VALIDITY MUST BE ONE (1) YEAR	
▪ CERTIFICATE OF AUTHENTICITY (MUST BE ATTACHED)	
▪ AUTHORIZATION LETTER AS A DISTRIBUTOR (MUST BE ATTACHED)	
▪ KINDLY SUBMIT OFFICIAL BROCHURES FOR OFFERED ITEMS	
▪ SAMPLE MUST BE SUBMITTED WITHIN TWO (2) WEEKS AFTER THE CLOSING DATE OR ELSE THE PARTICIPATION WILL BE INVALID	
▪ DO NOT SUBMIT EXPIRED SAMPLE	

ITEM NO.	ITEM DESCRIPTION	SPECIFICATIONS
1	Sterilization Pouches In Fleet Reel Size: 75mm x 200m (R1)	▪ View pack flat peelable sterilization pouch
		▪ Have steam autoclave indicator
		▪ Have formaldehyde gas indicator
		▪ Consists of shatterless film and paper at the back (manufactured from high wet strength bleach kraft paper)
		▪ Size: 75mm x 200m
		▪ Not easily torn
		▪ Fit to the machine of Steam Autoclave, Formaldehyde Gas
2	Sterilization Pouches In Fleet Reel Size: 100mm x 200m (R2)	▪ View pack flat peelable sterilization pouch
		▪ Have steam autoclave indicator
		▪ Have formaldehyde gas indicator
		▪ Consists of shatterless film and paper at the back (manufactured from high wet strength bleach kraft paper)
		▪ Size: 100mm x 200m
		▪ Not easily torn
		▪ Fit to the machine of Steam Autoclave, Formaldehyde Gas
3	Sterilization Pouches in Fleet Reel: Size: 150mm x 200m (R4)	▪ View pack flat peelable sterilization pouch
		▪ Have steam autoclave indicator
		▪ Have formaldehyde gas indicator
		▪ Consists of shatterless film and paper at the back (manufactured from high wet strength bleach kraft paper)

ITEM NO.	ITEM DESCRIPTION	SPECIFICATIONS
		<ul style="list-style-type: none"> <li>Size: 150mm x 200m</li> <li>Not easily torn</li> <li>Fit to the machine of Steam Autoclave, Formaldehyde Gas</li> </ul>
4	Sterilization Pouches in Fleet Reel: Size: 250mm x 200m (R5)	<ul style="list-style-type: none"> <li>View pack flat peelable sterilization pouch</li> <li>Have steam autoclave indicator</li> <li>Have formaldehyde gas indicator</li> <li>Consists of shatterless film and paper at the back (manufactured from high wet strength bleach kraft paper)</li> <li>Size: 250mm x 200m</li> <li>Not easily torn</li> <li>Fit to the machine of Steam Autoclave, Formaldehyde Gas</li> </ul>
5	Sterilization Pouches in Fleet Reel: Size: 300mm x 200m (R6)	<ul style="list-style-type: none"> <li>View pack flat peelable sterilization pouch</li> <li>Have steam autoclave indicator</li> <li>Have formaldehyde gas indicator</li> <li>Consists of shatterless film and paper at the back (manufactured from high wet strength bleach kraft paper)</li> <li>Size: 300mm x 200m</li> <li>Not easily torn</li> <li>Fit to the machine of Steam Autoclave, Formaldehyde Gas</li> </ul>
6	Heat Seal Sterilization Bag: Code C (Size: 110mm x 30mm x 190mm)	<ul style="list-style-type: none"> <li>Heat seal sterilization bag</li> <li>Have steam autoclave indicator</li> <li>High quality paper bag</li> <li>Manufactured from wet strength bleach kraft paper</li> <li>Size: 110mm x 30mm x 190mm</li> <li>Not easily torn</li> <li>Fit to machine of Steam Autoclave, Formaldehyde Gas</li> </ul>
7	Heat Seal Sterilization Bag: Code D (Size: 140mm x 30mm x 250mm)	<ul style="list-style-type: none"> <li>Heat seal sterilization bag</li> <li>Have steam autoclave indicator</li> <li>High quality paper bag</li> <li>Manufactured from wet strength bleach kraft paper</li> <li>Size: 140mm x 30mm x 250mm</li> <li>Not easily torn</li> <li>Fit to machine of Steam Autoclave, Formaldehyde Gas</li> </ul>
8	Heat Seal Sterilization Bag: Code L (Size: 250mm x 100mm x 380mm) - Code M34	<ul style="list-style-type: none"> <li>Heat seal sterilization bag</li> <li>Have steam autoclave indicator</li> <li>High quality paper bag</li> </ul>

ITEM NO.	ITEM DESCRIPTION	SPECIFICATIONS
		<ul style="list-style-type: none"> <li>▪ Manufactured from wet strength bleach kraft paper</li> <li>▪ Size: 250mm x 100mm x 380mm</li> <li>▪ Not easily torn</li> <li>▪ Fit to machine of Steam Autoclave, Formaldehyde Gas</li> </ul>
9	Heat Seal Sterilization Bag: Code T (Size: 380mm x 125mm x 510mm)	<ul style="list-style-type: none"> <li>▪ Heat seal sterilization bag</li> <li>▪ Have steam autoclave indicator</li> <li>▪ High quality paper bag</li> <li>▪ Manufactured from wet strength bleach kraft paper</li> <li>▪ Size: 380mm x 125mm x 510mm</li> <li>▪ Not easily torn</li> <li>▪ Fit to machine of Steam Autoclave, Formaldehyde Gas</li> </ul>
10	Sterilization Pouches with Gussetted in Fleet Reel: Size: 300mm x 50mm / 65mm x 200m (RG12)	<ul style="list-style-type: none"> <li>▪ View pack flat peelable sterilization pouch</li> <li>▪ Have steam autoclave indicator</li> <li>▪ Have formaldehyde gas indicator</li> <li>▪ Consists of shatterless film and paper at the back</li> <li>▪ Manufactured from wet strength black kraft paper</li> <li>▪ Size: 300mm x 50mm / 65mm x 200mm</li> <li>▪ Not easily torn</li> <li>▪ Fit to the machine of Steam Autoclave, Formaldehyde Gas</li> </ul>
11	Sterilization Pouches with Gussetted in Fleet Reel: Size: 380mm x 125mm / 150mm x 200m (RG15)	<ul style="list-style-type: none"> <li>▪ View pack flat peelable sterilization pouch</li> <li>▪ Have steam autoclave indicator</li> <li>▪ Have formaldehyde gas indicator</li> <li>▪ Consists of shatterless film and paper at the back</li> <li>▪ Manufactured from wet strength black kraft paper</li> <li>▪ Size: 380mm x 125mm / 150mm x 200mm</li> <li>▪ Not easily torn</li> <li>▪ Fit to the machine of Steam Autoclave, Formaldehyde Gas</li> </ul>

**SECTION 3**  
**FORMS TO BE USED**

**CONTENTS**

SCHEDULE 1 – TENDER FORM .....	2-10
SCHEDULE 2 – INFORMATION SUMMARY .....	11
SCHEDULE 3 –SUB-CONTRACTS .....	12
SCHEDULE 4 –COMPANY BACKGROUND .....	13
SCHEDULE 5 –REFERENCES .....	14
SCHEDULE 6 –SUBMISSION OF SAMPLE ... ..	15-17
SCHEDULE 7 - LETTER OF DECLARATION .....	



**SCHEDULE 1**

**TENDER FORM**

To:

**TENDER REFERENCE NO: KK/195/2025/SMS(TC)**

**INVITATION TO TENDER  
SUPPLY AND DELIVERY OF MEDICAL CONSUMABLES ITEM (CSSD CONSUMABLES) FOR THE STATE MEDICAL STORE FOR A PERIOD OF  
THREE (3) YEARS**

---

**TENDER OF (*name of tenderer*)** \_\_\_\_\_

Company/Business Registration No \_\_\_\_\_

Tender Closing Date: \_\_\_\_\_

ITEM NO.	DESCRIPTION	ESTIMATED YEARLY USAGE	BUFFER STOCK	BRAND	PACKING SIZE	UNIT PRICE (B\$)	TOTAL PRICE (B\$)	DELIVERY PERIOD
1	Sterilization Pouches In Fleet Reel Size:75mm x 200m (R1)  <b>SPECIFICATIONS: REFER ATTACHMENT</b>	250 ROLLS	40 ROLLS					
2	Sterilization Pouches In Fleet Reel Size: 100mm x 200m (R2)  <b>SPECIFICATIONS: REFER ATTACHMENT</b>	250 ROLLS	40 ROLLS					
3	Sterilization Pouches in Fleet Reel: Size: 150mm x 200m (R4)  <b>SPECIFICATIONS: REFER ATTACHMENT</b>	300 ROLLS	50 ROLLS					

ITEM NO.	DESCRIPTION	ESTIMATED YEARLY USAGE	BUFFER STOCK	BRAND	PACKING SIZE	UNIT PRICE (B\$)	TOTAL PRICE (B\$)	DELIVERY PERIOD
4	Sterilization Pouches in Fleet Reel: Size: 250mm x 200m (R5)  <b>SPECIFICATIONS: REFER ATTACHMENT</b>	250 ROLLS	40 ROLLS					
5	Sterilization Pouches in Fleet Reel: Size: 300mm x 200m (R6)  <b>SPECIFICATIONS: REFER ATTACHMENT</b>	300 ROLLS	50 ROLLS					
6	Heat Seal Sterilization Bag: Code C (Size: 110mm x 30mm x 190mm)  <b>SPECIFICATIONS: REFER ATTACHMENT</b>	220,000 PCS	36,000 PCS					
7	Heat Seal Sterilization Bag: Code D (Size: 140mm x 30mm x 250mm)  <b>SPECIFICATIONS: REFER ATTACHMENT</b>	140,000 PCS	23,000 PCS					
8	Heat Seal Sterilization Bag: Code L (Size: 250mm x 100mm x 380mm) - Code M34  <b>SPECIFICATIONS: REFER ATTACHMENT</b>	40,000 PCS	6,500 PCS					

ITEM NO.	DESCRIPTION	ESTIMATED YEARLY USAGE	BUFFER STOCK	BRAND	PACKING SIZE	UNIT PRICE (B\$)	TOTAL PRICE (B\$)	DELIVERY PERIOD
9	Heat Seal Sterilization Bag: Code T (Size: 380mm x 125mm x 510mm)  <b>SPECIFICATIONS: REFER ATTACHMENT</b>	15,000 PCS	2,500 PCS					
10	Sterilization Pouches with Gussetted in Fleet Reel: Size: 300mm x 50mm / 65mm x 200m (RG12)  <b>SPECIFICATIONS: REFER ATTACHMENT</b>	50 ROLLS	8 ROLLS					
11	Sterilization Pouches with Gussetted in Fleet Reel: Size: 380mm x 125mm / 150mm x 200m (RG15)  <b>SPECIFICATIONS: REFER ATTACHMENT</b>	50 ROLLS	8 ROLLS					
<b>TOTAL PRICE PER YEAR (B\$)</b>								
<b>TOTAL PRICE FOR THREE (3) YEARS (B\$)</b>								

## COMPULSORY REQUIREMENTS:

COMPULSORY REQUIREMENTS	YES (√)	NO (X)
▪ FOR MEDICAL PURPOSES		
▪ PRICE QUOTED SHOULD NOT BE MORE THAN 2 DECIMAL PLACES		
▪ PRICE VALIDITY MUST BE ONE (1) YEAR		
▪ CERTIFICATE OF AUTHENTICITY (MUST BE ATTACHED)		
▪ AUTHORIZATION LETTER AS A DISTRIBUTOR (MUST BE ATTACHED)		
▪ KINDLY SUBMIT OFFICIAL BROCHURES FOR OFFERED ITEMS		
▪ SAMPLE MUST BE SUBMITTED WITHIN TWO (2) WEEKS AFTER THE CLOSING DATE OR ELSE THE PARTICIPATION WILL BE INVALID		
▪ DO NOT SUBMIT EXPIRED SAMPLE		

ITEM NO.	ITEM DESCRIPTION	SPECIFICATIONS	YES (√)	NO (X)
1	Sterilization Pouches In Fleet Reel Size:75mm x 200m (R1)	▪ View pack flat peelable sterilization pouch		
		▪ Have steam autoclave indicator		
		▪ Have formaldehyde gas indicator		
		▪ Consists of shatterless film and paper at the back (manufactured from high wet strength bleach kraft paper)		
		▪ Size: 75mm x 200m		
		▪ Not easily torn		
		▪ Fit to the machine of Steam Autoclave, Formaldehyde Gas		
2	Sterilization Pouches In Fleet Reel Size: 100mm x 200m (R2)	▪ View pack flat peelable sterilization pouch		
		▪ Have steam autoclave indicator		
		▪ Have formaldehyde gas indicator		
		▪ Consists of shatterless film and paper at the back (manufactured from high wet strength bleach kraft		

ITEM NO.	ITEM DESCRIPTION	SPECIFICATIONS	YES (√)	NO (X)
		paper)		
		▪ Size: 100mm x 200m		
		▪ Not easily torn		
		▪ Fit to the machine of Steam Autoclave, Formaldehyde Gas		
3	Sterilization Pouches in Fleet Reel: Size: 150mm x 200m (R4)	▪ View pack flat peelable sterilization pouch		
		▪ Have steam autoclave indicator		
		▪ Have formaldehyde gas indicator		
		▪ Consists of shatterless film and paper at the back (manufactured from high wet strength bleach kraft paper)		
		▪ Size: 150mm x 200m		
		▪ Not easily torn		
		▪ Fit to the machine of Steam Autoclave, Formaldehyde Gas		
4	Sterilization Pouches in Fleet Reel: Size: 250mm x 200m (R5)	▪ View pack flat peelable sterilization pouch		
		▪ Have steam autoclave indicator		
		▪ Have formaldehyde gas indicator		
		▪ Consists of shatterless film and paper at the back (manufactured from high wet strength bleach kraft paper)		
		▪ Size: 250mm x 200m		
		▪ Not easily torn		
		▪ Fit to the machine of Steam Autoclave, Formaldehyde Gas		
5	Sterilization Pouches in Fleet Reel: Size: 300mm x 200m (R6)	▪ View pack flat peelable sterilization pouch		
		▪ Have steam autoclave indicator		
		▪ Have formaldehyde gas indicator		

ITEM NO.	ITEM DESCRIPTION	SPECIFICATIONS	YES (√)	NO (X)
		▪ Consists of shatterless film and paper at the back (manufactured from high wet strength bleach kraft paper)		
		▪ Size: 300mm x 200m		
		▪ Not easily torn		
		▪ Fit to the machine of Steam Autoclave, Formaldehyde Gas		
6	Heat Seal Sterilization Bag: Code C (Size: 110mm x 30mm x 190mm)	▪ Heat seal sterilization bag		
		▪ Have steam autoclave indicator		
		▪ High quality paper bag		
		▪ Manufactured from wet strength bleach kraft paper		
		▪ Size: 110mm x 30mm x 190mm		
		▪ Not easily torn		
		▪ Fit to machine of Steam Autoclave, Formaldehyde Gas		
7	Heat Seal Sterilization Bag: Code D (Size: 140mm x 30mm x 250mm)	▪ Heat seal sterilization bag		
		▪ Have steam autoclave indicator		
		▪ High quality paper bag		
		▪ Manufactured from wet strength bleach kraft paper		
		▪ Size: 140mm x 30mm x 250mm		
		▪ Not easily torn		
		▪ Fit to machine of Steam Autoclave, Formaldehyde Gas		
8	Heat Seal Sterilization Bag: Code L (Size: 250mm x 100mm x 380mm) - Code M34	▪ Heat seal sterilization bag		
		▪ Have steam autoclave indicator		
		▪ High quality paper bag		
		▪ Manufactured from wet strength bleach kraft paper		

ITEM NO.	ITEM DESCRIPTION	SPECIFICATIONS	YES (√)	NO (X)
		▪ Size: 250mm x 100mm x 380mm		
		▪ Not easily torn		
		▪ Fit to machine of Steam Autoclave, Formaldehyde Gas		
9	Heat Seal Sterilization Bag: Code T (Size: 380mm x 125mm x 510mm)	▪ Heat seal sterilization bag		
		▪ Have steam autoclave indicator		
		▪ High quality paper bag		
		▪ Manufactured from wet strength bleach kraft paper		
		▪ Size: 380mm x 125mm x 510mm		
		▪ Not easily torn		
		▪ Fit to machine of Steam Autoclave, Formaldehyde Gas		
10	Sterilization Pouches with Gussetted in Fleet Reel: Size: 300mm x 50mm / 65mm x 200m (RG12)	▪ View pack flat peelable sterilization pouch		
		▪ Have steam autoclave indicator		
		▪ Have formaldehyde gas indicator		
		▪ Consists of shatterless film and paper at the back		
		▪ Manufactured from wet strength black kraft paper		
		▪ Size: 300mm x 50mm / 65mm x 200mm		
		▪ Not easily torn		
		▪ Fit to the machine of Steam Autoclave, Formaldehyde Gas		
11	Sterilization Pouches with Gussetted in Fleet Reel: Size: 380mm x 125mm / 150mm x 200m (RG15)	▪ View pack flat peelable sterilization pouch		
		▪ Have steam autoclave indicator		
		▪ Have formaldehyde gas indicator		
		▪ Consists of shatterless film and paper at the back		

ITEM NO.	ITEM DESCRIPTION	SPECIFICATIONS	YES (√)	NO (X)
		▪ Manufactured from wet strength black kraft paper		
		▪ Size: 380mm x 125mm / 150mm x 200mm		
		▪ Not easily torn		
		▪ Fit to the machine of Steam Autoclave, Formaldehyde Gas		



1. We offer and undertake on your acceptance of our Tender to supply and deliver the above mentioned goods in accordance with your Invitation To Tender.
2. Our Tender is fully consistent with and does not contradict or derogate from anything in your Invitation To Tender. We have not qualified or changed any of the provisions of your Invitation To Tender.
3. We shall execute a formal agreement in the appropriate form set out in Section 4 – Contract of the Invitation to Tender together with such further terms and conditions, if any, agreed between the Government and us.
4. OUR OFFER IS VALID FOR **TWELVE (12) MONTHS** CALENDER MONTHS FROM THE TENDER CLOSING DATE.
5. When requested by you, we shall extend the validity of this offer.
6. We further undertake to give you any further information which you may require.

Dated this \_\_\_\_\_ day of \_\_\_\_\_, 20\_\_\_\_\_.

\_\_\_\_\_  
*[Signature of authorised officer of Tenderer]*

Name:

Designation:

Tenderer's official stamp:

## **SCHEDULE 2 - INFORMATION SUMMARY**

1.1 Tenderers shall provide in this Schedule the following information:

- (a) Management summary
- (b) Company profile (including Contractor and sub-contractor(s), if any)
- (c) Years of experience (as of the Tender Closing Date) of the Contractor and sub-contractor(s) in the:
  - *Supply and Delivery of Medical Supplies*
- (d) Other information which is considered relevant

### SCHEDULE 3 – SUB-CONTRACTS

- 3.1 Tenderers shall complete Table 3.1 with information about all the companies involved in the provision of the services and items specified in this tender. This shall include details about the Contractor and each sub-contractor involved, as well as their respective responsibilities.
- 3.2 Tenderers shall also indicate in Table 3.1 any alliance relationship established with each sub-contractor. An alliance is defined as a formal and binding business relationship between the allied parties.

Table 3.1      Responsibility Table

Company Name	Responsibility Description	Alliance Relationship between Contractor and Sub-contractor(s)		
		Alliance Exists? (Y/N)	Date Established	Alliance Description
Contractor				
		Not Applicable	Not Applicable	Not Applicable
Sub-contractor(s)				

#### **SCHEDULE 4 – COMPANY’S BACKGROUND**

- 4.1 Each of the companies involved in this tender, including Contractor and sub-contractor(s) (if any), shall provide information on the company’s background, scope of operations, financial standing and certified copy of its Certificate of Incorporation or Certificate of Registration (as the case may be).

## SCHEDULE 5 - REFERENCES

- 5.1 Tenderers shall submit a list of customers in Table 5.1 to whom the Contractor has provided similar services and items as specified in this tender in the recent 5 years as of the Tender Closing Date.

Table 5.1 References of previous customers

Customer Name and Address	Customer Type (Govt or Quasi Govt)*	Contact Person	Title	Contact Number, Fax Number and E-mail Address

**\*Note:** Tenderers shall indicate whether the customer is a Government or Quasi Government organisation. A Quasi Government is defined as an organisation which (1) is managed and controlled by the Government; or (2) has at least 50% shares being held by the Government. Please leave the column blank if the customer is neither a Government or Quasi Government organisation.

- 5.2 The Ministry of Health shall treat all the information submitted under this schedule in strict confidence.
- 5.3 The Ministry of Health reserves the right to contact the references for tender assessment purposes.

#### **SCHEDULE 6 - SUBMISSION OF SAMPLE**

- 6.1 Tenderers shall submit the Submission of Sample form below in respect of the items specified in this tender.
- 6.2 Samples of the items to be submitted shall be:
  - a. identical in packing and manufacture to the items to be offered by the Tenderer; and
  - b. marked with the corresponding item number of the tender.

**SUBMISSION OF SAMPLE FORM**

To:

**TENDER REFERENCE NO: KK/195/2025/SMS(TC)**

**INVITATION TO TENDER  
SUPPLY AND DELIVERY OF MEDICAL CONSUMABLES ITEM (CSSD CONSUMABLES) FOR  
THE STATE MEDICAL STORE FOR A PERIOD OF THREE (3) YEARS**

**SUBMISSION OF SAMPLE FORM OF (NAME OF TENDERER)**

---

ITEM NO.	DESCRIPTION	SAMPLE SUBMITTED (indicate with ✓)	SAMPLE NOT SUBMITTED (indicate with X)	OFFERED/NOT OFFERED (indicate as appropriate)
1	Sterilization Pouches In Fleet Reel Size:75mm x 200m (R1) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	250 ROLLS	40 ROLLS	
2	Sterilization Pouches In Fleet Reel Size: 100mm x 200m (R2) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	250 ROLLS	40 ROLLS	
3	Sterilization Pouches in Fleet Reel: Size: 150mm x 200m (R4) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	300 ROLLS	50 ROLLS	
4	Sterilization Pouches in Fleet Reel: Size: 250mm x 200m (R5) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	250 ROLLS	40 ROLLS	
5	Sterilization Pouches in Fleet Reel: Size: 300mm x 200m (R6) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	300 ROLLS	50 ROLLS	
6	Heat Seal Sterilization Bag: Code C (Size: 110mm x 30mm x 190mm) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	220,000 PCS	36,000 PCS	
7	Heat Seal Sterilization Bag: Code D (Size: 140mm x 30mm x 250mm) <b>SPECIFICATIONS: REFER ATTACHMENT</b>	140,000 PCS	23,000 PCS	
8	Heat Seal Sterilization Bag: Code L (Size: 250mm x 100mm x 380mm) - Code M34	40,000 PCS	6,500 PCS	

ITEM NO.	DESCRIPTION	SAMPLE SUBMITTED (indicate with ✓)	SAMPLE NOT SUBMITTED (indicate with ✕)	OFFERED/NOT OFFERED (indicate as appropriate)
	<b>SPECIFICATIONS: REFER ATTACHMENT</b>			
9	Heat Seal Sterilization Bag: Code T (Size: 380mm x 125mm x 510mm)  <b>SPECIFICATIONS: REFER ATTACHMENT</b>	15,000 PCS	2,500 PCS	
10	Sterilization Pouches with Gussetted in Fleet Reel: Size: 300mm x 50mm / 65mm x 200m (RG12)  <b>SPECIFICATIONS: REFER ATTACHMENT</b>	50 ROLLS	8 ROLLS	
11	Sterilization Pouches with Gussetted in Fleet Reel: Size: 380mm x 125mm / 150mm x 200m (RG15)  <b>SPECIFICATIONS: REFER ATTACHMENT</b>	50 ROLLS	8 ROLLS	

We understand as stated in the Instructions to Tenderers that Tenders without samples shall not be considered.

Tenderer's official stamp:

\_\_\_\_\_  
[Signature of authorized officer of Tenderer]

Name:

Designation:

Date:

**FOR OFFICE USE**

Date of receipt : \_\_\_\_\_

Receiving Officer : \_\_\_\_\_