

LAMPIRAN - A1

RUJ: BME/A2K/(1481)/2025/HTD (B8)

4/8/2025

TO REPAIR, REPLACE AND INSTALL SPARE PARTS ONE (1) UNIT OF C-MAC INTUBATION FIBERSCOPE FOR CMOS ENDOSCOPES, BRAND: KARL STORZ, MODEL: 8403ZX, S/N: ON2460, BME29654 AT OPERATING THEATRE RIPAS HOSPITAL.

SCOPE OF WORK

To repair, replace and install spare parts one (1) unit of C-Mac intubation Fibrescope for C-Mos endoscope, Brand: Karl Storz, Model: 8403ZX at OT Ripas Hospital. Detail are as follow:

1. BME29654, Brand: Karl Storz, Model: 8403ZX, SN: ON2460, at Operating Theatre, W/order 2025-6287

- Scope of work to follow equipment manual/recommendation specific for this system which include the following where if necessary:

- (i) installation of items listed in spareparts list (Lampiran A2)
- (ii) alignment
- (iii) calibration
- (iv) optimisation
- (vi) any other related maintenance works required after installation of items listed in spareparts list (Lampiran A2)

- Perform functional test, verification and performance check post-maintenance.

- Disposal of any used spare parts, unless otherwise instructed.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.

LAMPIRAN - A2

RUJ: BME/A2K/(1481)/2025/HTD (B8)

TO REPAIR, REPLACE AND INSTALL SPARE PARTS ONE (1) UNIT OF C-MAC INTUBATION FIBERSCOPE FOR CMOS ENDOSCOPES, BRAND: KARL STORZ, MODEL: 8403ZX, S/N: ON2460, BME29654 AT OPERATING THEATRE RIPAS HOSPITAL.

| | SPAREPART | PART NUMBER | QUANTITY |
|---|---------------------------------|--------------------|-----------------|
| 1 | C-MAC VIDEO LARYNGOSCOPE MAC #4 | 8404BX | 1 |
| 2 | ADAPTER 12V 2500mA | | 1 |