

Rujukan Kami: **(110) MOH/HQ/P/IKLAN-SH/2025**

LAMPIRAN 7

BIL	Quotation Reference	Description	Advertisement Date	Closing Date (Not Later Than 09.00AM)	Quotation Fee	Requesting Department
7	PPM/PROC/2025/<50K/007(SSBH) - IKN	TO SUPPLY, DELIVER, INSTALL AND COMMISSION ELECTRIC HYDRAULIC BODYLIFTER FOR MORTUARY, SURI SERI BEGAWAN HOSPITAL, DEPARTMENT OF LABORATORY SERVICES, MINISTRY OF HEALTH	05/08/2025	23/08/2025	\$5.00	JABATAN PERKHIDMATAN MAKMAL, KEMENTERIAN KESIHATAN

EQUIPMENT PURCHASE SPECIFICATION

QUOTATION /TENDER REFERENCE NO.	
TITLE OF QUOTATION / TENDER	TO SUPPLY, DELIVER, INSTALL AND COMMISSION ELECTRIC HYDRAULIC BODYLIFTER FOR MORTUARY, SURI SERI BEGAWAN HOSPITAL , DEPARTMENT OF LABORATORY SERVICES, MINISTRY OF HEALTH.

NO.	USER'S REQUIREMENT		VENDOR'S OFFER			
	ITEM DESCRIPTIONS AND SPECIFICATIONS	QUANTITY	ITEM DESCRIPTIONS AND SPECIFICATIONS	COMPLY / NOT COMPLY	COST PER UNIT (B\$)	TOTAL COSTS (B\$)
1	ELECTRIC HYDRAULIC BODY LIFTER <ul style="list-style-type: none"> The lifter is composed of base, lifting support and a body tray It is specially for autopsy table or body freezer The body tray is a concave shape with drainage, convenient for cleaning X-shaped lifting pushrod, smooth, steady, and noiseless The lifting controller is easy to operate Height adjustable range: 420 – 1300 mm The lifter can work 30-50 times when it is fully charged 4 silent universal castors, 2 with brake With electricity leakage protector and waterproof socket Dimensions (L x W x H): 220 x 65 x 42~130 cm (detachable body tray) Body tray size: 190x61 cm 	1 UNIT				
11	PRODUCT PRESENTATION AND/OR DEMONSTRATION					

NO.	USER'S REQUIREMENT		VENDOR'S OFFER			
	ITEM DESCRIPTIONS AND SPECIFICATIONS	QUANTITY	ITEM DESCRIPTIONS AND SPECIFICATIONS	COMPLY / NOT COMPLY	COST PER UNIT (B\$)	TOTAL COSTS (B\$)
	Vendor shall provide a presentation and/or demonstration on the offered product no later than two (2) weeks after the closing date of this advertisement.					
12	INTERNATIONAL MARKING The offered equipment shall be CE Marked or FDA approved or have equivalent international marking that is acceptable by the User.					
13	SAFETY STANDARDS The offered equipment shall be certified according to any international safety standards that is acceptable by the User. Please state the safety standards and submit a copy of the certificate.					
14	WARRANTY Vendor shall provide an extended warranty for the offered instruments up to eight (8) years of its life. This warranty shall cover all service works, technical support, periodic and preventive maintenance, spare parts and replacements.					
15	INSTALLATION AND COMMISSIONING					
15.1	Vendor shall verify the installation of offered equipment and submit to the User a copy of signed Delivery Order and a copy of user verified Installation Qualification Report.					
15.2	Vendor shall perform a performance verification upon commissioning and ensure the performance of the installed equipment is within the acceptable limit of performance or as per manufacturer's recommendations or as per User's acceptance criteria. Vendor shall submit a copy of user verified Performance Qualification Report.					
15.3	Electrical Safety – Vendor shall conduct electrical safety testing upon installation of purchased equipment(s) using calibrated device. Electrical safety testing report shall be submitted to the laboratory for acceptance.					
16	LITERATURE					

NO.	USER'S REQUIREMENT		VENDOR'S OFFER			
	ITEM DESCRIPTIONS AND SPECIFICATIONS	QUANTITY	ITEM DESCRIPTIONS AND SPECIFICATIONS	COMPLY / NOT COMPLY	COST PER UNIT (B\$)	TOTAL COSTS (B\$)
16.1	To supply one (1) CD or one (1) set of hard copy of the Operating Manual and Service Manual including circuit diagrams of the equipment shall be provided upon commissioning.					
16.2	To supply hardcopy of maintenance log with list of details of daily, weekly or scheduled maintenance					
17**	TRAINING Training shall be provided, at no additional cost, as follows:					
17.1	On-site training for ALL staff members expected to handle the machine. Please ensure that adequate time is allocated such that training will take place in small groups to minimize staff shortage in the laboratory.					
17.2	Certificate of attendance and competence shall be issued to all trainees after completion of training.					
18	DELIVERY PERIOD: Preferably 4 – 8 weeks and no later than 12 weeks after issue of Purchase Order		(Yes / No) (If No, please specify)			
19	PRICE VALIDITY: The quotation shall remain valid for 6 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).					

* 6 months validity required for <\$50K

No.	GENERAL SPECIFICATIONS	VENDOR'S OFFER
A	Model & Brand	
B	Country of Origin	
C	Total Price Per Test (CIF): B\$	
D	Price Ranking:	
E	Where marketed	
F	Year of Manufacture	
G	Warranty:	
H	Delivery Time:	
I	Power Requirements:	
J	Battery Back-up:	
K	International Safety Standard:	
L	Technical Support:	
M	Equipment Whole Life Support	
N	Dimensions (WxHxD) cm:	
O	Weight (kg):	
P	User Manuals	
Q	Service Manuals	
R	Spare-parts & Consumables Listing	
S	Technical Training On-Site:	
T	Site Requirements:	

*To all participating companies, please fill in the table above along with your other documents during submission of tender.

DELIVERY PERIOD AFTER PO ISSUED	4-8 weeks and no longer than 12 weeks		
Lab/Section/Unit	SSBH LABORATORY		
Lab/Section/Unit Ref No.:	DLS/PU/SSBH/2025/003		
Person to Contact	Name : SALMAH ZAHM		
	E-mail : salmah.zahm@moh.gov.bn		
	Tel.No. : 2325431 ext 4114	Fax No. : 2325100	
FOR ADMINISTRATION USE ONLY			
PPM/PROC Ref.No.	PPM/PROC/2025/<50K/007(SSBH)		
Advertisement Ref. No.		Date :	