

RUJ: BME/A2K/(1551)/2025/HTD (B8)**10/30/2025****TO REPAIR ONE (1) UNIT OF BABY INCUBATOR, BRAND: GE, MODEL: OMNIBED AT NICU WCC RIPAS HOSPITA****SCOPE OF WORK**

To repair one (1) unit of Baby Incubator, Brand: GE, Model: Omnibed at NICU WCC Ripas Hospital. Detail are as follow:

1. BME17952, Brand: GE, Model: Omnibed, SN: HDGU60011, at NICU, W/order 2025-5724

- Scope of work to follow equipment manual/recommendation specific for this system which include the following where if necessary:

(i) installation of items listed in spareparts list (Lampiran A2)

(ii) alignment

(iii) calibration

(iv) optimisation

(vi) any other related maintenance works required after installation of items listed in spareparts list (Lampiran A2)

- Perform functional test, verification and performance check post-maintenance.

- Disposal of any used spare parts, unless otherwise instructed.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipmen

RUJ: BME/A2K/(1551)/2025/HTD (B8)

TO REPAIR ONE (1) UNIT OF BABY INCUBATOR, BRAND: GE, MODEL: OMNIBED AT NICU WCC RIPAS HOSPITA

SPAREPART		PART NUMBER	QUANTITY
1	GIRAFFE EXAM LIGHT	2105361-001	1