

LAMPIRAN - A1

RUJ: BME/A2K/(1577)/2025/HTD (B4)

24/12/2025

TO REPAIR ONE (1) UNIT OF ENT WORKSTATION MODEL: S61 SERVANT, BME8513 AT
OTORHINOLARYNGOLOGY CLINIC, RIPAS HOSPITAL

SCOPE OF WORK

EQUIPMENT: ENT WORKSTATION
BRAND: ATMOS
MODEL NO. : S61 SERVANT
SERIAL NO. : 030038453-090959-2012-01
BME NO. : BME8513
LOCATION: OTORHINOLARYNGOLOGY CLINIC, RIPAS HOSPITAL

SCOPE OF WORKS:-

1. SPARE PART TO BE REPLACED
 - Heating Block 4 CPL S61 -1
 - CIRCUIT BOARD PAC-MOD -1
 - HEAT COND. FOIL ADHESIVE 1
 - HALOGEN LAMP 15V/150W 1
 - TRANSISTOR RET. SPRING F. T0220 1

LAMPIRAN - A2

RUJ: BME/A2K/(1577)/2025/HTD (B4)

**TO REPAIR ONE (1) UNIT OF ENT WORKSTATION MODEL: S61 SERVANT, BME8513 AT
OTORHINOLARYNGOLOGY CLINIC, RIPAS HOSPITAL**

	SPAREPART	PART NUMBER	QUANTITY
1	Heating Block 4 CPL S61	530.1166.0	1
2	CIRCUIT BOARD PAC-MOD	531.0102.0	1
3	HEAT COND. FOIL ADHESIVE	011.1079.0	1
4	TRANSISTOR RET. SPRING F. T0220	011.1289.0	1
5	HALOGEN LAMP 15V/150W	507.0905.0	10