

LAMPIRAN - A1

RUJ: BME/A2K/(1578)/2025/HTD (A16)

27/12/2025

TO CALIBRATE ONE (1) UNIT AUDIOMETRIC BOOTH AND TWO (2) UNITS AUDIOMETER AT OCCUPATIONAL HEALTH DIVISION, BERAKAS HEALTH SCREENING.

SCOPE OF WORK

To calibrate one (1) unit Audiometric Booth and two (2) units audiometer at Occupational Health Division, Berakas Health screening. The details of the equipment are as follows:

1. BME18727, Brand: Sibelmed, Model: S40, S/N: 017-D938 at Occupational Health Division Berakas Health Screening, W/order 2025-7662
2. BME20065, Brand: Interacoustics, Model: AD226, S/N: 0960589 at Occupational Health Division Berakas Health Screening, W/order 2025-7663
3. BME8088, Brand: Interacoustics, Model: AC40, S/N: 891038 at Occupational Health Division Berakas Health Screening, W/order 2025-7661

- Scope of work to follow maintenance manual/recommendation specific for models, which include the following where necessary:

- (i) inspection
- (ii) cleaning
- (iii) installation of PM kits, consumables and spare parts
- (iv) alignment
- (v) calibration
- (vi) any other related preventive maintenance works required

- Functional check post-maintenance.

- Apply preventive maintenance sticker/tag after successful service.

- Disposal of any used spare parts, unless otherwise instructed.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.

- Vendor(s) to submit maintenance checklist/details/scope specific for the equipment along with other necessary documents to join this tender.