

LAMPIRAN - A1

RUJ: BME/A2K/(1580)/2025/HTD (B4)

30/12/2025

TO REPAIR ONE (1) UNIT OF DIGITIZER FOR AGFA COMPUTED R SYSTEM (CR) SYSTEM, BMEXXXXX, RADIOLOGY DEPARTMENT, RIPAS HOSPITAL

SCOPE OF WORK

Repair one (1) unit of Digitizer for agfa computed r system (CR) machine, Brand: Agfa Healthcare NV, Model: DX-M, detail are as follow:

1. BME29943, SN; 13097, Workorder no. 2025-10445 at Radiology Department
 - TO REPLACE ONE (1) UNIT CPU EQUIVALENT TO THE EXISITING CPU AND COMPATIBLE WITH AGFA CR MACHINE
 - THE INSTALLATION OF HARDWARE AND SOFTWARE, ALONG WITH TESTING AND COMMISSIONING