

LAMPIRAN - A1

RUJ: BME/A2K/(1582)/2026/HTD (A20)

14/1/2026

REPAIR THREE (3) UNITS OF DYNAFLOW FUMEHOODS AT DEPARTMENT OF SCIENTIFIC SERVICES

SCOPE OF WORK

Repair services for the following three (3) units of Dynaflow fumehoods at Department of Scientific Services:

1. BME0002439, S/N: 15591 at Food Chemistry laboratory
2. BME0011676, S/N: 14005/06 at Environment laboratory
3. BME0001731, S/N: 14674 at Environment laboratory

- Scope of work to follow equipment manual/recommendation specific for this system which include the following where necessary:
 - (i) installation of items listed in spareparts list (Item 1-4 for BME0002349, Item 1, 2 & 7 for BME0011676 and Item 3, 5 & 6 for BME0001731 in Lampiran A2)
 - (i) alignment
 - (iii) calibration
 - (iv) optimisation
 - (vi) any other related maintenance works required after installation of items listed in spareparts list
- Functional check post-maintenance.
- Disposal of any used spare parts/pm kits, unless otherwise instructed.
- Work to be carried out by engineer(s) authorised in providing maintenance to the equipment.

LAMPIRAN - A2

RUJ: BME/A2K/(1582)/2026/HTD (A20)

REPAIR THREE (3) UNITS OF DYNAFLOW FUMEHOODS AT DEPARTMENT OF SCIENTIFIC SERVICES

	SPAREPART	PART NUMBER	QUANTITY
1	Control Module MK5		2
2	Variable Speed Drive 0.75kw		2
3	Lighting		2
4	Contactor for 3 pin outlet		1
5	Removable DURCON Epoxy Worktop with Sink cut-out and sink cover		1
6	3 pin socket		2
7	Circuit breaker		1