

ITEM(S) SPECIFICATIONS FOR ADVERTISEMENT (ABOVE \$2000)

QUOTATION/TENDER REFERENCE NO:	(02)IKLAN-QTN/UPP.HRIPAS/2026/ICU
QUOTATION/TENDER NAME	<i>SUPPLY AND DELIVERY MEDICAL ITEMS (DRESSING TROLLEY AND DRIP STAND) FOR INTENSIVE CARE UNIT AT RAJA ISTERI PENGIRAN ANAK SALEHA HOSPITAL (NON-CLUSTERING)</i>

USER'S REQUIREMENTS				VENDOR'S OFFER					
NO	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY OFFERED	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS (B\$)
1	<p><u>DRESSING TROLLEY (SIZE SMALL: L450mm X W450mm X H860mm)</u></p> <ul style="list-style-type: none"> • Fully stainless steel • 2 fixed shelves • Mobile on 4 x 75mm non-marking castor <ul style="list-style-type: none"> • No guardrail • No drawer • No handle <p>Technical Details:</p> <ul style="list-style-type: none"> • Overall dimension: Approx. L450 x W450 x H860 (mm) OR EQUIVALENT <p>* Please refer to the attachment for the picture of the item</p>	PER UNIT	3 UNITS						

USER'S REQUIREMENTS				VENDOR'S OFFER					
NO	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY OFFERED	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS (B\$)
2	<p><u>DRESSING TROLLEY (SIZE MEDIUM: L760mm X W450mm X H860mm)</u></p> <ul style="list-style-type: none"> • Fully stainless steel • 2 fixed shelves • Mobile on 4 x 75mm non-marking castor <ul style="list-style-type: none"> • No guardrail • No drawer • No handle <p>Technical Details:</p> <ul style="list-style-type: none"> • Overall dimension: Approx. L760 x W450 x H860 (mm) OR EQUIVALENT <p>* Please refer to the attachment for the picture of the item</p>	PER UNIT	3 UNITS						

USER'S REQUIREMENTS				VENDOR'S OFFER					
NO	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY OFFERED	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS (B\$)
3.	<p><u>DRESSING TROLLEY (SIZE LARGE: L900mm X W450mm X H860mm)</u></p> <ul style="list-style-type: none"> • Fully stainless steel • 2 fixed shelves • Mobile on 4 x 75mm non-marking castor <ul style="list-style-type: none"> • No guardrail • No drawer • No handle <p>Technical Details:</p> <ul style="list-style-type: none"> • Overall dimension: Approx. L900 x W450 x H860 (mm) OR EQUIVALENT <p>* Please refer to the attachment for the picture of the item</p>	PER UNIT	3 UNITS						

USER'S REQUIREMENTS				VENDOR'S OFFER					
NO	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY OFFERED	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS (B\$)
4.	<p><u>DRIP STAND</u></p> <p>1. Stability and Mobility</p> <ul style="list-style-type: none"> - Heavy, wide base (typically 5 legs) with high-quality, often anti-static and braked, castor wheels to prevent tipping when heavily loaded and to allow for safe patient transport. <p>2. Multiple Hooks</p> <ul style="list-style-type: none"> - Equipped with at least four robust hooks to support the concurrent administration of various fluids, medications, or blood products. <p>3. Durability and Hygiene</p> <ul style="list-style-type: none"> - Constructed from materials like high-quality stainless steel, easy to clean, resistant to corrosion, and meet strict infection control standards required in sterile environments. 	PER UNIT	20 UNITS						


USER'S REQUIREMENTS				VENDOR'S OFFER					
NO	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY USAGE	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY OFFERED	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS (B\$)
	<p>4. Adjustable Height</p> <ul style="list-style-type: none"> - Telescopic poles to allow adjustment of height precisely, ensuring proper gravity-assisted flow rates and ergonomic use for staff. <p>5. High Load Capacity</p> <ul style="list-style-type: none"> - Built with a robust total load capacity to securely hold multiple heavy items of equipment. <p>* Please refer to the attachment for the picture of the item</p>								
								TOTAL PRICE (B\$)	

NO	<u>TERMS AND CONDITIONS</u>	VENDOR'S OFFER (PLEASE STATE)
1	Tenderer must be registered with the Ministry of Health.	
2	QUOTATION/TENDER FORM should be filled completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form MAY cause DISQUALIFICATION OF QUOTATION/TENDER.	
3	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF QUOTATION/TENDER.	
4	All consumables supplied throughout this tender <u>shall</u> have a minimum expiry date of twelve (12) months / on delivery . Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made.	
5	Brochures / catalogues should be submitted / attached with quotation/tender document.	
5	Samples should be submitted together with quotation/ tender or within fourteen (14 days) of the quotation/tender closing date (if applicable).	
7	DELIVERY PERIOD: (Please state) Not later than 4 weeks	(Yes / No) (If No, please specify)
8	PRICE VALIDITY: The quotation shall remain valid for 12 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).	

Section/Unit		Section/Unit Ref No.:	
Person to Contact	Name :	Tel.No. :	
	E-mail :	Fax No.:	

FOR QUOTATION ONLY

TERMS AND CONDITIONS		
a.	Tenderer must be registered with the Ministry of Health	Company's Official Stamp
b.	Please fill in the QUOTATION FORM completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form MAY cause DISQUALIFICATION OF QUOTATION	
c.	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF QUOTATION	
d.	Please do not use TIPPEX for amendment	
		Acknowledgement: Company Ref. No.: I hereby certify the above quote to be correct. Signature: Name: Designation: Date:



ITEM NO.1:

DRESSING TROLLEY


(SIZE: L450mm X W450mm X H860mm)



ITEM NO.2:

DRESSING TROLLEY


(SIZE: L760mm X W450mm X H860mm)



ITEM NO.3:

DRESSING TROLLEY

(SIZE: L900mm X W450mm X H860mm)



ITEM NO.4 : DRIP STAND