

LAMPIRAN - A1

RUJ: BME/A2K/(1607)/2026/HTD (B8)

9/2/2026

TO REPAIR FOUR (4) UNITS OF BABY INCUBATOR, BRAND: GE, MODEL: OMNIBED AT WCC RIPAS HOSPITAL.

SCOPE OF WORK

To repair four (4) units of Baby Incubator, Brand: GE, Model: Omnibed at WCC Ripas Hospital. Detail are as follow:

1. BME17961, Brand: GE, Model: Omnibed, SN: HDGU60028, at NICU, W/order 2026-11026 (item 1)
2. BME16717, Brand: GE, Model: Omnibed, SN: HDHT50764, at SCBU, W/order 2026-10979 (item 2)
3. BME16720, Brand: GE, Model: Omnibed, SN: HDHT50767, at SCBU, W/order 2026-10967 (item 2)
4. BME10766, Brand: GE, Model: Omnibed, SN: HDHT50772, at SCBU, W/order 2026-10766 (item 2)

- Scope of work to follow equipment manual/recommendation specific for this system which include the following where if necessary:
 - (i) installation of items listed in spareparts list (Lampiran A2)
 - (ii) alignment
 - (iii) calibration
 - (iv) optimisation
 - (v) Top up gas
 - (vi) any other related maintenance works required after installation of items listed in spareparts list (Lampiran A2)
- Functional check post-maintenance.
- Disposal of any used spare parts, unless otherwise instructed.
- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.
- Vendor(s) to submit maintenance checklist/details/scope specific for the equipment along with other necessary documents to join this tender.

LAMPIRAN - A2

RUJ: BME/A2K/(1607)/2026/HTD (B8)

TO REPAIR FOUR (4) UNITS OF BABY INCUBATOR, BRAND: GE, MODEL: OMNIBED AT WCC RIPAS HOSPITAL.

	SPAREPART	PART NUMBER	QUANTITY
1	INNER WALL HYBRID GH	6600-1201-500	1
2	KIT POWER SUPPLY OMNIBED ASSEMBLY	2081522-001	3