

TENDER REFERENCE NO.: KK/06/2026/LAB(TC)

**MINISTRY OF HEALTH
NEGARA BRUNEI DARUSSALAM**

**TO SUPPLY AND DELIVER LABORATORY CONSUMABLES
& REAGENTS FOR NATIONAL HISTOLOGY AND
CYTOLOGY LABORATORY & MORTUARY SERVICES,
DEPARTMENT OF LABORATORY SERVICES, MINISTRY
OF HEALTH FOR A PERIOD OF THREE (3) YEARS USAGE**

TENDER FEES : \$30.00

RECEIPT NO. :

CLOSING DATE : ON TUESDAY, 03 MARCH 2026

TIME : 12.00 PM

FOA :

**THE CHAIRMAN
MINI TENDER BOARD, TENDER BOX
GROUND FLOOR, MINISTRY OF HEALTH
COMMONWEALTH DRIVE
BANDAR SERI BEGAWAN BB3910
NEGARA BRUNEI DARUSSALAM**

(CLUSTERING)

SECTION 2

SPECIFICATIONS AND REQUIREMENTS

TENDER REFERENCE NO.: KK/06/2026/LAB(TC)

**INVITATION TO TENDER
TO SUPPLY AND DELIVER LABORATORY CONSUMABLES & REAGENTS FOR NATIONAL
HISTOLOGY AND CYTOLOGY LABORATORY & MORTUARY SERVICES, DEPARTMENT OF
LABORATORY SERVICES, MINISTRY OF HEALTH FOR A PERIOD OF THREE (3) YEARS
USAGE**

DELIVERY PERIOD AFTER PURCHASE ORDER ISSUED	4-8 WEEKS AND NO LONGER THAN 12 WEEKS
--	--

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKAGING SIZE	TOTAL ESTIMATE USAGE / YEAR
1	<p>VFM COVERSLEIPS NO. 1.5 SIZE: 24 X 50MM (AUTOGLAS MICRO COVER GLASS)</p> <ul style="list-style-type: none"> - Size 24 x 50mm with thickness 0.13-0.16mm - Clear, pre-washed - Smooth surfaced with Uniform flatness and size - Tropical packaging with vacuum sealed, aluminium foil packed - Dust free, moisture and anti-fungal/ bacterial growth - Compatible with Automated Coverslippers 	100 PIECES /PACK OR EQUIVALENT	1,400 PACKS
2	<p>POLY-L-LYSINE (POLYSINE) PRE-COATED MICROSCOPE SLIDES SIZE: 25(W) X 75(L) X 1(D) MM</p> <ul style="list-style-type: none"> - Packing-72pcs/pack or equivalent - Premium Grade & pre-cleaned, ready to use - Ground edges, 90 degrees corners - White colour frosted one end markable with conventional labelling system, pencils, designated mark pens and compatible with 2-D barcode based slide printers - Dimensions: 25.00mm (W) x 75.0mm (L) x approx. 1.1mm thickness - hydrophobic surface with strong adhesion for tissue sections and body fluid 	72 PIECES/PACK OR EQUIVALENT	1,400 PACKS
3	<p>POSITIVE CHARGED MICROSCOPE SLIDES SIZE: 25(W) X 75(L) X 1(D) MM</p> <ul style="list-style-type: none"> - MUST BE compatible with ROCHE Benchmark Ultra and ROCHE Benchmark SS automated IHC and Special stainers - Packaging: 72 pcs per pack or equivalent - Color: White edge polishing - Size: Approx. 25mm x 75mm - Thickness: 1.0mm - Frosted one end, Frost Width: 15 x 26 mm - Ground edges, 90 degrees - Corner style: Traditional/ 90 degrees 	72 PIECES/ PACK OR EQUIVALENT	280 PACKS

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKAGING SIZE	TOTAL ESTIMATE USAGE / YEAR
4	<p>FORMALDEHYDE SOLUTION</p> <ul style="list-style-type: none"> - 37-41% w/v concentration - chemical formula CH₂O - pH 2.5-3.5 - Methanol content <5% - packaging 5-30 Litres per bottle - complete solubility in water - for use in mixing concentrated formalin solutions 	BETWEEN 5-30L PER BOTTLE	375 LITRES
5	<p>HARD CARDBOARD 6 ROWS STORAGE BOX</p> <ul style="list-style-type: none"> - cardboard storage box with cover - for storage of paraffin blocks generated from embedded patient's specimens. - Box outer dimensions: approx. 22 (W) x 46 (D) x 5 (H) cm. - enables systemic archiving and retrieval of patient's paraffin blocks - must be similarly compatible with existing storage boxes to ensure continuity of storage in the laboratory. 	PER SET OR EQUIVALENT	150 SETS
6	<p>DRAGER GAS DETECTION TUBES</p> <p>XYLENE 10/a</p> <p>Packing: 10 tubes per box Measuring range: 10 to 400ppm For use with Drager Hand Pump</p>	PER BOX OF 10 TESTS	1 BOX
7	<p>DRAGER GAS DETECTION TUBES</p> <p>FORMALDEHYDE 0.2/a</p> <p>Packing: 10 tubes per box Measuring range: 0.2 to 2.5ppm or 0.5-5ppm For use with Drager Hand Pump</p>	PER BOX OF 10 TUBES	1 BOX
8	<p>DRAGER GAS DETECTION TUBES</p> <p>ETHANOL 100/a</p> <p>Packing: 10 tubes per box Measuring range: 100 to 3000ppm For use with Drager Hand Pump</p>	PER BOX OF 10 TUBES	1 BOX
9	<p>CYTOLOGY FIXATIVE</p> <ul style="list-style-type: none"> - spray fixative for Cytological applications. - polyethylene glycol formulation for slide fixation. - enables rapid high quality fixation of cytological cell spreads. - protects cell morphology for microscopic examination. 	PER BOTTLE	6 BOTTLES
10	<p>VACUUM BLOOD COLLECTION TUBE</p> <ul style="list-style-type: none"> - vacuum plain glass tubes with cap - 10mL capacity, no additives 	PER RACK OF 100PCS / 12 RACKS PER CARTON	4,800 TUBES

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKAGING SIZE	TOTAL ESTIMATE USAGE / YEAR
	<ul style="list-style-type: none"> - tube size approximately 16 x 100mm, c/w red cap - 100pcs/rack, 12 racks/carton 		
11	<p>BLADE REMOVAL SYSTEM</p> <ul style="list-style-type: none"> - self contained unit compatible with most brands of general surgery blades and flat scalpel handles - able to remove, contain and dispose of scalpel blades with one-click system - able to remove up to 100 scalpel blades visible via blade counter onboard the system - OSHA compliant with single handed activation - automatic shut-off system upon reaching maximum capacity - puncture-proof sharps containment - bright colour for safe visibility 	24'S PER CARTON	1 CARTON
12	<p>MICROTOME TRIMMING KNIFE HANDLE</p> <ul style="list-style-type: none"> - to fit most high profile microtome blades - easy to clean - user-friendly operation 	PER UNIT/PC	1 UNIT/PC
13	<p>SCALPEL BLADE PM40</p> <ul style="list-style-type: none"> - to fit Swann-Morton screw-type autopsy blade handle - construction from carbon steel - individually wrapped, per box of 10 pieces 	10'S PC BOX	480 PCS
14	<p>SCALPEL BLADE No. 22</p> <ul style="list-style-type: none"> - surgical scalpel blade, disposable - to fit most surgical blade holders - curved cutting edge and flat unsharpened back edge - can fit type 4, 4L, 4 Graduated and 6B surgical blade handles - individually wrapped 100 pieces per box 	100 PCS PER BOX	12 BOXES
15	<p>SUTURE NEEDLE HALF CURVED FOR POST MORTEM</p> <ul style="list-style-type: none"> - high quality stainless steel needle for autopsy post mortem usage - durable and maintains sharpness and durability - for use with heavier post mortem thread - 4 inches - 12 per pack 	PER PACK OF 12'S	1 PACK
16	<p>SUTURE NEEDLE S-SHAPED FOR POST MORTEM</p> <ul style="list-style-type: none"> - high quality stainless steel needle for autopsy post mortem usage - durable and maintains sharpness and durability - for use with heavier post mortem thread 	PER PACK OF 12	33 PACKS

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKAGING SIZE	TOTAL ESTIMATE USAGE / YEAR
	<ul style="list-style-type: none"> - 140mm - 12 per pack 		
17	<p>EVIDENCE TAPE FOR AUTOPSY CASE SPECIMENS</p> <ul style="list-style-type: none"> - per roll approximate dimensions: 1.37 inch x 108 feet - red colour with tamper-evident construction and tamper-indicating properties - writable and option for package in dispenser box 	PER ROLL	100 ROLLS
18	<p>SUPER GLUE</p> <ul style="list-style-type: none"> - packaging of 3-5g per tube per pack - for total adhesion of mated surfaces under most environmental conditions 	PER TUBE	600 TUBES
19	<p>FLOOR CLEANER SOLUTION</p> <ul style="list-style-type: none"> - 3kg per bottle or equivalent to Ajax Fabuloso - kills up to 99% harmful bacteria and captures grease - can be used on marble, tiles, parquet, terrazzo, cement flooring 	PER BOTTLE/LITTER	100 BOTTLES
20	<p>DISHWASHING LIQUID</p> <ul style="list-style-type: none"> - 5L per bottle or equivalent - bactericide and quick grease dissolving properties - capable of cleaning grease from plastic and metal surfaces 	5L PER BOTTLE	20 BOTTLES
21	<p>LEO WIPE STAIN REMOVER</p> <ul style="list-style-type: none"> - 2L per bottle or equivalent - capable of removing formation of scale, algae, mould, urine and water marks from mosaic flooring and drains - chemical formulation requires minimal scrubbing or agitation 	2L PER BOTTLE	10 BOTTLES
22	<p>BODY SHOWER CREAM</p> <ul style="list-style-type: none"> - 2L per bottle or equivalent - anti-bacterial properties - moisturizing properties, non-drying to skin - neutral scent profile preferred 	2L PER BOTTLE	15 BOTTLES
23	<p>Rapid Decalcifier TBD-1</p> <ul style="list-style-type: none"> - used for rapid decalcification of bony tissue - permits processing of bone marrow biopsies as well as thin pieces of dense bone - additives protect cells from action of the decalcifying acids 	PER BOTTLE	2 BOTTLES

NO.	SPECIFICATIONS AND REQUIREMENTS
1	All reagent test kits / consumables supplied throughout this tender <u>shall</u> have a minimum expiry date of twelve (12) months on delivery. Should the reagent or consumable be urgently needed, provision of a reagent test kit or consumable with expiry date of less than twelve (12) months should be first agreed by the User of the particular laboratory before delivery is made.
2	Letter of Undertaking (LOU) shall be produced upon each delivery of test kit or consumable with expiry date of less than twelve (12) months and vendor shall declare in the LOU that unused, unopened, expired kits will be replaced accordingly. For items which are known to have short expiry date such as those containing red blood cells, list down all such items and vendor shall declare in this tender submission of such items and shall be exempted from submitting LOU upon delivery.
3	Staggered delivery every 3 months period directly to the User.
4	Product inserts and MSDS (if applicable) shall be provided to the User.
5	International markings: products must be CE marked or equivalent standards that are acceptable to the users.
6	User shall have the rights to refuse delivery of items that do not meet the acceptance criteria such as, but not limited to, the following: 1. Tampered or damaged box 2. Leakage upon delivery 3. Items stored pre-delivery not in accordance to manufacturer's instructions 4. Expiry date not meeting requirement
7	User shall have the rights to return any items, and to be replaced at no extra cost, if found not meeting the acceptance criteria upon opening a pack such as, but not limited to, the following: 1. Tampered or damaged packaging 2. Evident of leakage or damaged products 3. Expired products that are evidently less than the requirement mentioned in para 1 calculated from delivery date 4. Leakage upon delivery
8	Vendor shall submit samples of the offered items directly to the Users no later than 7 days after the Closing Date of this advertisement or as required by the Users.
9	FINANCIAL AGREEMENT
9.1	Supply of the test kit including reagents, consumables and/or accessories is based on the number of kits required in the Purchase Order according to an agreed schedule period as stated in para 2.
9.2	Buffer stock of the test kit including reagents, consumables and accessories shall be available at the local representative as contingency.
9.3	Should there be any discontinuity of reagents / consumables due to non-compliance in the manufacturing of reagents; the vendor must be able to provide an alternative so that the test requests / services are still available for the customers.
10	EXIT CLAUSE: The tender contract shall be automatically terminated even though tender has not yet expired and this shall be in effect due to, but not limited to, the following: 1. When the testing is no longer required or relevant i.e. test is obsolete, to the laboratory or the Department. 2. When the item(s) set out in this tender is/are no longer required by the laboratory or the Department. 3. When the approved budget allocation for this tender contract has been used up before the tender contract expires whereby a renewal of tender shall be submitted by the user for an open advertisement subject to approval by the Mini Tender Board (<i>Lembaga Tawaran Kecil</i>).
11	DELIVERY PERIOD: Preferably 4 – 8 weeks and no later than 12 weeks after issue of Purchase Order
12	PRICE VALIDITY: The quotation shall remain valid for 12 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).

* 6 months validity required for <\$50K or 12 months for >\$50K

SECTION 3
FORMS TO BE USED
CONTENTS

SCHEDULE 1 - TENDER FORM

SCHEDULE 2 - INFORMATION SUMMARY

SCHEDULE 3 - SUB-CONTRACTS

SCHEDULE 4 - COMPANY BACKGROUND

SCHEDULE 5 - REFERENCES

SCHEDULE 6 - LETTER OF DECLARATION

USER'S REQUIREMENTS				VENDOR'S OFFER					
NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKAGING SIZE	TOTAL ESTIMATE USAGE / YEAR	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKAGING SIZE	TOTAL QUANTITY OFFERED / YEAR	*COST PER UNIT (B\$)	TOTAL COSTS (B\$)
2	<p>POLY-L-LYSINE (POLYSINE) PRE-COATED MICROSCOPE SLIDES SIZE: 25(W) X 75(L) X 1(D) MM</p> <ul style="list-style-type: none"> - Packing-72pcs/pack or equivalent - Premium Grade & pre-cleaned, ready to use - Ground edges, 90 degrees corners - White colour frosted one end markable with conventional labelling system, pencils, designated mark pens and compatible with 2-D barcode based slide printers - Dimensions: 25.00mm (W) x 75.0mm (L) x approx. 1.1mm thickness - hydrophobic surface with strong adhesion for tissue sections and body fluid 	72 PIECES/PACK OR EQUIVALENT	1,400 PACKS						
3	<p>POSITIVE CHARGED MICROSCOPE SLIDES SIZE: 25(W) X 75(L) X 1(D) MM</p> <ul style="list-style-type: none"> - MUST BE compatible with ROCHE Benchmark Ultra and ROCHE Benchmark SS automated IHC and Special stainers - Packaging: 72 pcs per pack or equivalent - Color: White edge polishing - Size: Approx. 25mm x 75mm - Thickness: 1.0mm 	72 PIECES/ PACK OR EQUIVALENT	280 PACKS						

USER'S REQUIREMENTS				VENDOR'S OFFER					
NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKAGING SIZE	TOTAL ESTIMATE USAGE / YEAR	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKAGING SIZE	TOTAL QUANTITY OFFERED / YEAR	*COST PER UNIT (B\$)	TOTAL COSTS (B\$)
	<ul style="list-style-type: none"> - Frosted one end, Frost Width: 15 x 26 mm - Ground edges, 90 degrees - Corner style: Traditional/ 90 degrees 								
4	<p>FORMALDEHYDE SOLUTION</p> <ul style="list-style-type: none"> - 37-41% w/v concentration - chemical formula CH₂O - pH 2.5-3.5 - Methanol content <5% - packaging 5-30 Litres per bottle - complete solubility in water - for use in mixing concentrated formalin solutions 	BETWEEN 5-30L PER BOTTLE	375 LITRES						
5	<p>HARD CARDBOARD 6 ROWS STORAGE BOX</p> <ul style="list-style-type: none"> - cardboard storage box with cover - for storage of paraffin blocks generated from embedded patient's specimens. - Box outer dimensions: approx. 22 (W) x 46 (D) x 5 (H) cm. - enables systemic archiving and retrieval of patient's paraffin blocks - must be similarly compatible with existing storage boxes to ensure continuity of storage in the laboratory. 	PER SET OR EQUIVALENT	150 SETS						
6	<p>DRAGER GAS DETECTION TUBES</p> <p>XYLENE 10/a</p>	PER BOX OF 10 TESTS	1 BOX						

USER'S REQUIREMENTS				VENDOR'S OFFER					
NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKAGING SIZE	TOTAL ESTIMATE USAGE / YEAR	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKAGING SIZE	TOTAL QUANTITY OFFERED / YEAR	*COST PER UNIT (B\$)	TOTAL COSTS (B\$)
	Packing: 10 tubes per box Measuring range: 10 to 400ppm For use with Drager Hand Pump								
7	DRAGER GAS DETECTION TUBES FORMALDEHYDE 0.2/a Packing: 10 tubes per box Measuring range: 0.2 to 2.5ppm or 0.5-5ppm For use with Drager Hand Pump	PER BOX OF 10 TUBES	1 BOX						
8	DRAGER GAS DETECTION TUBES ETHANOL 100/a Packing: 10 tubes per box Measuring range: 100 to 3000ppm For use with Drager Hand Pump	PER BOX OF 10 TUBES	1 BOX						
9	CYTOLOGY FIXATIVE - spray fixative for Cytological applications. - polyethylene glycol formulation for slide fixation. - enables rapid high quality fixation of cytological cell spreads. - protects cell morphology for microscopic examination.	PER BOTTLE	6 BOTTLES						
10	VACUUM BLOOD COLLECTION TUBE - vacuum plain glass tubes with cap - 10mL capacity, no additives	PER RACK OF 100PCS / 12 RACKS PER CARTON	4,800 TUBES						

USER'S REQUIREMENTS				VENDOR'S OFFER					
NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKAGING SIZE	TOTAL ESTIMATE USAGE / YEAR	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKAGING SIZE	TOTAL QUANTITY OFFERED / YEAR	*COST PER UNIT (B\$)	TOTAL COSTS (B\$)
	<ul style="list-style-type: none"> - tube size approximately 16 x 100mm, c/w red cap - 100pcs/rack, 12 racks/carton 								
11	<p>BLADE REMOVAL SYSTEM</p> <ul style="list-style-type: none"> - self contained unit compatible with most brands of general surgery blades and flat scalpel handles - able to remove, contain and dispose of scalpel blades with one-click system - able to remove up to 100 scalpel blades visible via blade counter onboard the system - OHSA compliant with single handed activation - automatic shut-off system upon reaching maximum capacity - puncture-proof sharps containment - bright colour for safe visibility 	24'S PER CARTON	1 CARTON						
12	<p>MICROTOME TRIMMING KNIFE HANDLE</p> <ul style="list-style-type: none"> - to fit most high profile microtome blades - easy to clean - user-friendly operation 	PER UNIT/PC	1 UNIT/PC						
13	<p>SCALPEL BLADE PM40</p> <ul style="list-style-type: none"> - to fit Swann-Morton screw-type autopsy blade handle - construction from carbon steel - individually wrapped, per box of 10 pieces 	10'S PC BOX	480 PCS						

USER'S REQUIREMENTS				VENDOR'S OFFER					
NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKAGING SIZE	TOTAL ESTIMATE USAGE / YEAR	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKAGING SIZE	TOTAL QUANTITY OFFERED / YEAR	*COST PER UNIT (B\$)	TOTAL COSTS (B\$)
14	SCALPEL BLADE No. 22 - surgical scalpel blade, disposable to fit most surgical blade holders - curved cutting edge and flat unsharpened back edge - can fit type 4, 4L, 4 Graduated and 6B surgical blade handles - individually wrapped 100 pieces per box	100 PCS PER BOX	12 BOXES						
15	SUTURE NEEDLE HALF CURVED FOR POST MORTEM - high quality stainless steel needle for autopsy post mortem usage - durable and maintains sharpness and durability - for use with heavier post mortem thread - 4 inches 12 per pack	PER PACK OF 12'S	1 PACK						
16	SUTURE NEEDLE S-SHAPED FOR POST MORTEM - high quality stainless steel needle for autopsy post mortem usage - durable and maintains sharpness and durability - for use with heavier post mortem thread - 140mm - 12 per pack	PER PACK OF 12	33 PACKS						
17	EVIDENCE TAPE FOR AUTOPSY CASE SPECIMENS	PER ROLL	100 ROLLS						

USER'S REQUIREMENTS				VENDOR'S OFFER					
NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKAGING SIZE	TOTAL ESTIMATE USAGE / YEAR	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKAGING SIZE	TOTAL QUANTITY OFFERED / YEAR	*COST PER UNIT (B\$)	TOTAL COSTS (B\$)
	<ul style="list-style-type: none"> - per roll approximate dimensions: 1.37 inch x 108 feet - red colour with tamper-evident construction and tamper-indicating properties - writable and option for package in dispenser box 								
18	<p>SUPER GLUE</p> <ul style="list-style-type: none"> - packaging of 3-5g per tube per pack - for total adhesion of mated surfaces under most environmental conditions 	PER TUBE	600 TUBES						
19	<p>FLOOR CLEANER SOLUTION</p> <ul style="list-style-type: none"> - 3kg per bottle or equivalent to Ajax Fabuloso - kills up to 99% harmful bacteria and captures grease - can be used on marble, tiles, parquet, terrazzo, cement flooring 	PER BOTTLE/LITTER	100 BOTTLES						
20	<p>DISHWASHING LIQUID</p> <ul style="list-style-type: none"> - 5L per bottle or equivalent - bactericide and quick grease dissolving properties - capable of cleaning grease from plastic and metal surfaces 	5L PER BOTTLE	20 BOTTLES						
21	<p>LEO WIPE STAIN REMOVER</p> <ul style="list-style-type: none"> - 2L per bottle or equivalent 	2L PER BOTTLE	10 BOTTLES						

USER'S REQUIREMENTS				VENDOR'S OFFER					
NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	PACKAGING SIZE	TOTAL ESTIMATE USAGE / YEAR	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/ CATALOGUE NUMBER AND BRAND	PACKAGING SIZE	TOTAL QUANTITY OFFERED / YEAR	*COST PER UNIT (B\$)	TOTAL COSTS (B\$)
	<ul style="list-style-type: none"> - capable of removing formation of scale, algae, mould, urine and water marks from mosaic flooring and drains - chemical formulation requires minimal scrubbing or agitation 								
22	BODY SHOWER CREAM <ul style="list-style-type: none"> - 2L per bottle or equivalent - anti-bacterial properties - moisturizing properties, non-drying to skin - neutral scent profile preferred 	2L PER BOTTLE	15 BOTTLES						
23	Rapid Decalcifier TBD-1 <ul style="list-style-type: none"> - used for rapid decalcification of bony tissue - permits processing of bone marrow biopsies as well as thin pieces of dense bon - additives protect cells from action of the decalcifying acids 	PER BOTTLE	2 BOTTLES						
TOTAL PRICE PER YEAR (B\$)									
TOTAL PRICE FOR THREE (3) YEARS (B\$)									

*Cost per test should include the kit, control, calibrator and accessories/consumables required to run the test.

NO.	SPECIFICATIONS AND REQUIREMENTS	VENDOR'S OFFER (PLEASE STATE)
1	All reagent test kits / consumables supplied throughout this tender <u>shall</u> have a minimum expiry date of twelve (12) months on delivery. Should the reagent or consumable be urgently needed, provision of a reagent test kit or consumable with expiry date of less than twelve (12) months should be first agreed by the User of the particular laboratory before delivery is made.	
2	Letter of Undertaking (LOU) shall be produced upon each delivery of test kit or consumable with expiry date of less than twelve (12) months and vendor shall declare in the LOU that unused, unopened, expired kits will be replaced accordingly. For items which are known to have short expiry date such as those containing red blood cells, list down all such items and vendor shall declare in this tender submission of such items and shall be exempted from submitting LOU upon delivery.	
3	Staggered delivery every 3 months period directly to the User.	
4	Product inserts and MSDS (if applicable) shall be provided to the User.	
5	International markings: products must be CE marked or equivalent standards that are acceptable to the users.	
6	User shall have the rights to refuse delivery of items that do not meet the acceptance criteria such as, but not limited to, the following: 1. Tampered or damaged box 2. Leakage upon delivery 3. Items stored pre-delivery not in accordance to manufacturer's instructions 4. Expiry date not meeting requirement	
7	User shall have the rights to return any items, and to be replaced at no extra cost, if found not meeting the acceptance criteria upon opening a pack such as, but not limited to, the following: 1. Tampered or damaged packaging 2. Evident of leakage or damaged products 3. Expired products that are evidently less than the requirement mentioned in para 1 calculated from delivery date 4. Leakage upon delivery	
8	Vendor shall submit samples of the offered items directly to the Users no later than 7 days after the Closing Date of this advertisement or as required by the Users.	

NO.	SPECIFICATIONS AND REQUIREMENTS	VENDOR'S OFFER (PLEASE STATE)
9	FINANCIAL AGREEMENT	
9.1	Supply of the test kit including reagents, consumables and/or accessories is based on the number of kits required in the Purchase Order according to an agreed schedule period as stated in para 2.	
9.2	Buffer stock of the test kit including reagents, consumables and accessories shall be available at the local representative as contingency.	
9.3	Should there be any discontinuity of reagents / consumables due to non-compliance in the manufacturing of reagents; the vendor must be able to provide an alternative so that the test requests / services are still available for the customers.	
10	<p>EXIT CLAUSE: The tender contract shall be automatically terminated even though tender has not yet expired and this shall be in effect due to, but not limited to, the following:</p> <ol style="list-style-type: none"> 1. When the testing is no longer required or relevant i.e. test is obsolete, to the laboratory or the Department. 2. When the item(s) set out in this tender is/are no longer required by the laboratory or the Department. 3. When the approved budget allocation for this tender contract has been used up before the tender contract expires whereby a renewal of tender shall be submitted by the user for an open advertisement subject to approval by the Mini Tender Board (<i>Lembaga Tawaran Kecil</i>). 	
11	<p>DELIVERY PERIOD: Preferably 4 – 8 weeks and no later than 12 weeks after issue of Purchase Order</p>	
12	<p>PRICE VALIDITY: The quotation shall remain valid for 12 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).</p>	

1. We offer and undertake on your acceptance of our Tender to supply and deliver the above mentioned goods in accordance with your Invitation To Tender.
2. Our Tender is fully consistent with and does not contradict or derogate from anything in your Invitation To Tender. We have not qualified or changed any of the provisions of your Invitation To Tender.
3. We shall execute a formal agreement in the appropriate form set out in Section 4 – Contract of the Invitation to Tender together with such further terms and conditions, if any, agreed between the Government and us.
4. OUR OFFER IS VALID FOR **TWELVE (12)** CALENDER MONTHS FROM THE TENDER CLOSING DATE.
5. When requested by you, we shall extend the validity of this offer.
6. We further undertake to give you any further information which you may require.

Dated this _____ day of _____, 20_____

[Signature of authorised officer of Tenderer]

Name:

Designation:

Tenderer's official stamp:

SCHEDULE 2 - INFORMATION SUMMARY

2.1 Tenderers shall provide in this Schedule the following information:

- a. Management summary
- b. Company profile (including Contractor and sub-contractor(s), if any)
- c. Years of experience (as of the Tender Closing Date) of the Contractor and sub-contractor(s) in the:
 - *Supply & Delivery Of Laboratory Equipment, Test Kits and Consumables.*
- d. Other information which is considered relevant

SCHEDULE 3 – SUB-CONTRACTS

- 3.1 Tenderers shall complete Table 3.1 with information about all the companies involved in the provision of the services and items specified in this tender. This shall include details about the Contractor and each sub-contractor involved, as well as their respective responsibilities.
- 3.2 Tenderers shall also indicate in Table 3.1 any alliance relationship established with each sub-contractor. An alliance is defined as a formal and binding business relationship between the allied parties.

Table 3.1 - Responsibility Table

Company Name	Responsibility Description	Alliance Relationship between Contractor and Sub-contractor(s)		
		Alliance Exists? (Y/N)	Date Established	Alliance Description
Contractor				
		Not Applicable	Not Applicable	Not Applicable
Sub-contractor(s)				

SCHEDULE 4 – COMPANY’S BACKGROUND

- 4.1 Each of the companies involved in this tender, including Contractor and sub-contractor(s) (if any), shall provide information on the company’s background, scope of operations, financial standing and certified copy of its Certificate of Incorporation or Certificate of Registration (as the case may be).

SCHEDULE 5 – REFERENCES

- 5.1 Tenderers shall submit a list of customers in Table 5.1 to whom the Contractor has provided similar services and items as specified in this tender in the recent 5 years as of the Tender Closing Date.

Table 5.1 - References of previous customers

Customer Name and Address	Customer Type (Govt or Quasi Govt)*	Contact Person	Title	Contact Number, Fax Number and E-mail Address

***Note: Tenderers shall indicate whether the customer is a Government or Quasi Government organisation. A Quasi Government is defined as an organisation which (1) is managed and controlled by the Government; or (2) has at least 50% shares being held by the Government. Please leave the column blank if the customer is neither a Government or Quasi Government organisation.**

- 5.2 The Ministry of Health shall treat all the information submitted under this schedule in strict confidence.
- 5.3 The Ministry of Health reserves the right to contact the references for tender assessment purposes.

SCHEDULE 6 - SUBMISSION OF SAMPLE

- 6.1 Tenderers shall submit the Submission of Sample form below in respect of the items specified in this tender.
- 6.2 Samples of the items to be submitted shall be:
- a) identical in packing and manufacture to the items to be offered by the Tenderer; and
 - b) marked with the corresponding item number of the tender.

NO.	TEST/REAGENT NAME	SAMPLE SUBMITTED (indicate with ✓)	SAMPLE NOT SUBMITTED (indicate with X)	OFFERED/ NOT OFFERED (indicate as appropriate)
3	<p>POSITIVE CHARGED MICROSCOPE SLIDES SIZE: 25(W) X 75(L) X 1(D) MM</p> <ul style="list-style-type: none"> - MUST BE compatible with ROCHE Benchmark Ultra and ROCHE Benchmark SS automated IHC and Special stainers - Packaging: 72 pcs per pack or equivalent - Color: White edge polishing - Size: Approx. 25mm x 75mm - Thickness: 1.0mm - Frosted one end, Frost Width: 15 x 26 mm - Ground edges, 90 degrees - Corner style: Traditional/ 90 degrees 			
4	<p>FORMALDEHYDE SOLUTION</p> <ul style="list-style-type: none"> - 37-41% w/v concentration - chemical formula CH₂O - pH 2.5-3.5 - Methanol content <5% - packaging 5-30 Litres per bottle - complete solubility in water - for use in mixing concentrated formalin solutions 			
5	<p>HARD CARDBOARD 6 ROWS STORAGE BOX</p> <ul style="list-style-type: none"> - cardboard storage box with cover - for storage of paraffin blocks generated from embedded patient's specimens. - Box outer dimensions: approx. 22 (W) x 46 (D) x 5 (H) cm. - enables systemic archiving and retrieval of patient's paraffin blocks - must be similarly compatible with existing storage boxes to ensure continuity of storage in the laboratory. 			
6	<p>DRAGER GAS DETECTION TUBES</p> <p>XYLENE 10/a</p> <p>Packing: 10 tubes per box Measuring range: 10 to 400ppm</p> <ul style="list-style-type: none"> - For use with Drager Hand Pump 			

NO.	TEST/REAGENT NAME	SAMPLE SUBMITTED (indicate with ✓)	SAMPLE NOT SUBMITTED (indicate with X)	OFFERED/ NOT OFFERED (indicate as appropriate)
7	DRAGER GAS DETECTION TUBES FORMALDEHYDE 0.2/a Packing: 10 tubes per box Measuring range: 0.2 to 2.5ppm or 0.5-5ppm - For use with Drager Hand Pump			
8	DRAGER GAS DETECTION TUBES ETHANOL 100/a Packing: 10 tubes per box Measuring range: 100 to 3000ppm - For use with Drager Hand Pump			
9	CYTOLOGY FIXATIVE <ul style="list-style-type: none"> - spray fixative for Cytological applications. - polyethylene glycol formulation for slide fixation. - enables rapid high quality fixation of cytological cell spreads. - protects cell morphology for microscopic examination. 			
10	VACUUM BLOOD COLLECTION TUBE <ul style="list-style-type: none"> - vacuum plain glass tubes with cap - 10mL capacity, no additives - tube size approximately 16 x 100mm, c/w red cap - 100pcs/rack, 12 racks/carton 			
11	BLADE REMOVAL SYSTEM <ul style="list-style-type: none"> - self contained unit compatible with most brands of general surgery blades and flat scalpel handles - able to remove, contain and dispose of scalpel blades with one-click system - able to remove up to 100 scalpel blades visible via blade counter onboard the system - OHSA compliant with single handed activation - automatic shut-off system upon reaching maximum capacity - puncture-proof sharps containment - bright colour for safe visibility 			
12	MICROTOME TRIMMING KNIFE HANDLE <ul style="list-style-type: none"> - to fit most high profile microtome blades - easy to clean - user-friendly operation 			

NO.	TEST/REAGENT NAME	SAMPLE SUBMITTED (indicate with ✓)	SAMPLE NOT SUBMITTED (indicate with X)	OFFERED/ NOT OFFERED (indicate as appropriate)
13	SCALPEL BLADE PM40 <ul style="list-style-type: none"> - to fit Swann-Morton screw-type autopsy blade handle - construction from carbon steel - individually wrapped, per box of 10 pieces 			
14	SCALPEL BLADE No. 22 <ul style="list-style-type: none"> - surgical scalpel blade, disposable - to fit most surgical blade holders - curved cutting edge and flat unsharpened back edge - can fit type 4, 4L, 4 Graduated and 6B surgical blade handles - individually wrapped 100 pieces per box 			
15	SUTURE NEEDLE HALF CURVED FOR POST MORTEM <ul style="list-style-type: none"> - high quality stainless steel needle for autopsy post mortem usage - durable and maintains sharpness and durability - for use with heavier post mortem thread - 4 inches - 12 per pack 			
16	SUTURE NEEDLE S-SHAPED FOR POST MORTEM <ul style="list-style-type: none"> - high quality stainless steel needle for autopsy post mortem usage - durable and maintains sharpness and durability - for use with heavier post mortem thread - 140mm - 12 per pack 			
17	EVIDENCE TAPE FOR AUTOPSY CASE SPECIMENS <ul style="list-style-type: none"> - per roll approximate dimensions: 1.37 inch x 108 feet - red colour with tamper-evident construction and tamper-indicating properties - writable and option for package in dispenser box 			
18	SUPER GLUE <ul style="list-style-type: none"> - packaging of 3-5g per tube per pack - for total adhesion of mated surfaces under most environmental conditions 			

NO.	TEST/REAGENT NAME	SAMPLE SUBMITTED (indicate with ✓)	SAMPLE NOT SUBMITTED (indicate with X)	OFFERED/ NOT OFFERED (indicate as appropriate)
19	FLOOR CLEANER SOLUTION <ul style="list-style-type: none"> - 3kg per bottle or equivalent to Ajax Fabuloso - kills up to 99% harmful bacteria and captures grease - can be used on marble, tiles, parquet, terrazzo, cement flooring 			
20	DISHWASHING LIQUID <ul style="list-style-type: none"> - 5L per bottle or equivalent - bactericide and quick grease dissolving properties - capable of cleaning grease from plastic and metal surfaces 			
21	LEO WIPE STAIN REMOVER <ul style="list-style-type: none"> - 2L per bottle or equivalent - capable of removing formation of scale, algae, mould, urine and water marks from mosaic flooring and drains - chemical formulation requires minimal scrubbing or agitation 			
22	BODY SHOWER CREAM <ul style="list-style-type: none"> - 2L per bottle or equivalent - anti-bacterial properties - moisturizing properties, non-drying to skin - neutral scent profile preferred 			
23	Rapid Decalcifier TBD-1 <ul style="list-style-type: none"> - used for rapid decalcification of bony tissue - permits processing of bone marrow biopsies as well as thin pieces of dense bone - additives protect cells from action of the decalcifying acids 			

We understand as stated in the Instructions to Tenderers that Tenders without samples shall not be considered.

Tenderer's official stamp:

[signature of authorized officer of Tenderer]

Name:

Designation:

Date:

FOR OFFICE USE

Date of receipt : _____

Receiving Officer : _____