

**TENDER REFERENCE NO.: KK/22/2026/UPP(TC)**

**MINISTRY OF HEALTH  
NEGARA BRUNEI DARUSSALAM**

**SUPPLY AND DELIVERY OF ENDOSCOPIC  
CONSUMABLES FOR GASTROENTEROLOGY &  
HEPATOLOGY UNIT, RAJA ISTERI PENGIRAN ANAK  
SALEHA (RIPAS) HOSPITAL FOR A PERIOD OF THREE (3)  
YEARS**

**TENDER FEES : \$30.00**

**RECEIPT NO. :**

**CLOSING DATE : ON TUESDAY, 10/03/2026**

**TIME : 12.00 PM**

**FOA :**

**THE CHAIRMAN  
MINI TENDER BOARD, TENDER BOX  
GROUND FLOOR, MINISTRY OF HEALTH  
COMMONWEALTH DRIVE  
BANDAR SERI BEGAWAN BB3910  
NEGARA BRUNEI DARUSSALAM**

**(NON-CLUSTERING)**

**SECTION 2**

**SPECIFICATIONS**

**TENDER REFERENCE NO.: KK/22/2026/UPP(TC)**

**INVITATION TO TENDER  
THE SUPPLY AND DELIVERY ENDOSCOPIC CONSUMABLES FOR THE  
GASTROENTEROLOGY AND HEPATOLOGY UNIT, RAJA ISTERI PENGIRAN ANAK SALEHA  
(RIPAS) HOSPITAL, FOR A PERIOD OF THREE (3) YEARS**

|                        |   |
|------------------------|---|
| <b>DELIVERY PERIOD</b> | <b>NOT LATER THAN 4 WEEKS<br/>STAGGERED DELIVERY UPON REQUEST</b> |
|------------------------|---|

| NO. | ITEM DESCRIPTIONS AND SPECIFICATIONS   | PACKING SIZE | TOTAL QUANTITY USAGE PER YEAR |
|-----|--|--------------|-------------------------------|
| 1   | <b>HERCULES 3 STAGE WIRE GUIDED BALLOON</b><br>▪ P/N: HBD-W-8-9-10<br>Catheter 8Fr, catheter length 240cm, balloon length cm 5.5   | UNIT         | 4                             |
| 2   | <b>HERCULES 3 STAGE WIRE GUIDED BALLOON</b><br>▪ P/N: HBD-W-10-11-12<br>Catheter 8Fr, catheter length 240cm, balloon length 5.5cm  | UNIT         | 4                             |
| 3   | <b>HERCULES 3 STAGE WIRE GUIDED BALLOON</b><br>▪ P/N: HBD-W-12-13.5-15<br>Catheter 8Fr, catheter length 240cm, balloon length 5.5cm                                      | UNIT         | 4                             |
| 4   | <b>HERCULES 3 STAGE WIRE GUIDED BALLOON</b><br>▪ P/N: HBD-W-15-16.5-18<br>Catheter 8Fr, catheter length 240cm, balloon diameter mm 15-16.5-18, balloon length 5.5 cm     | UNIT         | 4                             |
| 5   | <b>HERCULES 3 STAGE WIRE GUIDED BALLOON</b><br>▪ P/N: HBD-W-18-19-20<br>▪ Catheter 8Fr, Catheter length 240cm, balloon 18-19-20 balloon length 5.5cm                     | UNIT         | 4                             |
| 6   | <b>ACHALASIA 8CM BALLOON CATHETER</b><br>▪ 16Fr- 75cm<br>▪ Balloon Diameter 30mm, 8cm Length<br>Balloon Diameter 90 Fr   | UNIT         | 8                             |
| 7   | <b>ACHALASIA 8CM BALLOON CATHETER</b><br>▪ 16fr- 75cm, Balloon Diameter 35 mm, 8cm Length<br>Balloon Diameter 105 Fr   | UNIT         | 8                             |
| 8   | <b>CYTOMAX II DOUBLE LUMEN CYTOLOGY</b><br>▪ Catheter 8Fr, 200cm<br>▪ Length Wire Guide Diameter 0.035"<br>▪ Flipform Tipsize 1.5cm/3.5cm<br>Product code: G22674/G22675 | UNIT         | 72                            |
| 9   | <b>CRE PRO WIRE GUIDED BALLOON DILATATION CATHETER</b><br>▪ Esophageal/Pyloric/Colonic/Biliary<br>Inflated OD 8-10mm   | UNIT         | 10                            |
| 10  | <b>CRE PRO WIRE GUIDED BALLOON DILATATION CATHETER</b><br>▪ Esophageal/Pyloric/Colonic/Biliary<br>Inflated OD 10-12mm  | UNIT         | 10                            |

| NO. | ITEM DESCRIPTIONS AND SPECIFICATIONS  | PACKING SIZE          | TOTAL QUANTITY USAGE PER YEAR |
|-----|---|-----------------------|-------------------------------|
| 11  | <b>HERCULES 3 STAGE BALLOON</b><br>7FR 8-9-10 mm Balloon Diameter   | UNIT                  | 4                             |
| 12  | <b>HERCULES 3 STAGE BALLOON</b><br>7FR 10-11-12 mm Balloon Diameter   | UNIT                  | 4                             |
| 13  | <b>HERCULES 3 STAGE BALLOON</b><br>7FR 12-13.5-15 mm Balloon Diameter   | UNIT                  | 4                             |
| 14  | <b>HERCULES 3 STAGE BALLOON</b><br>7FR 15-16.5-18 mm Balloon Diameter   | UNIT                  | 4                             |
| 15  | <b>HERCULES 3 STAGE BALLOON</b><br>7FR 18-19-20 mm Balloon Diameter   | UNIT                  | 4                             |
| 16  | <b>FiAPC PROBE</b><br>▪ <b>2200-SC-Side Fire, Flexible</b><br>Must be compatible with machine <b>ERBE, Germany</b>                    | Per box<br>(10 units) | 1 box                         |
| 17  | <b>FiAPC PROBE</b><br>▪ <b>2200A- Straight Fire</b><br>Must be compatible with machine <b>ERBE, Germany</b>                           | Per box<br>(10 units) | 1 box                         |
| 18  | <b>FiAPC PROBE</b><br>▪ <b>2200C Circumferential Fire, Flexible</b><br>Must be compatible with machine <b>ERBE, Germany</b>           | Per box<br>(10 units) | 1 box                         |
| 19  | <b>STENT SET</b><br>▪ Drain obstructed biliary ducts<br>▪ wire guide diameter 0.035inch<br>stent various size ranges (french- length) | UNIT                  | 100                           |

| NO. | TERMS AND CONDITIONS  |
|-----|---|
| 1   | Tenderer must be registered with the Ministry of Health.  |
| 2   | <b>TENDER FORM should be filled</b> completely including the <b>USER REQUIREMENT FORM</b> (if available). Submission of incomplete form <b><u>MAY</u></b> cause <b>DISQUALIFICATION OF TENDER</b> .   |
| 3   | Each tenderer is allowed to quote <b>ONE BRAND WITH ONE PRICE ONLY</b> for each item. Submission of more than one brand and price will cause <b>DISQUALIFICATION OF TENDER</b> .  |
| 4   | All consumables supplied throughout this tender shall have a minimum expiry date of <b>twelve (12) months / on delivery</b> . Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made.  |
| 5   | <b>Brochures / catalogues should be submitted / attached</b> with tender document.  |
| 6   | <b>Samples should be submitted together with tender or within fourteen (14 days)</b> of the tender closing date (if applicable).  |
| 7   | <b>DELIVERY PERIOD:</b><br><b>Not later than 4 weeks</b><br><b>Staggered delivery upon request</b>  |
| 8   | <b>PRICE VALIDITY:</b><br>The quotation shall remain valid for <b>12 MONTHS</b> from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s). |

| Section/Unit      | Unit Gastroenterologi & Hepatologi  | Section/Unit Ref No.: |                  |
|-------------------|---|-----------------------|------------------|
| Person to Contact | Name: Dr Hjh Dewi Norwani binti Haji Basir<br>Konsultan dan Ketua Unit Gastroenterologi & Hepatologi<br>Hospital Raja Isteri Pengiran Anak Saleha (RIPAS) | Tel.No.:              | 2242424 ext 6236 |
|                   | E-mail: -   | Fax No.:              |                  |

**SECTION 3**  
**FORMS TO BE USED**

**CONTENTS**

|   |           |
|---|-----------|
| <b>SCHEDULE 1 – TENDER FORM .....</b>           | <b>1</b>  |
| <b>SCHEDULE 2 – INFORMATION SUMMARY .....</b>   | <b>4</b>  |
| <b>SCHEDULE 3 –SUB-CONTRACTS .....</b>          | <b>5</b>  |
| <b>SCHEDULE 4 –COMPANY BACKGROUND .....</b>     | <b>6</b>  |
| <b>SCHEDULE 5 –REFERENCES .....</b>             | <b>7</b>  |
| <b>SCHEDULE 6 –SUBMISSION OF SAMPLE ... ..</b>  | <b>8</b>  |
| <b>SCHEDULE 7 - LETTER OF DECLARATION .....</b> | <b>11</b> |

**SCHEDULE 1**

**TENDER FORM**

To:

**TENDER REFERENCE NO.: KK/22/2026/UPP(TC)**

**INVITATION TO TENDER**

**THE SUPPLY AND DELIVERY ENDOSCOPIC CONSUMABLES FOR THE GASTROENTEROLOGY AND HEPATOLOGY UNIT, RAJA ISTERI  
PENGIRAN ANAK SALEHA (RIPAS) HOSPITAL, FOR A PERIOD OF THREE (3) YEARS**

**TENDER OF (*name of tenderer*)** :

Company/Business Registration No. :

Tender Closing Date :

|                        |  |
|------------------------|--|
| <b>DELIVERY PERIOD</b> |  |
|------------------------|--|

| USER'S REQUIREMENTS |  |              |                               | VENDOR'S OFFER                       |                                  |              |                                 |                                     |                            |                            |
|---------------------|--|--------------|-------------------------------|--------------------------------------|----------------------------------|--------------|---------------------------------|-------------------------------------|----------------------------|----------------------------|
| NO.                 | ITEM DESCRIPTIONS AND SPECIFICATIONS   | PACKING SIZE | TOTAL QUANTITY USAGE PER YEAR | ITEM DESCRIPTIONS AND SPECIFICATIONS | PART/ CATALOGUE NUMBER AND BRAND | PACKING SIZE | TOTAL QUANTITY OFFERED PER YEAR | COST PER PACK (COST PER UNIT) (B\$) | TOTAL COSTS (B\$) (1 YEAR) | TOTAL COST (B\$) (3 YEARS) |
| 1                   | <b>HERCULES 3 STAGE WIRE GUIDED BALLOON</b><br>▪ P/N: HBD-W-8-9-10<br>Catheter 8Fr, catheter length 240cm, balloon length cm 5.5 | UNIT         | 4                             |                                      |                                  |              |                                 |                                     |                            |                            |
| 2                   | <b>HERCULES 3 STAGE WIRE GUIDED BALLOON</b><br>▪ P/N: HBD-W-10-11-12   | UNIT         | 4                             |                                      |                                  |              |                                 |                                     |                            |                            |

| USER'S REQUIREMENTS |  |              |                               | VENDOR'S OFFER                       |                                  |              |                                 |                                     |                            |                            |
|---------------------|--|--------------|-------------------------------|--------------------------------------|----------------------------------|--------------|---------------------------------|-------------------------------------|----------------------------|----------------------------|
| NO.                 | ITEM DESCRIPTIONS AND SPECIFICATIONS   | PACKING SIZE | TOTAL QUANTITY USAGE PER YEAR | ITEM DESCRIPTIONS AND SPECIFICATIONS | PART/ CATALOGUE NUMBER AND BRAND | PACKING SIZE | TOTAL QUANTITY OFFERED PER YEAR | COST PER PACK (COST PER UNIT) (B\$) | TOTAL COSTS (B\$) (1 YEAR) | TOTAL COST (B\$) (3 YEARS) |
|                     | Catheter 8Fr, catheter length 240cm, balloon length 5.5cm  |              |                               |                                      |                                  |              |                                 |                                     |                            |                            |
| 3                   | <b>HERCULES 3 STAGE WIRE GUIDED BALLOON</b><br>▪ P/N: HBD-W-12-13.5-15<br>Catheter 8Fr, catheter length 240cm, balloon length 5.5cm                                  | UNIT         | 4                             |                                      |                                  |              |                                 |                                     |                            |                            |
| 4                   | <b>HERCULES 3 STAGE WIRE GUIDED BALLOON</b><br>▪ P/N: HBD-W-15-16.5-18<br>Catheter 8Fr, catheter length 240cm, balloon diameter mm 15-16.5-18, balloon length 5.5 cm | UNIT         | 4                             |                                      |                                  |              |                                 |                                     |                            |                            |
| 5                   | <b>HERCULES 3 STAGE WIRE GUIDED BALLOON</b><br>▪ P/N: HBD-W-18-19-20<br>▪ Catheter 8Fr, Catheter length 240cm, balloon 18-19-20<br>balloon length 5.5cm              | UNIT         | 4                             |                                      |                                  |              |                                 |                                     |                            |                            |
| 6                   | <b>ACHALASIA 8CM BALLOON CATHETER</b><br>▪ 16Fr- 75cm<br>▪ Balloon Diameter 30mm, 8cm Length<br>Balloon Diameter 90 Fr   | UNIT         | 8                             |                                      |                                  |              |                                 |                                     |                            |                            |
| 7                   | <b>ACHALASIA 8CM BALLOON CATHETER</b><br>▪ 16fr- 75cm, Balloon Diameter 35 mm, 8cm Length<br>Balloon Diameter 105 Fr   | UNIT         | 8                             |                                      |                                  |              |                                 |                                     |                            |                            |

| USER'S REQUIREMENTS |   |                    |                               | VENDOR'S OFFER                       |                                  |              |                                 |                                     |                            |                            |
|---------------------|---|--------------------|-------------------------------|--------------------------------------|----------------------------------|--------------|---------------------------------|-------------------------------------|----------------------------|----------------------------|
| NO.                 | ITEM DESCRIPTIONS AND SPECIFICATIONS  | PACKING SIZE       | TOTAL QUANTITY USAGE PER YEAR | ITEM DESCRIPTIONS AND SPECIFICATIONS | PART/ CATALOGUE NUMBER AND BRAND | PACKING SIZE | TOTAL QUANTITY OFFERED PER YEAR | COST PER PACK (COST PER UNIT) (B\$) | TOTAL COSTS (B\$) (1 YEAR) | TOTAL COST (B\$) (3 YEARS) |
| 8                   | <b>CYTOMAX II DOUBLE LUMEN CYTOLOGY</b><br><ul style="list-style-type: none"> <li>▪ Catheter 8Fr, 200cm</li> <li>▪ Length Wire Guide Diameter 0.035"</li> <li>▪ Flipform Tipsize 1.5cm/3.5cm</li> </ul> Product code: G22674/G22675 | UNIT               | 72                            |                                      |                                  |              |                                 |                                     |                            |                            |
| 9                   | <b>CRE PRO WIRE GUIDED BALLOON DILATATION CATHETER</b><br><ul style="list-style-type: none"> <li>▪ Esophageal/Pyloric/Colonic/Biliary Inflated OD 8-10mm</li> </ul>   | UNIT               | 10                            |                                      |                                  |              |                                 |                                     |                            |                            |
| 10                  | <b>CRE PRO WIRE GUIDED BALLOON DILATATION CATHETER</b><br><ul style="list-style-type: none"> <li>▪ Esophageal/Pyloric/Colonic/Biliary Inflated OD 10-12mm</li> </ul>  | UNIT               | 10                            |                                      |                                  |              |                                 |                                     |                            |                            |
| 11                  | <b>HERCULES 3 STAGE BALLOON</b><br>7FR 8-9-10 mm Balloon Diameter   | UNIT               | 4                             |                                      |                                  |              |                                 |                                     |                            |                            |
| 12                  | <b>HERCULES 3 STAGE BALLOON</b><br>7FR 10-11-12 mm Balloon Diameter   | UNIT               | 4                             |                                      |                                  |              |                                 |                                     |                            |                            |
| 13                  | <b>HERCULES 3 STAGE BALLOON</b><br>7FR 12-13.5-15 mm Balloon Diameter   | UNIT               | 4                             |                                      |                                  |              |                                 |                                     |                            |                            |
| 14                  | <b>HERCULES 3 STAGE BALLOON</b><br>7FR 15-16.5-18 mm Balloon Diameter   | UNIT               | 4                             |                                      |                                  |              |                                 |                                     |                            |                            |
| 15                  | <b>HERCULES 3 STAGE BALLOON</b><br>7FR 18-19-20 mm Balloon Diameter   | UNIT               | 4                             |                                      |                                  |              |                                 |                                     |                            |                            |
| 16                  | <b>FiAPC PROBE</b><br><ul style="list-style-type: none"> <li>▪ 2200-SC-Side Fire, Flexible</li> </ul>   | Per box (10 units) | 1 box                         |                                      |                                  |              |                                 |                                     |                            |                            |

| USER'S REQUIREMENTS     |  |                       | VENDOR'S OFFER                |                                      |                                  |              |                                 |                                     |                            |                            |
|-------------------------|--|-----------------------|-------------------------------|--------------------------------------|----------------------------------|--------------|---------------------------------|-------------------------------------|----------------------------|----------------------------|
| NO.                     | ITEM DESCRIPTIONS AND SPECIFICATIONS   | PACKING SIZE          | TOTAL QUANTITY USAGE PER YEAR | ITEM DESCRIPTIONS AND SPECIFICATIONS | PART/ CATALOGUE NUMBER AND BRAND | PACKING SIZE | TOTAL QUANTITY OFFERED PER YEAR | COST PER PACK (COST PER UNIT) (B\$) | TOTAL COSTS (B\$) (1 YEAR) | TOTAL COST (B\$) (3 YEARS) |
|                         | Must be compatible with machine<br><b>ERBE, Germany</b>  |                       |                               |                                      |                                  |              |                                 |                                     |                            |                            |
| 17                      | <b>FiAPC PROBE</b><br>▪ <b>2200A- Straight Fire</b><br>Must be compatible with machine<br><b>ERBE, Germany</b>                       | Per box<br>(10 units) | 1 box                         |                                      |                                  |              |                                 |                                     |                            |                            |
| 18                      | <b>FiAPC PROBE</b><br>▪ <b>2200C Circumferential Fire, Flexible</b><br>Must be compatible with machine<br><b>ERBE, Germany</b>       | Per box<br>(10 units) | 1 box                         |                                      |                                  |              |                                 |                                     |                            |                            |
| 19                      | <b>STENT SET</b><br>▪ Drain obstructed biliary ducts<br>▪ wire guide diameter 0.035inch<br>stent various size ranges (french-length) | UNIT                  | 100                           |                                      |                                  |              |                                 |                                     |                            |                            |
| <b>TOTAL COST (B\$)</b> |  |                       |                               |                                      |                                  |              |                                 |                                     |                            |                            |

| NO. | TERMS AND CONDITIONS  | VENDOR'S OFFER<br>(PLEASE STATE)                    |
|-----|---|---|
| 1   | Tenderer must be registered with the Ministry of Health.  |   |
| 2   | <b>QUOTATION FORM should be filled</b> completely including the <b>USER REQUIREMENT FORM</b> (if available). Submission of incomplete form <b>MAY</b> cause <b>DISQUALIFICATION OF QUOTATION.</b>   |   |
| 3   | Each tenderer is allowed to quote <b>ONE BRAND WITH ONE PRICE ONLY</b> for each item. Submission of more than one brand and price will cause <b>DISQUALIFICATION OF QUOTATION.</b>  |   |
| 4   | All consumables supplied throughout this tender shall have a minimum expiry date of <b>twelve (12) months / on delivery.</b> Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made.   |   |
| 5   | <b>Brochures / catalogues should be submitted / attached</b> with quotation/tender document.  |   |
| 6   | <b>Samples should be submitted together with quotation or within fourteen (14 days)</b> of the quotation/tender closing date.   |   |
| 7   | <b>DELIVERY PERIOD:</b><br>(Please state)<br>Not later than 4 weeks   | <b>(Yes / No)</b><br><b>(If No, please specify)</b> |
| 8   | <b>PRICE VALIDITY:</b><br>The quotation shall remain valid for <b>12 MONTHS</b> from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s). |   |

1. We offer and undertake on your acceptance of our Tender to supply and deliver the above mentioned goods in accordance with your Invitation To Tender.
2. Our Tender is fully consistent with and does not contradict or derogate from anything in your Invitation To Tender. We have not qualified or changed any of the provisions of your Invitation To Tender.
3. We shall execute a formal agreement in the appropriate form set out in Section 4 – Contract of the Invitation to Tender together with such further terms and conditions, if any, agreed between the Government and us.
4. OUR OFFER IS VALID FOR **TWELVE (12)** CALENDER MONTHS FROM THE TENDER CLOSING DATE.
5. When requested by you, we shall extend the validity of this offer.
6. We further undertake to give you any further information which you may require.

Dated this                      day of    20

\_\_\_\_\_  
**[Signature of authorised officer of Tenderer]**  
Name:  
Designation:

Tenderer's official stamp:

## SCHEDULE 2 - INFORMATION SUMMARY

2.1 Tenderers shall provide in this Schedule the following information:

- a. Management summary
- b. Company profile (including Contractor and sub-contractor(s), if any)
- c. Years of experience (as of the Tender Closing Date) of the Contractor and sub-contractor(s) in the:
  - ***Supply and Delivery of Medical Consumables***
- d. Other information which is considered relevant

**SCHEDULE 3 – SUB-CONTRACTS**

- 3.1 Tenderers shall complete Table 3.1 with information about all the companies involved in the provision of the services and items specified in this tender. This shall include details about the Contractor and each sub-contractor involved, as well as their respective responsibilities.
- 3.2 Tenderers shall also indicate in Table 3.1 any alliance relationship established with each sub-contractor. An alliance is defined as a formal and binding business relationship between the allied parties.

Table 3.1 - Responsibility Table

| Company Name             | Responsibility<br>'Description | Alliance Relationship between<br>Contractor and Sub-contractor(s) |                           |                           |
|--------------------------|--------------------------------|---|---------------------------|---------------------------|
|                          |                                | Alliance<br>Exists?<br>(Y/N)                                      | Date<br>Established       | Alliance<br>Description   |
| <b>Contractor</b>        |                                |   |                           |                           |
|                          |                                | <b>Not<br/>Applicable</b>   | <b>Not<br/>Applicable</b> | <b>Not<br/>Applicable</b> |
| <b>Sub-contractor(s)</b> |                                |   |                           |                           |
|                          |                                |   |                           |                           |

#### **SCHEDULE 4 – COMPANY’S BACKGROUND**

- 4.1 Each of the companies involved in this tender, including Contractor and sub-contractor(s) (if any), shall provide information on the company’s background, scope of operations, financial standing and certified copy of its Certificate of Incorporation or Certificate of Registration (as the case may be).

**SCHEDULE 5 – REFERENCES**

5.1 Tenderers shall submit a list of customers in Table 5.1 to whom the Contractor has provided similar services and items as specified in this tender in the recent 5 years as of the Tender Closing Date.

Table 5.1 - References of previous customers

| Customer Name and Address | Customer Type (Govt or Quasi Govt)* | Contact Person | Title | Contact Number, Fax Number and E-mail Address |
|---------------------------|-------------------------------------|----------------|-------|---|
|                           |                                     |                |       |   |

**\*Note: Tenderers shall indicate whether the customer is a Government or Quasi Government organisation. A Quasi Government is defined as an organisation which (1) is managed and controlled by the Government; or (2) has at least 50% shares being held by the Government. Please leave the column blank if the customer is neither a Government or Quasi Government organisation.**

5.2 The Ministry of Health shall treat all the information submitted under this schedule in strict confidence.

5.3 The Ministry of Health reserves the right to contact the references for tender assessment purposes.

## **SCHEDULE 6 - SUBMISSION OF SAMPLE**

- 6.1 Tenderers shall submit the Submission of Sample form below in respect of the items specified in this tender.
- 6.2 Samples of the items to be submitted shall be:
- a) identical in packing and manufacture to the items to be offered by the Tenderer; and
  - b) marked with the corresponding item number of the tender.

## SUBMISSION OF SAMPLE FORM

TENDER REFERENCE NO.: KK/22/2026/UPP(TC)

### INVITATION TO TENDER

**THE SUPPLY AND DELIVERY ENDOSCOPIC CONSUMABLES FOR THE  
GASTROENTEROLOGY AND HEPATOLOGY UNIT, RAJA ISTERI PENGIRAN ANAK SALEHA  
(RIPAS) HOSPITAL, FOR A PERIOD OF THREE (3) YEARS**

SUBMISSION OF SAMPLE FORM OF (NAME OF TENDERER)

| ITEM NO. | DESCRIPTION  | SAMPLE SUBMITTED<br>(indicate with<br>✓) | SAMPLE NOT SUBMITTED<br>(indicate with<br>✗) | OFFERED/NOT OFFERED<br>(indicate as appropriate) |
|----------|--|--|--|--|
| 1        | <b>HERCULES 3 STAGE WIRE GUIDED BALLOON</b><br>▪ P/N: HBD-W-8-9-10<br>Catheter 8Fr, catheter length 240cm, balloon length cm 5.5                                     |  |  |  |
| 2        | <b>HERCULES 3 STAGE WIRE GUIDED BALLOON</b><br>▪ P/N: HBD-W-10-11-12<br>Catheter 8Fr, catheter length 240cm, balloon length 5.5cm                                    |  |  |  |
| 3        | <b>HERCULES 3 STAGE WIRE GUIDED BALLOON</b><br>▪ P/N: HBD-W-12-13.5-15<br>Catheter 8Fr, catheter length 240cm, balloon length 5.5cm                                  |  |  |  |
| 4        | <b>HERCULES 3 STAGE WIRE GUIDED BALLOON</b><br>▪ P/N: HBD-W-15-16.5-18<br>Catheter 8Fr, catheter length 240cm, balloon diameter mm 15-16.5-18, balloon length 5.5 cm |  |  |  |
| 5        | <b>HERCULES 3 STAGE WIRE GUIDED BALLOON</b><br>▪ P/N: HBD-W-18-19-20<br>▪ Catheter 8Fr, Catheter length 240cm, balloon 18-19-20<br>balloon length 5.5cm              |  |  |  |
| 6        | <b>ACHALASIA 8CM BALLOON CATHETER</b><br>▪ 16Fr- 75cm<br>▪ Balloon Diameter 30mm, 8cm Length<br>Balloon Diameter 90 Fr   |  |  |  |
| 7        | <b>ACHALASIA 8CM BALLOON CATHETER</b><br>▪ 16fr- 75cm, Balloon Diameter 35 mm, 8cm Length<br>Balloon Diameter 105 Fr   |  |  |  |

| ITEM NO. | DESCRIPTION  | SAMPLE SUBMITTED<br>(indicate with ✓) | SAMPLE NOT SUBMITTED<br>(indicate with X) | OFFERED/NOT OFFERED<br>(indicate as appropriate) |
|----------|--|---------------------------------------|---|--|
| 8        | <b>CYTOMAX II DOUBLE LUMEN CYTOLOGY</b> <ul style="list-style-type: none"> <li>▪ Catheter 8Fr, 200cm</li> <li>▪ Length Wire Guide Diameter 0.035"</li> <li>▪ Flipform Tipsize 1.5cm/3.5cm</li> </ul> Product code: G22674/G22675 |                                       |   |  |
| 9        | <b>CRE PRO WIRE GUIDED BALLOON DILATATION CATHETER</b> <ul style="list-style-type: none"> <li>▪ Esophageal/Pyloric/Colonic/Biliary</li> </ul> Inflated OD 8-10mm   |                                       |   |  |
| 10       | <b>CRE PRO WIRE GUIDED BALLOON DILATATION CATHETER</b> <ul style="list-style-type: none"> <li>▪ Esophageal/Pyloric/Colonic/Biliary</li> </ul> Inflated OD 10-12mm  |                                       |   |  |
| 11       | <b>HERCULES 3 STAGE BALLOON</b><br>7FR 8-9-10 mm Balloon Diameter  |                                       |   |  |
| 12       | <b>HERCULES 3 STAGE BALLOON</b><br>7FR 10-11-12 mm Balloon Diameter  |                                       |   |  |
| 13       | <b>HERCULES 3 STAGE BALLOON</b><br>7FR 12-13.5-15 mm Balloon Diameter  |                                       |   |  |
| 14       | <b>HERCULES 3 STAGE BALLOON</b><br>7FR 15-16.5-18 mm Balloon Diameter  |                                       |   |  |
| 15       | <b>HERCULES 3 STAGE BALLOON</b><br>7FR 18-19-20 mm Balloon Diameter  |                                       |   |  |
| 16       | <b>FiAPC PROBE</b> <ul style="list-style-type: none"> <li>▪ <b>2200-SC-Side Fire, Flexible</b></li> </ul> Must be compatible with machine <b>ERBE, Germany</b>   |                                       |   |  |
| 17       | <b>FiAPC PROBE</b> <ul style="list-style-type: none"> <li>▪ <b>2200A- Straight Fire</b></li> </ul> Must be compatible with machine <b>ERBE, Germany</b>  |                                       |   |  |
| 18       | <b>FiAPC PROBE</b> <ul style="list-style-type: none"> <li>▪ <b>2200C Circumferential Fire, Flexible</b></li> </ul> Must be compatible with machine <b>ERBE, Germany</b>  |                                       |   |  |
| 19       | <b>STENT SET</b> <ul style="list-style-type: none"> <li>▪ Drain obstructed biliary ducts</li> <li>▪ wire guide diameter 0.035inch</li> </ul> stent various size ranges (french- length)  |                                       |   |  |

We understand as stated in the Instructions to Tenderers that Tenders without samples shall not be considered.

Tenderer's official stamp:

\_\_\_\_\_  
***[Signature of authorized officer of Tenderer]***

Name:

Designation:

Date:

---

**FOR OFFICE USE**

Date of receipt: \_\_\_\_\_

Receiving Officer: \_\_\_\_\_