

LAMPIRAN - A1

RUJ: BME/A2K/(1624)/2026/HTD (A19)

25/2/2026

ONE-OFF PREVENTIVE MAINTENANCE SERVICE OF DAAVLIN 3 SERIES PC PHOTOTHERAPY MACHINE, S/N: P3514-ASNBPC-2424 BME7688, IN DERMATOLOGY CLINIC, SSB HOSPITAL KUALA BELAIT

SCOPE OF WORK

EQUIPMENT DETAILS:

- PHOTOTHERAPY MACHINE
- BRAND: DAAVLIN, MODEL: 3 SERIES PC
- BME NO: 7688
- SN: P3514-ASNBPC-2424
- LOCATION: DERMATOLOGY CLINIC, SSB HOSPITAL

SCOPE OF WORK:

- 1) PERFORM PREVENTIVE MAINTENANCE FOR BOTH EQUIPMENTS.
 - CHECKING AND TESTING OF 48 NOS. OF UV TUBES (UVA+NB UVB)
 - CHECK AND TEST THE FUNCTION OF CONTROL PANEL
 - CHECKING OF UVA SENSOR
 - CHECKING OF UVB SENSOR
 - CHECK AND TEST THE DOOR SWITCHES
 - CHECK THE STARTER, CAPACITOR AND BALLAST
 - CHECK AND TEST THE INPUT/OUTPUT POWER
 - DISMANTLE AND CLEANING OF 48 NOS. UV TUBES
 - CLEANING OF REFLECTORS
 - CLEANING OF SOCKET COVERS
 - CLEANING OF CONTROL PANEL
 - VACCUUM THE DUST AND DIRT IN THE CABIN
 - CALIBRATION OF THE SYSTEM USING DAAVLIN'S CALIBRATOR (UV METER)
 - CALIBRATION FOR EVERY 100 LAMP HOURS LIMITED TO 3 TIMES ONLY
- 2) REPLACE A TOTAL OF 24 UNITS OF UVB TUBES
(24 NOS OF UVB TUBES FOR EACH PHOTOTHERAPY MACHINE)

LAMPIRAN - A2

RUJ: BME/A2K/(1624)/2026/HTD (A19)

ONE-OFF PREVENTIVE MAINTENANCE SERVICE OF DAAVLIN 3 SERIES PC PHOTOTHERAPY MACHINE, S/N: P3514-ASNBPC-2424 BME7688, IN DERMATOLOGY CLINIC, SSB HOSPITAL KUALA BELAIT

	SPAREPART	PART NUMBER	QUANTITY
1	UVB TUBES		24
2	PM CHARGES		1