

## LAMPIRAN - A1

RUJ: BME/A2K/(1628)/2026/HTD (A20)

2/3/2026

ONE OFF PREVENTIVE MAINTENANCE SERVICES FOR ONE (1) UNIT OF AUTO DISINTEGRATOR EQUIPMENT, BRAND: PHARMA TEST, MODEL: PTZ AUTO 3EZ, S/N: 20222, BME17822, AT PHARMACY SECTION, DEPARTMENT SCIENTIFIC SERVICES.

### SCOPE OF WORK

One off preventive maintenance services for one (1) unit of Auto Disintegrator, Brand: Pharma Test, Model: PTZ Auto 3EZ, S/N: 20222, BME17822 at Pharmacy Section, Department of Scientific Services, Madaras.

- Scope of work to follow Pharma Test preventive maintenance manual/recommendation specific for model PTZ Auto 3EZ which include the following where necessary:

- (i) inspection
- (ii) cleaning
- (iii) installation of PM kits, consumables and spare parts listed in spareparts list (if any in Lampiran A2)
- (iv) alignment
- (v) calibration
- (vi) any other related preventive maintenance works required.

- Functional check post-maintenance.

- Apply preventive maintenance sticker/tag after successful service.

- Disposal of any used spare parts/pm kits, unless otherwise instructed.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to Pharma Test disintegrators.

- Vendor(s) to submit maintenance checklist/details/scope specific for Pharma Test PTZ Auto 3EZ disintegrator along with other necessary documents to join this tender.

- Upon approval, the earliest the preventive maintenance work can be carried out on the equipment is one month before the next PM due date, and not anytime earlier.