

# LAMPIRAN - A1

RUJ: BME/A2K/(1629)/2026/HTD (A20)

9/3/2026

TO REPAIR ONE (1) UNIT OF ULTRAPURE WATER PURIFICATION SYSTEM, BRAND: ELGA, MODEL: PURELAB CHORUS 1 COMPLETE 10BP AT TOXICOLOGY LABORATORY, DEPARTMENT OF SCIENTIFIC SERVICES

## SCOPE OF WORK

To repair one (1) unit of Ultrapure Water Purification system, Brand: Elga, Model: Purelab Chorus 1 complete 10BP, S/N: CRS00006916, BME24828 at Toxicology Laboratory, Department of Scientific Services.

- Scope of work to follow equipment manual/recommendation specific for this system which include the following where if necessary:
  - (i) installation of items listed in spareparts list (Lampiran A2)
  - (ii) alignment
  - (iii) calibration
  - (iv) optimisation
  - (v) any other related maintenance works required after installation of items listed in spareparts list (Lampiran A2)
- Functional check post-maintenance.
- Disposal of any used spare parts, unless otherwise instructed.
- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.
- Vendor(s) to submit maintenance checklist/details/scope specific for the equipment along with other necessary documents to join this tender.

## LAMPIRAN - A2

RUJ: BME/A2K/(1629)/2026/HTD (A20)

TO REPAIR ONE (1) UNIT OF ULTRAPURE WATER PURIFICATION SYSTEM, BRAND: ELGA, MODEL: PURELAB CHORUS 1 COMPLETE 10BP AT TOXICOLOGY LABORATORY, DEPARTMENT OF SCIENTIFIC SERVICES

	<b>SPAREPART</b>	<b>PART NUMBER</b>	<b>QUANTITY</b>
1	PUMP DIAPHRAGM BYPASS SET	EL/SV/O/PUMP3594	1
2	BOOST PUMP	EL/PUMP/39821-02	1