

TAWARAN SEBUTHARGA  
 JABATAN PERKHIDMATAN KESIHATAN

Rujukan Tawaran: **2/QTN/JPK/2026**

Nama Tawaran:

**SUPPLY AND DELIVERY OF ONE HUNDRED (100) UNITS OF REUSABLE UNIVERSAL  
 PAEDIATRIC SPO2 SENSOR FOR DEPARTMENT OF HEALTH SERVICES, MINISTRY OF HEALTH**

<b>SPECIFICATIONS</b>	
<b>STANDARD FEATURES/ USER REQUIREMENTS</b>	<b>Latex free</b>
	<b>9-pin Male connector</b>
	<b>Suitable for measuring SPO2 and pulse rate</b>
	<b>Finger clip type</b>
	<b>REUSABLE TYPE</b>
	<b>Easy to clean</b>
	<b>Suitable for all types of pulse oximeter machine (e.g. Nellcor, Edan, Creative Medical and Mindray brands). Attached are picture sample for the SPO2 sensor</b>
	<b>Accuracy of SPO2 sensor 4%-5% for standard SPO2 measurement</b>
	<b>High sensitivity</b>
	<b>Ease of use &amp; comfort with secure fit with minimal pressure on patient finger</b>
	<b>Sample should be provided for testing on existing SPO2 machine (unable to provide will be considered non-compliance)</b>

**PROCUREMENT REQUIREMENT**

- Brand/ Model
- Accessories
- Country of origin
- International Safety Standard
- Year Manufactured
- Price per unit
- Total price
- Delivery period
- Price Validity
- Sample & Brochure
- Warranty

• Authorized by:  
.....

TAWARAN SEBUTHARGA  
JABATAN PERKHIDMATAN KESIHATAN

Rujukan Tawaran: **2/QTN/JPK/2026**

Nama Tawaran:

**SUPPLY AND DELIVERY OF ONE HUNDRED (100) UNITS OF REUSABLE UNIVERSAL  
PAEDIATRIC SPO2 SENSOR FOR DEPARTMENT OF HEALTH SERVICES, MINISTRY OF HEALTH**

	ITEMS	REMARKS TENDERER SHOULD SPECIFY (YES OR NO)
<b>SPECIFICATIONS/ USER REQUIREMENTS</b>	Latex free	
	9-pin Male connector	
	Suitable for measuring SPO2 and pulse rate	
	Finger clip type	
	REUSABLE TYPE	
	Easy to clean	
	Suitable for all types of pulse oximeter machine (e.g. Nellcor, Edan, Creative Medical and Mindray brands). Attached are picture sample for the SPO2 sensor	
	Accuracy of SPO2 sensor 4%-5% for standard SPO2 measurement	
	High sensitivity	
	Ease of use & comfort with secure fit with minimal pressure on patient finger	
	Sample should be provided for testing on existing SPO2 machine (unable to provide will be considered non-compliance)	

PROCUREMENT REQUIREMENT	
Brand/ Model	
Country of origin	
International Safety Standard	
Year Manufactured	
Price per unit	
Total price	
Delivery period	
Price Validity	
Sample & Brochure	
Warranty	