

LAMPIRAN - A1

RUJ: BME/A2K/(1584)/2026/HTD (A20)

18/3/2026

REPAIR FOR ONE (1) UNIT OF HYDROGEN GENERATOR BRAND: LEMAN, MODEL: FID STATION RC250/3000, S/N: 40721, BME29458 AT ILLICIT DRUGS LABORATORY, DEPARTMENT OF SCIENTIFIC SERVICES

SCOPE OF WORK

Repair service for one (1) unit of hydrogen generator, Brand: Lemman Instruments, Model: FID Station RC250/3000 S/N: 40721, BME29458 at Illicit Drugs laboratory, inclusive but not limit to:

- Scope of work to follow equipment manual/recommendation specific for this system which include the following where necessary:
 - (i) installation of items listed in spareparts list (if any in Lampiran A2)
 - (i) alignment
 - (iii) calibration
 - (iv) optimisation
 - (v) any other related maintenance works required after installation of items listed in spareparts list
- Functional check post-maintenance.
- Disposal of any used spare parts/pm kits, unless otherwise instructed.
- Work to be carried out by engineer(s) authorised in providing maintenance to the equipment.

LAMPIRAN - A2

RUJ: BME/A2K/(1584)/2026/HTD (A20)

REPAIR FOR ONE (1) UNIT OF HYDROGEN GENERATOR BRAND: LEMAN, MODEL: FID STATION RC250/3000, S/N: 40721, BME29458 AT ILLICIT DRUGS LABORATORY, DEPARTMENT OF SCIENTIFIC SERVICES

	SPAREPART	PART NUMBER	QUANTITY
1	Water Level Sensor	91-0057	2
2	Electro Valve	91-0009	1
3	MKH2-1 deionization bag + water filter	71-1080	2