

TENDER REFERENCE NO.: KK/63/2026/PPN(TC)

**MINISTRY OF HEALTH
NEGARA BRUNEI DARUSSALAM**

**SUPPLY AND DELIVERY OF ORTHODONTIC
CONSUMABLE ITEMS FOR DEPARTMENT OF DENTAL
SERVICES, MINISTRY OF HEALTH FOR A PERIOD OF
THREE (3) YEARS**

TENDER FEES : \$10.00

RECEIPT NO. :

CLOSING DATE : 05th May 2026

TIME : 2.00 PM

FOA :

**THE CHAIRMAN
MINI TENDER BOARD, TENDER BOX
GROUND FLOOR, MINISTRY OF HEALTH
COMMONWEALTH DRIVE
BANDAR SERI BEGAWAN BB3910
NEGARA BRUNEI DARUSSALAM**

(NON-CLUSTERING)

SECTION 2

SPECIFICATIONS

TENDER REF. NO.: KK/63/2026/PPN(TC)

**INVITATION TO TENDER
SUPPLY AND DELIVERY OF ORTHODONTIC CONSUMABLE ITEMS FOR DEPARTMENT OF
DENTAL SERVICES, MINISTRY OF HEALTH FOR A PERIOD OF THREE (3) YEARS**

DELIVERY PERIOD	NOT MORE THAN 60 DAYS UPON CONFIRMATION
------------------------	--

NO.	ITEM DESCRIPTION	QUANTITY
1	<p>BRACKET REMOVING PLIERS, ANGLED Should have the following properties:</p> <ul style="list-style-type: none"> a) Medical grade stainless steel b) Corrosion and stain resistant c) For use intra-orally d) For the removal of anterior and posterior brackets e) Angled cutting surface (maximum 45 degrees) f) Sterilizable and able to be autoclaved g) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) 	15 PIECES
2	<p>PIN AND LIGATURE CUTTER</p> <ul style="list-style-type: none"> a) Medical grade stainless steel b) Corrosion and stain resistant c) For use intra-orally d) To cut soft or ligature wires (maximum 0.015" or 0.30mm) e) Sterilizable and able to be autoclaved h) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) 	15 PIECES
3	<p>MOSQUITO FORCEPS WITHOUT HOOK</p> <ul style="list-style-type: none"> a) Medical grade stainless steel b) Corrosion and stain resistant c) For use intra-orally d) For placing elastomeric ligatures e) Narrow, serrated tips with no hook f) Locking handles g) Maximum 5 ½" long h) Sterilizable and able to be autoclaved i) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) 	20 PIECES
4	<p>MATHEIU NEEDLE HOLDER</p> <ul style="list-style-type: none"> a) Medical grade stainless steel b) Corrosion and stain resistant c) For use intra-orally d) For tightening ligatures e) Rounded handles to allow spinning of instrument in the palm of the hand f) Narrow, serrated tips g) Safety locking handles- squeezable locking mechanism to 'lock' or 'release' the ligature h) To hold up to 0.015" or 0.30mm soft ligature wires i) Sterilizable and able to be autoclaved j) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) 	5 PIECES

NO.	ITEM DESCRIPTION	QUANTITY
5	<p>DISTAL END CUTTER</p> <ul style="list-style-type: none"> a) Medical grade stainless steel b) Corrosion and stain resistant c) For use intra- or extra- orally d) Used to cut orthodontic archwires e) Safely holds onto the cut distal end portions f) Able to cut flush to the buccal tube g) Smooth design for patient comfort h) To cut wire up to maximum 0.012"-0.021" x 0.025" or 0.030mm -0.53mm x 0.65mm wide i) Sterilizable and able to be autoclaved j) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) k) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) 	20 PIECES
6	<p>BAND SEATING INSTRUMENT</p> <ul style="list-style-type: none"> a) Medical grade plastic and stainless steel b) Corrosion and stain resistant c) For use intra-orally d) Used to fit and seat orthodontic bands e) Handle can also be used to apply occlusal pressure onto band during seating f) Serrated triangular metal tip to allow application of occlusal pressure to edges of band g) Sterilizable and able to be autoclaved h) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) 	5 PIECES
7	<p>DOUBLE ENDED LIGATING INSTRUMENT</p> <ul style="list-style-type: none"> a) Medical grade stainless steel b) Corrosion and stain resistant c) For use intra-orally d) Used to push steel ligatures onto archwires e) Notched on both ends f) Both ends must be of different angles/ orientation- to allow use anteriorly and posteriorly g) Sterilizable and able to be autoclaved h) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) 	15 PIECES

SECTION 3
FORMS TO BE USED
CONTENTS

SCHEDULE 1 – TENDER FORM1

SCHEDULE 2 – INFORMATION SUMMARY6

SCHEDULE 3 –SUB-CONTRACTS7

SCHEDULE 4 –COMPANY BACKGROUND8

SCHEDULE 5 –REFERENCES9

SCHEDULE 6 –SUBMISSION OF SAMPLE10

SCHEDULE 7 - LETTER OF DECLARATION13

SECTION 3

TENDER FORM

To:

TENDER REF. NO.: KK/63/2026/PPN(TC)

INVITATION TO TENDER

SUPPLY AND DELIVERY OF ORTHODONTIC CONSUMABLE ITEMS FOR DEPARTMENT OF DENTAL SERVICES, MINISTRY OF HEALTH FOR A PERIOD OF THREE (3) YEARS

TENDER OF (*name of tenderer*) _____

Company/Business Registration No _____

Tender Closing Date: _____

DELIVERY PERIOD

NO.	ITEM DESCRIPTION	QUANTITY	BRAND	MANUFACTURER AND ORIGIN	UNIT PRICE (B\$)	TOTAL PRICE (B\$)
1	BRACKET REMOVING PLIERS, ANGLED Should have the following properties: a) Medical grade stainless steel b) Corrosion and stain resistant c) For use intra-orally d) For the removal of anterior and posterior brackets e) Angled cutting surface (maximum 45 degrees) f) Sterilizable and able to be autoclaved g) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date)	15 PIECES				

NO.	ITEM DESCRIPTION	QUANTITY	BRAND	MANUFACTURER AND ORIGIN	UNIT PRICE (B\$)	TOTAL PRICE (B\$)
2	<p>PIN AND LIGATURE CUTTER</p> <ul style="list-style-type: none"> a) Medical grade stainless steel b) Corrosion and stain resistant c) For use intra-orally d) To cut soft or ligature wires (maximum 0.015" or 0.30mm) e) Sterilizable and able to be autoclaved f) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) 	15 PIECES				
3	<p>MOSQUITO FORCEPS WITHOUT HOOK</p> <ul style="list-style-type: none"> a) Medical grade stainless steel b) Corrosion and stain resistant c) For use intra-orally d) For placing elastomeric ligatures e) Narrow, serrated tips with no hook f) Locking handles g) Maximum 5 ½" long h) Sterilizable and able to be autoclaved i) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) 	20 PIECES				
4	<p>MATHEIU NEEDLE HOLDER</p> <ul style="list-style-type: none"> a) Medical grade stainless steel b) Corrosion and stain resistant c) For use intra-orally d) For tightening ligatures e) Rounded handles to allow spinning of instrument in the palm of the hand f) Narrow, serrated tips g) Safety locking handles- squeezable locking mechanism to 'lock' or 'release' the ligature h) To hold up to 0.015" or 0.30mm soft ligature wires i) Sterilizable and able to be autoclaved 	5 PIECES				

NO.	ITEM DESCRIPTION	QUANTITY	BRAND	MANUFACTURER AND ORIGIN	UNIT PRICE (B\$)	TOTAL PRICE (B\$)
	j) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date)					
5	DISTAL END CUTTER a) Medical grade stainless steel b) Corrosion and stain resistant c) For use intra- or extra- orally d) Used to cut orthodontic archwires e) Safely holds onto the cut distal end portions f) Able to cut flush to the buccal tube g) Smooth design for patient comfort h) To cut wire up to maximum 0.012"-0.021" x 0.025" or 0.030mm -0.53mm x 0.65mm wide i) Sterilizable and able to be autoclaved j) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) k) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date)	20 PIECES				
6	BAND SEATING INSTRUMENT a) Medical grade plastic and stainless steel b) Corrosion and stain resistant c) For use intra-orally d) Used to fit and seat orthodontic bands e) Handle can also be used to apply occlusal pressure onto band during seating f) Serrated triangular metal tip to allow application of occlusal pressure to edges of band g) Sterilizable and able to be autoclaved h) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date)	5 PIECES				

NO.	ITEM DESCRIPTION	QUANTITY	BRAND	MANUFACTURER AND ORIGIN	UNIT PRICE (B\$)	TOTAL PRICE (B\$)
7	DOUBLE ENDED LIGATING INSTRUMENT a) Medical grade stainless steel b) Corrosion and stain resistant c) For use intra-orally d) Used to push steel ligatures onto archwires e) Notched on both ends f) Both ends must be of different angles/ orientation- to allow use anteriorly and posteriorly g) Sterilizable and able to be autoclaved h) Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date)	15 PIECES				
TOTAL PRICE PER YEAR (B\$)						
TOTAL PRICE FOR THREE (3) YEARS (B\$)						

NO.	USER REQUIREMENTS	YES	NO	PLEASE SPECIFY THE SPECIFICATION/ SCOPE OFFERED FAILURE TO FILL THE OFFER DETAIL SHALL BE IDENTIFIED AS NON-COMPLIANCE TO THE OFFER	
1	BRACKET REMOVING PLIERS, ANGLED	Should have the following properties:			
		Medical grade stainless steel			
		Corrosion and stain resistant			
		For use intra-orally			
		For the removal of anterior and posterior brackets			
		Angled cutting surface (maximum 45 degrees)			
		Sterilizable and able to be autoclaved			
		Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) 15 pieces			
2	PIN LIGATURE CUTTER AND	Medical grade stainless steel			
		Corrosion and stain resistant			
		For use intra-orally			
		To cut soft or ligature wires (maximum 0.015" or 0.30mm)			
		Sterilizable and able to be autoclaved			
		Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date)			
		15 pieces			
3	MOSQUITO FORCEPS WITHOUT HOOK	Medical grade stainless steel			
		Corrosion and stain resistant			
		For use intra-orally			
		For placing elastomeric ligatures			
		Narrow, serrated tips with no hook			
		Locking handles			
		Maximum 5 ½" long			
		Sterilizable and able to be			

NO.	USER REQUIREMENTS	YES	NO	PLEASE SPECIFY THE SPECIFICATION/ SCOPE OFFERED FAILURE TO FILL THE OFFER DETAIL SHALL BE IDENTIFIED AS NON-COMPLIANCE TO THE OFFER
	autoclaved Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) 20 pieces			
4	MATHEIU NEEDLE HOLDER Medical grade stainless steel Corrosion and stain resistant For use intra-orally For tightening ligatures Rounded handles to allow spinning of instrument in the palm of the hand Narrow, serrated tips Safety locking handles- squeezable locking mechanism to 'lock' or 'release' the ligature To hold up to 0.015" or 0.30mm soft ligature wires Sterilizable and able to be autoclaved Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) 5 pieces			
5	DISTAL END CUTTER Medical grade stainless steel Corrosion and stain resistant For use intra- or extra- orally Used to cut orthodontic archwires Safely holds onto the cut distal end portions Able to cut flush to the buccal tube Smooth design for patient comfort To cut wire up to maximum 0.012"-0.021" x 0.025" or 0.030mm - 0.53mm x 0.65mm wide			

NO.	USER REQUIREMENTS	YES	NO	PLEASE SPECIFY THE SPECIFICATION/ SCOPE OFFERED FAILURE TO FILL THE OFFER DETAIL SHALL BE IDENTIFIED AS NON-COMPLIANCE TO THE OFFER
	Sterilizable and able to be autoclaved Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) 20 pieces			
6	BAND SEATING INSTRUMENT Medical grade plastic and stainless steel Corrosion and stain resistant For use intra-orally Used to fit and seat orthodontic bands Handle can also be used to apply occlusal pressure onto band during seating Serrated triangular metal tip to allow application of occlusal pressure to edges of band Sterilizable and able to be autoclaved Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date) 5 pieces			
7	DOUBLE ENDED LIGATING INSTRUMENT Medical grade stainless steel Corrosion and stain resistant For use intra-orally Used to push steel ligatures onto archwires Notched on both ends Both ends must be of different angles/ orientation- to allow use anteriorly and posteriorly Sterilizable and able to be autoclaved			

NO.	USER REQUIREMENTS		YES	NO	PLEASE SPECIFY THE SPECIFICATION/ SCOPE OFFERED FAILURE TO FILL THE OFFER DETAIL SHALL BE IDENTIFIED AS NON-COMPLIANCE TO THE OFFER
		Samples must be provided for accurate evaluation upon submission of quotation (30 days after closing date)			
		15 pieces			

1. We offer and undertake on your acceptance of our Tender to supply and deliver the above mentioned goods in accordance with your Invitation To Tender.
2. Our Tender is fully consistent with and does not contradict or derogate from anything in your Invitation To Tender. We have not qualified or changed any of the provisions of your Invitation To Tender.
3. We shall execute a formal agreement in the appropriate form set out in Section 4 – Contract of the Invitation to Tender together with such further terms and conditions, if any, agreed between the Government and us.
4. OUR OFFER IS VALID FOR **TWELVE (12)** CALENDER MONTHS FROM THE TENDER CLOSING DATE.
5. When requested by you, we shall extend the validity of this offer.
6. We further undertake to give you any further information which you may require.

Dated this _____ day of _____, _____.

Signature of authorised officer of Tenderer

Name:

Designation:

Tenderer's official stamp:

SCHEDULE 2 - INFORMATION SUMMARY

2.1 Tenderers shall provide in this Schedule the following information:

- a. Management summary
- b. Company profile (including Contractor and sub-contractor(s), if any)
- c. Years of experience (as of the Tender Closing Date) of the Contractor and sub-contractor(s) in the:
 - *Supply and Delivery of Consumables Items*
- d. Other information which is considered relevant

SCHEDULE 3 – SUB-CONTRACTS

- 3.1 Tenderers shall complete Table 3.1 with information about all the companies involved in the provision of the services and items specified in this tender. This shall include details about the Contractor and each sub-contractor involved, as well as their respective responsibilities.
- 3.2 Tenderers shall also indicate in Table 3.1 any alliance relationship established with each sub-contractor. An alliance is defined as a formal and binding business relationship between the allied parties.

Table 3.1 - Responsibility Table

Company Name	Responsibility Description	Alliance Relationship between Contractor and Sub-contractor(s)		
		Alliance Exists? (Y/N)	Date Established	Alliance Description
Contractor				
		Not Applicable	Not Applicable	Not Applicable
Sub-contractor(s)				

SCHEDULE 4 – COMPANY’S BACKGROUND

- 4.1 Each of the companies involved in this tender, including Contractor and sub-contractor(s) (if any), shall provide information on the company’s background, scope of operations, financial standing and certified copy of its Certificate of Incorporation or Certificate of Registration (as the case may be).

SCHEDULE 5 – REFERENCES

5.1 Tenderers shall submit a list of customers in Table 4.1 to whom the Contractor has provided similar services and items as specified in this tender in the recent 5 years as of the Tender Closing Date.

Table 5.1 References of previous customers

Customer Name and Address	Customer Type (Govt or Quasi Govt)*	Contact Person	Title	Contact Number, Fax Number and E-mail Address

***Note: Tenderers shall indicate whether the customer is a Government or Quasi Government organisation. A Quasi Government is defined as an organisation which (1) is managed and controlled by the Government; or (2) has at least 50% shares being held by the Government. Please leave the column blank if the customer is neither a Government or Quasi Government organisation.**

5.2 The Ministry of Health shall treat all the information submitted under this schedule in strict confidence.

5.3 The Ministry of Health reserves the right to contact the references for tender assessment purposes.

SCHEDULE 6 - SUBMISSION OF SAMPLE

- 6.1 Tenderers shall submit the Submission of Sample form below in respect of the items specified in this tender.
- 6.2 Samples of the items to be submitted shall be:
- a) identical in packing and manufacture to the items to be offered by the Tenderer; and
 - b) marked with the corresponding item number of the tender.

SUBMISSION OF SAMPLE FORM

To:

TENDER REF. NO.: KK/63/2026/PPN(TC)

**INVITATION TO TENDER
SUPPLY AND DELIVERY OF ORTHODONTIC CONSUMABLE ITEMS FOR DEPARTMENT OF
DENTAL SERVICES, MINISTRY OF HEALTH FOR A PERIOD OF THREE (3) YEARS**

SUBMISSION OF SAMPLE FORM OF (NAME OF TENDERER)

ITEM NO.	DESCRIPTION	SAMPLE SUBMITTED (indicate with ✓)	SAMPLE NOT SUBMITTED (indicate with X)	OFFERED/ NOT OFFERED (indicate as appropriate)
1	BRACKET REMOVING PLIERS, ANGLED			
2	PIN AND LIGATURE CUTTER			
3	MOSQUITO FORCEPS WITHOUT HOOK			
4	MATHEIU NEEDLE HOLDER			
5	DISTAL END CUTTER			
6	BAND SEATING INSTRUMENT			
7	DOUBLE ENDED LIGATING INSTRUMENT			

We understand as stated in the Instructions to Tenderers that Tenders without samples shall not be considered.

(Signature of authorized officer of Tenderer)

Name:

Designation:

Date:

Tenderer's official stamp:

FOR OFFICE USE

Date of Receipt: _____

Receiving Officer: _____