

TENDER REFERENCE NO.: KK/64/2026/HTD

**MINISTRY OF HEALTH
NEGARA BRUNEI DARUSSALAM**

**THE PROVISION OF MAINTENANCE SERVICES FOR GE
HEALTHCARE, MAC 2000 AND MAC 5 ECGs IN MINISTRY
OF HEALTH**

TENDER FEES : \$10.00

RECEIPT NO. :

CLOSING DATE : 05th May 2026

TIME : 2.00 PM

FOA :

**THE CHAIRMAN
MINI TENDER BOARD, TENDER BOX
GROUND FLOOR, MINISTRY OF HEALTH
COMMONWEALTH DRIVE
BANDAR SERI BEGAWAN BB3910
NEGARA BRUNEI DARUSSALAM**

(CLUSTERING)

SECTION 2

SPECIFICATIONS AND REQUIREMENTS

TENDER REFERENCE NO: KK/64/2026/HTD

INVITATION TO TENDER THE PROVISION OF MAINTENANCE SERVICES FOR GE HEALTHCARE, MAC 2000 AND MAC 5 ECGs IN MINISTRY OF HEALTH

The provision of maintenance services for:

1. Eighty-Nine (89) units of GE MAC 2000 ECGs and its related accessories:
 - i. BME29776 - BME29862, and BME19747
 - ii. 3 units DICOM gateway server
2. Eighty-Two (82) units of GE MAC 5 ECGs
 - i. BME30414 - BME30416, BME31256 - BME31328 and BME31421 - BME31426

Scope of Work:

1. To perform PPM once a year
2. To conduct electrical safety test
3. To conduct performance verification using multiparameter patient simulator
4. To conduct DICOM gateway/ CPACS connectivity test
5. To perform service maintenance of three (3) DICOM gateway server in RIPASH, PMMPMHAMBH and SSBH6) To apply preventive maintenance sticker/tag after successful service.
6. Disposal of any used spare parts/pm kits, unless otherwise instructed.
7. Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.
8. Vendor(s) to submit maintenance checklist/details/scope specific for the equipment along with other necessary documents to join this tender.

SECTION 3
TENDER FORM

To:

TENDER REFERENCE NO: KK/64/2026/HTD

INVITATION TO TENDER
THE PROVISION OF MAINTENANCE SERVICES FOR GE HEALTHCARE, MAC 2000
AND MAC 5 ECGs IN MINISTRY OF HEALTH

SCOPE OF WORK	PRICE (B\$)
<p>The provision of maintenance services for:</p> <ol style="list-style-type: none"> 1. Eighty-Nine (89) units of GE MAC 2000 ECGs and its related accessories: <ol style="list-style-type: none"> i. BME29776 - BME29862, and BME19747 ii. 3 units DICOM gateway server 2. Eighty-Two (82) units of GE MAC 5 ECGs <ol style="list-style-type: none"> i. BME30414 - BME30416, BME31256 - BME31328 and BME31421 - BME31426 <p>Scope of Work:</p> <ol style="list-style-type: none"> 1. To perform PPM once a year 2. To conduct electrical safety test 3. To conduct performance verification using multiparameter patient simulator 4. To conduct DICOM gateway/ CPACS connectivity test 5. To perform service maintenance of three (3) DICOM gateway server in RIPASH, PMMPMHAMBH and SSBH6) To apply preventive maintenance sticker/tag after successful service. 6. Disposal of any used spare parts/pm kits, unless otherwise instructed. 7. Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment. 8. Vendor(s) to submit maintenance checklist/details/scope specific for the equipment along with other necessary documents to join this tender. 	
TOTAL PRICE (B\$)	