

# LAMPIRAN - A1

RUJ: BME/A2K/(1661)/2026/HTD (B4)

20/4/2026

ONE-OFF PREVENTIVE MAINTENANCE FOR ONE (1) UNIT OF OXYGEN GENERATOR SYSTEM BRAND: ERREDUE SPA, MODEL: GO6LABCT8, S/N: 1993, BME29416 AT ENVIRONMENT SECTION, DEPARTMENT OF SCIENTIFIC SERVICES.

## SCOPE OF WORK

One-off preventive maintenance service for oxygen generator system at Environment Section of Department of Scientific Services:

- Scope to follow preventive maintenance manual/recommendation specific for the equipment, which include the following where necessary:

(i) inspection

(ii) cleaning

(iii) installation of PM kits, consumables and spare parts listed in spareparts list (Lampiran A2)

(iv) alignment

(v) calibration

(vi) any other related preventive maintenance works required.

- Functional check post-maintenance.

- Apply preventive maintenance sticker/tag after successful service.

- Disposal of any used spare parts/pm kits, unless otherwise instructed.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.

- Vendor(s) to submit maintenance checklist/details/scope specific for the equipment along with other necessary documents to join this tender.

- Upon approval, the earliest the preventive maintenance work can be carried out on the equipment is one month before the next PM due date, and not anytime earlier.

## LAMPIRAN - A2

RUJ: BME/A2K/(1661)/2026/HTD (B4)

ONE-OFF PREVENTIVE MAINTENANCE FOR ONE (1) UNIT OF OXYGEN GENERATOR SYSTEM BRAND: ERREDUE SPA, MODEL: GO6LABCT8, S/N: 1993, BME29416 AT ENVIRONMENT SECTION, DEPARTMENT OF SCIENTIFIC SERVICES.

	<b>SPAREPART</b>	<b>PART NUMBER</b>	<b>QUANTITY</b>
1	FILTER CARTRIDGE FOR PRD004_SMC	FCF237	2
2	SAFETY VALVE 10 BAR 1/4 ONU + INX AIRTEK_GHLAB	VZZ099	1
3	KIT342 + KIT348 - COMPRESSOR MAINTENANCE KIT FOR DOF250_LA		2