

LAMPIRAN - A1

RUJ: BME/A2K/(1677)/2026/HTD (B8)

5/5/2026

TO REPAIR ONE (1) UNIT OF R.O. WATER TREATMENT SYSTEM, BRAND : COREMED, MODEL : MX-X-VIII AT RIMBA DIALYSIS CENTRE

SCOPE OF WORK

To repair one (1) unit of R.O. Water Treatment System, Brand: Coremed, Model: MX-X-VIII at Rimba Dialysis Centre. Detail are as follow:

1. BME6905, Brand: Coremed, Model: MX-X-VIII, SN: 200901-0320, 200901-0321, 200901-0012, at Rimpa Dialysis Centre, W/order 2026-12175

- Scope of work to follow equipment manual/recommendation specific for this system which include the following where if necessary:

(i) installation of items listed in spareparts list (Lampiran A2)

(ii) cleaning

(iii) alignment

(iv) calibration

(v) optimisation

(vi) any other related maintenance works required after installation of items listed in spareparts list (Lampiran A2)

- Perform functional test, verification and performance check post-maintenance.

- Disposal of any used spare parts, unless otherwise instructed.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.

LAMPIRAN - A2

RUJ: BME/A2K/(1677)/2026/HTD (B8)

TO REPAIR ONE (1) UNIT OF R.O. WATER TREATMENT SYSTEM, BRAND : COREMED, MODEL : MX-X-VIII AT RIMBA DIALYSIS CENTRE

	SPAREPART	PART NUMBER	QUANTITY
1	UV LIGHT COMPLETE SET	ATC-UV0001	1
2	SOLENOID VALVE 1/2"	ATC-VF0048-RO	2
3	S/S BALL VALVE 1/2"	ATC-VF0028-RO	1
4	MECHANICAL SEAL RO PUMP KIT SHAFT SEAL	ATC-EL 0218	1
5	S.STEEL BOLT 12M X 40MM		40
6	S.STEEL BOLT, NUT AND WASHER 10M X 40MM		20