

**TENDER REFERENCE NO.: KK/121/2026/UPP**

**MINISTRY OF HEALTH  
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND  
COMMISSIONING OF TWO (2) UNITS OF CLINICAL  
MICROSCOPES FOR PATHOLOGY DEPARTMENT, RAJA  
ISTERI PENGIRAN ANAK SALEHA HOSPITAL, MINISTRY  
OF HEALTH**

**TENDER FEES : \$30.00**

**RECEIPT NO. :**

**CLOSING DATE : ON Tuesday, 16th June 2026**

**TIME : 2.00 PM**

**FOA :**

**THE CHAIRMAN  
MINI TENDER BOARD, TENDER BOX  
GROUND FLOOR, MINISTRY OF HEALTH  
COMMONWEALTH DRIVE  
BANDAR SERI BEGAWAN BB3910  
NEGARA BRUNEI DARUSSALAM**

**(CLUSTERING)**

**SECTION 2**

**SPECIFICATIONS AND REQUIREMENTS**

**TENDER REFERENCE NO: KK/121/2026/UPP**

**INVITATION TO TENDER**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF TWO (2) UNITS OF CLINICAL MICROSCOPES FOR PATHOLOGY DEPARTMENT, RAJA ISTERI PENGIRAN ANAK SALEHA HOSPITAL, MINISTRY OF HEALTH**

<b>NO.</b>	<b>ITEMS DESCRIPTIONS AND SPECIFICATIONS</b>	<b>PACKING SIZE</b>	<b>TOTAL QUANTITY</b>
1	CLINICAL MICROSCOPY COMPLETE WITH ACCESSORIES AND DESKTOP	UNIT	2 UNITS

**SCOPE OF WORK**

- 1. Supply of TWO (2) units of Clinical Microscopes, Pathology Department, RIPAS Hospital**

SECTION 1 – USER REQUIREMENTS				
Please <input checked="" type="checkbox"/> Tick where appropriate		Yes	No	Remarks
1	<b>CLINICAL MICROSCOPES</b>			
1.1	A <b>clinical upright microscope</b> suitable for routine and advanced <b>pathology applications</b> , including Histopathology and Cytopathology.			
1.2	The microscope shall be constructed from robust, chemical-resistant materials with anti-fungal optical coating.			
1.3	<b>Observation tubes</b>			
1.3.1	Binocular: Interchangeable, ergonomic, with dual/multi-observer options			
1.3.2	Eyepieces: Widefield 10x with diopter adjustment			
1.3.3	Interpupillary distance adjustment: In the range of 55mm to 75mm or equivalent or better			
1.4	<b>Nosepiece configuration: Encoded, 6x or better</b>			
1.4.1	Objectives magnification: 1x-100x or better			
1.4.2	Objectives time: <b>Infinity color corrected system (ICS) including EC Plan-Neofluar</b>			
1.4.3	Objective set up requirement: <ul style="list-style-type: none"> <li>- EC Plan-Neofluar 2.5x/0.085 WD=8.8mm or equivalent</li> <li>- EC Plan-Neofluar 40x/0.75 WD=0.71mm or equivalent</li> <li>- EC Plan-Neofluar 100x/1.3 WD=0.2mm or equivalent</li> <li>- EC Plan-Neofluar 5x/0.16 WD=12.1mm or equivalent</li> <li>- EC Plan-Neofluar 10x/0.3 WD=2.0mm or equivalent</li> <li>- EC Plan-Neofluar 20x/0.8 or equivalent</li> </ul>			
1.5	<b>Contrast Methods:</b> <ul style="list-style-type: none"> <li>- Brightfield</li> <li>- Darkfield</li> <li>- Fluorescence</li> <li>- Phase</li> <li>- Polarization</li> </ul>			
1.5.1	<b>Comes with blue filter</b>			
1.6	<b>Illumination</b>			
1.6.1	Condenser type: Abbe condenser or better			
1.6.2	Numerical aperture: 0.9-1.25 or better			
1.6.3	Light source: LED			
1.6.4	Compatible with brightfield, phase contrast, polarization and other contrast methods			
1.7	<b>Focus Mechanism:</b> Moving stage, manual or better.			

SECTION 1 – USER REQUIREMENTS				
Please <input checked="" type="checkbox"/> Tick where appropriate		Yes	No	Remarks
1.7.1	Capable to do coarse and fine adjustment			
1.8	<b>Stage:</b> Manual Motorized mechanical type with hardcoat anodised surface or better			
1.8.1	Motion: - X-Y Direction - Low-stage handle position - Extendable			
1.8.2	Comes with tension adjustment			
1.8.3	Smooth, precise control via ergonomic right-hand drive			
1.9	Digital Imaging (To be supplied with each microscope)			
1.9.1	Microscope shall be compatible with digital microscope cameras for documentation			
1.9.2	Preferably supports standalone imaging (capture without PC) when paired with a dedicated camera			
1.9.3	Automatic camera exposure, white balance, and contrast optimization preferred			
1.9.4	Digital imaging software shall be upgradable when required			
1.10	Safety compliance: Equipment shall comply with electrical and safety standards for laboratory equipment			
1.11	<b><u>Two units of complete PC system shall be supplied with the following minimum requirements.</u></b>			
1.11.1	<b>Each PC system shall be connected to its respective microscopes.</b>			
1.11.2	Minimum Operating system: Microsoft Window 10 pro 64 bit or newer, upgradable as necessary, compatible with the offered workstation software.			
1.11.3	Enterprise Grade with Intel core i5 processor or better			
1.11.4	16GB DDR4 RAM or higher			
1.11.5	Hard drive capacity of 1TB or higher			
1.11.6	Appropriate video graphic card to support software visuals			
1.11.7	Connectivity through Ethernet, Wireless-N and Bluetooth 4.0			
1.11.8	Various ports such as USB 3.0, Display Port, RJ-45 and all the necessary ports for it to work as intended			
1.11.9	Security features: Antivirus, Trusted Platform Module 2.0 and remote support software.			
1.11.10	Inclusive of Display monitor suitable for working station			
1.11.11	Working System shall come with all the necessary standard accessories not mention above.			
1.12	<b>Accessories/Consumables</b> The Tenderer shall supply, but not be limited to, the accessories and consumables listed below, as well as any additional items required to ensure full functionality of the machine			

Please <input checked="" type="checkbox"/> Tick where appropriate		Yes	No	Remarks
<b>2</b>	<b>END-USER TRAINING</b>			
2.1	Conduct <b>user training</b> to the all-end users by an application specialist or competent local engineer including but not limited to: <ul style="list-style-type: none"> <li>▪ Basic user operation, user troubleshooting and user maintenance</li> <li>▪ Provide Operating manual (Hardcopy and/or Softcopy)</li> </ul>			
2.2	Tenderer must <b>prepare a training attendance or proof of training done to end user during commissioning and the refresher course (6) months after commissioning.</b>			
<b>3</b>	<b>TECHNICAL TRAINING</b>			
3.1	<b>Introductory Technical Training</b> to Biomedical Engineers and Technicians at BME Office by competent Tenderer's Engineer/Technicians that includes but not limited to: <ul style="list-style-type: none"> <li>▪ Troubleshooting and basic corrective maintenance</li> <li>▪ Handling and basic inspection maintenance</li> </ul> *(Two sessions/groups if required)			
<b>4</b>	<b>WARRANTY</b>			
4.1	Tenderer to include warranty period of <b>at least two (2) years</b>			
4.2	Tenderers to <b>ACKNOWLEDGE</b> the Warranty Undertaking Form in Section 4 stating the terms of warranty provided for the equipment in the tender for the period of two years. This includes but not limited to: <ul style="list-style-type: none"> <li>▪ Scope of Warranty</li> <li>▪ Planned Preventive Maintenance during warranty (one of which includes battery replacement at the end of warranty period).</li> </ul>			

**SECTION 2 – PRICE PROPOSAL**

<b>PURCHASE PRICE</b>	<b>CLINICAL MICROSCOPES</b>
	<b>TOTAL PRICE</b>

**SECTION 3 - PROCUREMENT AND TECHNICAL SPECIFICATION**

<b>BRAND:</b>	<b>MODEL:</b>
<b>COUNTRY OF ORIGIN:</b>	<b>YEAR INTRODUCED TO MARKET:</b>
<b>WARRANTY PERIOD:</b>	<b>LAST COUNTRY SOLD TO:</b>
<b>PRICE VALIDITY: [AT LEAST <u>ONE (1) YEAR</u> PRICE VALIDTY]</b>	<b>DELIVERY TIME:</b>

**SECTION 3 - PROCUREMENT AND TECHNICAL SPECIFICATION**

**AUTHORIZED DISTRIBUTOR:  
(AUTHORIZED DISTRIBUTOR LETTER ATTACHED)**

**DETAILED BROCHURE INCLUDED**

**USER AND SERVICE MANUALS:**

**MAINS POWER SUPPLY:**

**BATTERY:**

**POWER ADAPTER/CHARGER OUTPUT RATING:**

**EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:**

**INTERNATIONAL SAFETY STANDARD**

Must comply to at least 1 safety Standards and certification  
(Please attached the copy of stated standards and certifications)

**NUMBER OF TECHNICAL SUPPORT (ENGINEER/TECHNICIAN)**

Please provide training or certification for locals who is trained/certified

**DIMENSIONS AND WEIGHT OF MAIN UNIT:**

**EQUIPMENT WHOLE LIFE TIME SUPPORT:**

## SECTION 4 – WARRANTY UNDERTAKING FORM

Tenderer, on behalf of the manufacturer, acknowledged and agrees that when equipment is under the warranty period, must cover the scope of normal warranty below at no additional cost:

### NORMAL WARRANTY

- Warrants the supplied medical equipment and its accessories to be in good condition, in working order and free from defects to the extent such equipment do not comply with specifications, under normal use for the warranty period. The scope of warranty covers to its maximum extent permitted by applicable law.
- During warranty, tenderer must rectify issues arise from any mechanical, technical or software faulty as soon as it is reported.
- **Exchange warranty;** Providing replacement units:
  - A. Warranty against defects – Manufacturing defects or Equipment malfunction resulted from mechanical, electrical or software failure during Commissioning or within the first \_\_\_\_\_ months of use
  - B. Faulty workmanship or unsatisfactory condition during delivery or commissioning
  - C. If a unit or accessory is deemed used item or refurbished item (not a new unit) by the user and BME Unit.
- **Two time Planned Preventive Maintenance (PPM) PER YEAR** according to Manufacturer's Preventive Maintenance Guideline and to include one-time replacements of battery at the end of 2 years warranty period or any other relevant parts to prolong equipment lifespan.

### EXCLUSION FROM WARRANTY

MOH understand that the following circumstances are not covered in the warranty and Tenderer may quote for repair and subject to MOH approval:

- Unauthorized modifications - an alteration or repair by anyone other than the Manufacturer or Authorized agent during warranty period.
- Accidental damage or problems caused by negligence or mishandling, subject to appropriate justification by both parties.
- Vandalism and Natural disasters
- Normal wear and tear

### ANY OTHER EXCLUSION

Tenderer may propose below to include items or terms which is not listed in the exclusion list above for MOH consideration.

NO.	TERMS AND CONDITIONS
1	Tenderer must be registered with the Ministry of Health.
2	<b>TENDER FORM should be filled</b> completely including the <b>USER REQUIREMENT FORM</b> (if available). Submission of incomplete form <b>MAY</b> cause <b>DISQUALIFICATION OF TENDER.</b>
3	Each tenderer is allowed to quote <b>ONE BRAND WITH ONE PRICE ONLY</b> for each item. Submission of more than one brand and price will cause <b>DISQUALIFICATION OF TENDER.</b>
4	All consumables supplied throughout this tender <u>shall</u> have a minimum expiry date of <b>twelve (12) months / on delivery</b> (if applicable). Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made (if applicable).
5	Brochures / catalogues should be submitted / attached with tender document.
6	Any room renovation which may be required, it is mandatory to conduct site visit (if applicable)
7	Samples should be submitted together with tender or within fourteen (14 days) of the tender closing dates (if applicable).
8	<b>DELIVERY PERIOD: NOT MORE THAN 4 WEEKS STAGGERED DELIVERY UPON REQUEST</b>
9	<b>PRICE VALIDITY:</b> The quotation shall remain valid for <b>12 MONTHS</b> from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).
10	The equipment supplied must be newly manufactured, unused, and in its original, sealed packaging. The equipment must not be previously owned, refurbished, or reconditioned in any form. During delivery, the vendor is required to provide proof of manufacture date confirming the equipment is new.

SECTION 3

TENDER FORM

To:

TENDER REFERENCE NO: KK/121/2026/UPP

INVITATION TO TENDER

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF TWO (2) UNITS OF CLINICAL MICROSCOPES FOR PATHOLOGY DEPARTMENT, RAJA ISTERI PENGIRAN ANAK SALEHA HOSPITAL, MINISTRY OF HEALTH

---

TENDER OF (*name of tenderer*) : \_\_\_\_\_

Company/Business Registration No. : \_\_\_\_\_

Tender Closing Date : \_\_\_\_\_

USER'S REQUIREMENTS				VENDOR'S OFFER					
NO.	ITEMS DESCRIPTIONS AND SPECIFICATIONS	PACKING SIZE	TOTAL QUANTITY	ITEMS DESCRIPTIONS AND SPECIFICATIONS	PART/CATALOGUE NUMBER AND BRAND	PACKING SIZE	TOTAL QUANTITY	COST PER PACK (COST PER UNIT) (B\$)	TOTAL COSTS (B\$)
1	CLINICAL MICROSCOPY COMPLETE WITH ACCESSORIES AND DESKTOP	UNIT	2 UNITS						
TOTAL COST (B\$)									

SCOPE OF WORK			
Please <input checked="" type="checkbox"/> Tick where appropriate	Yes	No	Remarks
1. Supply of TWO (2) <u>units</u> of Clinical Microscopes, Pathology Department, RIPAS Hospital			

SECTION 1 – USER REQUIREMENTS			
Please <input checked="" type="checkbox"/> Tick where appropriate	Yes	No	Remarks
<b>1</b>	<b>CLINICAL MICROSCOPES</b>		
1.1	A <b>clinical upright microscope</b> suitable for routine and advanced <b>pathology applications</b> , including Histopathology and Cytopathology.		
1.2	The microscope shall be constructed from robust, chemical-resistant materials with anti-fungal optical coating.		
1.3	<b>Observation tubes</b>		
1.3.1	Binocular: Interchangeable, ergonomic, with dual/multi-observer options		
1.3.2	<i>Eyepieces: Widefield 10x with diopter adjustment</i>		
1.3.3	Interpupillary distance adjustment: In the range of 55mm to 75mm or equivalent or better		
1.4	<b>Nosepiece configuration: Encoded, 6x or better</b>		
1.4.1	Objectives magnification: 1x-100x or better		
1.4.2	Objectives time: <b>Infinity color corrected system (ICS) including EC Plan-Neofluar</b>		
1.4.3	Objective set up requirement: <ul style="list-style-type: none"> <li>- EC Plan-Neofluar 2.5x/0.085 WD=8.8mm or equivalent</li> <li>- EC Plan-Neofluar 40x/0.75 WD=0.71mm or equivalent</li> <li>- EC Plan-Neofluar 100x/1.3 WD=0.2mm or equivalent</li> <li>- EC Plan-Neofluar 5x/0.16 WD=12.1mm or equivalent</li> <li>- EC Plan-Neofluar 10x/0.3 WD=2.0mm or equivalent</li> <li>- EC Plan-Neofluar 20x/0.8 or equivalent</li> </ul>		
1.5	<b>Contrast Methods:</b> <ul style="list-style-type: none"> <li>- Brightfield</li> <li>- Darkfield</li> <li>- Fluorescence</li> <li>- Phase</li> <li>- Polarization</li> </ul>		
1.5.1	<b>Comes with blue filter</b>		
1.6	<b>Illumination</b>		

SECTION 1 – USER REQUIREMENTS				
Please <input checked="" type="checkbox"/> Tick where appropriate		Yes	No	Remarks
1.6.1	Condenser type: Abbe condenser or better			
1.6.2	Numerical aperture: 0.9-1.25 or better			
1.6.3	Light source: LED			
1.6.4	Compatible with brightfield, phase contrast, polarization and other contrast methods			
1.7	<b>Focus Mechanism:</b> Moving stage, manual or better.			
1.7.1	Capable to do coarse and fine adjustment			
1.8	<b>Stage:</b> Manual Motorized mechanical type with hardcoat anodised surface or better			
1.8.1	Motion: - X-Y Direction - Low-stage handle position - Extendable			
1.8.2	Comes with tension adjustment			
1.8.3	Smooth, precise control via ergonomic right-hand drive			
1.9	Digital Imaging (To be supplied with each microscope)			
1.9.1	Microscope shall be compatible with digital microscope cameras for documentation			
1.9.2	Preferably supports standalone imaging (capture without PC) when paired with a dedicated camera			
1.9.3	Automatic camera exposure, white balance, and contrast optimization preferred			
1.9.4	Digital imaging software shall be upgradable when required			
1.10	Safety compliance: Equipment shall comply with electrical and safety standards for laboratory equipment			
1.11	<b><u>Two units of complete PC system shall be supplied with the following minimum requirements.</u></b>			
1.11.1	<b>Each PC system shall be connected to its respective microscopes.</b>			
1.11.2	Minimum Operating system: Microsoft Window 10 pro 64 bit or newer, upgradable as necessary, compatible with the offered workstation software.			
1.11.3	Enterprise Grade with Intel core i5 processor or better			
1.11.4	16GB DDR4 RAM or higher			
1.11.5	Hard drive capacity of 1TB or higher			
1.11.6	Appropriate video graphic card to support software visuals			
1.11.7	Connectivity through Ethernet, Wireless-N and Bluetooth 4.0			
1.11.8	Various ports such as USB 3.0, Display Port, RJ-45 and all the necessary ports for it to work as			

SECTION 1 – USER REQUIREMENTS				
Please <input checked="" type="checkbox"/> Tick where appropriate		Yes	No	Remarks
	intended			
1.11.9	Security features: Antivirus, Trusted Platform Module 2.0 and remote support software.			
1.11.10	Inclusive of Display monitor suitable for working station			
1.11.11	Working System shall come with all the necessary standard accessories not mention above.			
1.12	<b>Accessories/Consumables</b> The Tenderer shall supply, but not be limited to, the accessories and consumables listed below, as well as any additional items required to ensure full functionality of the machine			

Please <input checked="" type="checkbox"/> Tick where appropriate		Yes	No	Remarks
<b>2</b>	<b>END-USER TRAINING</b>			
2.1	Conduct <b>user training</b> to the all-end users by an application specialist or competent local engineer including but not limited to: <ul style="list-style-type: none"> <li>▪ Basic user operation, user troubleshooting and user maintenance</li> <li>▪ Provide Operating manual (Hardcopy and/or Softcopy)</li> </ul>			
2.2	Tenderer must <b>prepare a training attendance or proof of training done to end user during commissioning and the refresher course (6) months after commissioning.</b>			
<b>3</b>	<b>TECHNICAL TRAINING</b>			
3.1	<b>Introductory Technical Training</b> to Biomedical Engineers and Technicians at BME Office by competent Tenderer's Engineer/Technicians that includes but not limited to: <ul style="list-style-type: none"> <li>▪ Troubleshooting and basic corrective maintenance</li> <li>▪ Handling and basic inspection maintenance</li> </ul> *(Two sessions/groups if required)			
<b>4</b>	<b>WARRANTY</b>			
4.1	Tenderer to include warranty period of <b>at least two (2) years</b>			
4.2	Tenderers to <b>ACKNOWLEDGE</b> the Warranty Undertaking Form in Section 4 stating the terms of warranty provided for the equipment in the tender for the period of two years. This includes but not limited to: <ul style="list-style-type: none"> <li>▪ Scope of Warranty</li> <li>▪ Planned Preventive Maintenance during warranty (one of which includes battery replacement at the end of warranty period).</li> </ul>			

SECTION 1 – USER REQUIREMENTS				
Please <input checked="" type="checkbox"/> Tick where appropriate		Yes	No	Remarks
1	<b>CLINICAL MICROSCOPES</b>			
1.1	A <b>clinical upright microscope</b> suitable for routine and advanced <b>pathology applications</b> , including Histopathology and Cytopathology.			
1.2	The microscope shall be constructed from robust, chemical-resistant materials with anti-fungal optical coating.			
1.3	<b>Observation tubes</b>			
1.3.1	Binocular: Interchangeable, ergonomic, with dual/multi-observer options			
1.3.2	Eyepieces: Widefield 10x with diopter adjustment			
1.3.3	Interpupillary distance adjustment: In the range of 55mm to 75mm or equivalent or better			
1.4	<b>Nosepiece configuration: Encoded, 6x or better</b>			
1.4.1	Objectives magnification: 1x-100x or better			
1.4.2	Objectives time: <b>Infinity color corrected system (ICS) including EC Plan-Neofluar</b>			
1.4.3	Objective set up requirement: <ul style="list-style-type: none"> <li>- EC Plan-Neofluar 2.5x/0.085 WD=8.8mm or equivalent</li> <li>- EC Plan-Neofluar 40x/0.75 WD=0.71mm or equivalent</li> <li>- EC Plan-Neofluar 100x/1.3 WD=0.2mm or equivalent</li> <li>- EC Plan-Neofluar 5x/0.16 WD=12.1mm or equivalent</li> <li>- EC Plan-Neofluar 10x/0.3 WD=2.0mm or equivalent</li> <li>- EC Plan-Neofluar 20x/0.8 or equivalent</li> </ul>			
1.5	<b>Contrast Methods:</b> <ul style="list-style-type: none"> <li>- Brightfield</li> <li>- Darkfield</li> <li>- Fluorescence</li> <li>- Phase</li> <li>- Polarization</li> </ul>			
1.5.1	<b>Comes with blue filter</b>			
1.6	<b>Illumination</b>			
1.6.1	Condenser type: Abbe condenser or better			
1.6.2	Numerical aperture: 0.9-1.25 or better			
1.6.3	Light source: LED			
1.6.4	Compatible with brightfield, phase contrast, polarization and other contrast methods			
1.7	<b>Focus Mechanism:</b> Moving stage, manual or better.			

SECTION 1 – USER REQUIREMENTS				
Please <input checked="" type="checkbox"/> Tick where appropriate		Yes	No	Remarks
1.7.1	Capable to do coarse and fine adjustment			
1.8	<b>Stage:</b> Manual Motorized mechanical type with hardcoat anodised surface or better			
1.8.1	Motion: - X-Y Direction - Low-stage handle position - Extendable			
1.8.2	Comes with tension adjustment			
1.8.3	Smooth, precise control via ergonomic right-hand drive			
1.9	Digital Imaging (To be supplied with each microscope)			
1.9.1	Microscope shall be compatible with digital microscope cameras for documentation			
1.9.2	Preferably supports standalone imaging (capture without PC) when paired with a dedicated camera			
1.9.3	Automatic camera exposure, white balance, and contrast optimization preferred			
1.9.4	Digital imaging software shall be upgradable when required			
1.10	Safety compliance: Equipment shall comply with electrical and safety standards for laboratory equipment			
1.11	<b>Two units of complete PC system shall be supplied with the following minimum requirements.</b>			
1.11.1	<b>Each PC system shall be connected to its respective microscopes.</b>			
1.11.2	Minimum Operating system: Microsoft Window 10 pro 64 bit or newer, upgradable as necessary, compatible with the offered workstation software.			
1.11.3	Enterprise Grade with Intel core i5 processor or better			
1.11.4	16GB DDR4 RAM or higher			
1.11.5	Hard drive capacity of 1TB or higher			
1.11.6	Appropriate video graphic card to support software visuals			
1.11.7	Connectivity through Ethernet, Wireless-N and Bluetooth 4.0			
1.11.8	Various ports such as USB 3.0, Display Port, RJ-45 and all the necessary ports for it to work as intended			
1.11.9	Security features: Antivirus, Trusted Platform Module 2.0 and remote support software.			
1.11.10	Inclusive of Display monitor suitable for working station			
1.11.11	Working System shall come with all the necessary standard accessories not mention above.			
1.12	<b>Accessories/Consumables</b> The Tenderer shall supply, but not be limited to, the accessories and consumables listed below, as well as any additional items required to ensure full functionality of the machine			

Please <input checked="" type="checkbox"/> Tick where appropriate		Yes	No	Remarks
<b>2</b>	<b>END-USER TRAINING</b>			
2.1	Conduct <b>user training</b> to the all-end users by an application specialist or competent local engineer including but not limited to: <ul style="list-style-type: none"> <li>▪ Basic user operation, user troubleshooting and user maintenance</li> <li>▪ Provide Operating manual (Hardcopy and/or Softcopy)</li> </ul>			
2.2	Tenderer must <b>prepare a training attendance or proof of training done to end user during commissioning and the refresher course (6) months after commissioning.</b>			
<b>3</b>	<b>TECHNICAL TRAINING</b>			
3.1	<b>Introductory Technical Training</b> to Biomedical Engineers and Technicians at BME Office by competent Tenderer's Engineer/Technicians that includes but not limited to: <ul style="list-style-type: none"> <li>▪ Troubleshooting and basic corrective maintenance</li> <li>▪ Handling and basic inspection maintenance</li> </ul> *(Two sessions/groups if required)			
<b>4</b>	<b>WARRANTY</b>			
4.1	Tenderer to include warranty period of <b>at least two (2) years</b>			
4.2	Tenderers to <b>ACKNOWLEDGE</b> the Warranty Undertaking Form in Section 4 stating the terms of warranty provided for the equipment in the tender for the period of two years. This includes but not limited to: <ul style="list-style-type: none"> <li>▪ Scope of Warranty</li> <li>▪ Planned Preventive Maintenance during warranty (one of which includes battery replacement at the end of warranty period).</li> </ul>			

SECTION 2 – PRICE PROPOSAL		
PURCHASE PRICE	CLINICAL MICROSCOPE PER UNIT	BND\$
	TOTAL	BND\$

SECTION 3 - PROCUREMENT AND TECHNICAL SPECIFICATION			
BRAND:		MODEL:	
COUNTRY OF ORIGIN:		YEAR INTRODUCED TO MARKET:	
WARRANTY PERIOD:		LAST COUNTRY SOLD TO:	
PRICE VALIDITY: [AT LEAST <u>ONE (1) YEAR</u> PRICE VALIDTY]		DELIVERY TIME:	

SECTION 3 - PROCUREMENT AND TECHNICAL SPECIFICATION						
AUTHORIZED DISTRIBUTOR: (AUTHORIZED DISTRIBUTOR LETTER ATTACHED)	APPOINTED BRUNEI DISTRIBUTOR					
		PROCURE FROM OVERSEA AUTHORIZED DISTRIBUTOR	COMPANY NAME:			
			COMPANY ORIGIN:			
DETAILED BROCHURE INCLUDED	YES		NO	<input checked="" type="checkbox"/> or specify where appropriate		
USER AND SERVICE MANUALS:	YES		NO	Tenderers to acknowledge that they must provide at least <b>TWO</b> sets of <b>USER AND SERVICE</b> manuals when applying commissioning form. One Set for End User, One Set for BME. (Please provide hardcopy or softcopy)		
MAINS POWER SUPPLY:			OTHERS:			
			OTHERS:			
BATTERY	RECHARGEABLE		SINGLE-USE		REPLACEABLE	
	OTHERS:					
	TYPE OF BATTERY:					
	RATING:					
POWER ADAPTER/CHARGER OUTPUT RATING:						
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:						
INTERNATIONAL SAFETY STANDARD Must comply to at least 1 safety Standards and certification (Please attached the copy of stated standards and certifications)				<input checked="" type="checkbox"/> Tick where appropriate <input type="checkbox"/> US FDA Standard, <input type="checkbox"/> European Union CE MARK, <input type="checkbox"/> Australian TGA Standard, <input type="checkbox"/> Canadian CSA Standard or <input type="checkbox"/> Japanese JIS Standard. Others (Please specify):		
NUMBER OF TECHNICAL SUPPORT (ENGINEER/TECHNICIAN)  Please provide training or certification for locals who is trained/certified	LOCAL		<input type="checkbox"/> Trained / Certified <input type="checkbox"/> Not yet trained on the product			
	OVERSEA (SPECIFY LOCATION)		NEAREST LOCATION:			
DIMENSIONS AND WEIGHT OF MAIN UNIT:			<input type="checkbox"/> mm <input type="checkbox"/> cm <input type="checkbox"/> inch	<input type="checkbox"/> Kilogram (Kg) <input type="checkbox"/> Gram(g) <input type="checkbox"/> Pound (lbs)		
EQUIPMENT WHOLE LIFE TIME SUPPORT:	The supplier shall ensure that spare parts for the equipment are available for a <b>minimum of 10 years</b> after installation, with the support period extending beyond the expected lifecycle of the equipment. No of years: _____ (Please specify)					

**SECTION 4 – WARRANTY UNDERTAKING FORM**

Tenderer, on behalf of the manufacturer, acknowledged and agrees that when equipment is under the warranty period, must cover the scope of normal warranty below at no additional cost:

**NORMAL WARRANTY**

- Warrants the supplied medical equipment and its accessories to be in good condition, in working order and free from defects to the extent such equipment do not comply with specifications, under normal use for the warranty period. The scope of warranty covers to its maximum extent permitted by applicable law.
- During warranty, tenderer must rectify issues arise from any mechanical, technical or software faulty as soon as it is reported.
- **Exchange warranty**; Providing replacement units:
  - A. Warranty against defects – Manufacturing defects or Equipment malfunction resulted from mechanical, electrical or software failure during Commissioning or within the first \_\_\_\_\_ months of use
  - B. Faulty workmanship or unsatisfactory condition during delivery or commissioning
  - C. If a unit or accessory is deemed used item or refurbished item (not a new unit) by the user and BME Unit.
- **Two time Planned Preventive Maintenance (PPM) PER YEAR** according to Manufacturer's Preventive Maintenance Guideline and to include one-time replacements of battery at the end of 2 years warranty period or any other relevant parts to prolong equipment lifespan.

**EXCLUSION FROM WARRANTY**

MOH understand that the following circumstances are not covered in the warranty and Tenderer may quote for repair and subject to MOH approval:

- Unauthorized modifications - an alteration or repair by anyone other than the Manufacturer or Authorized agent during warranty period.
- Accidental damage or problems caused by negligence or mishandling, subject to appropriate justification by both parties.
- Vandalism and Natural disasters
- Normal wear and tear

**ANY OTHER EXCLUSION**

Tenderer may propose below to include items or terms which is not listed in the exclusion list above for MOH consideration.

**TENDERER ACKNOWLEDGMENT**

**COMPANY CHOP AND SIGNATURE**

NO.	TERMS AND CONDITIONS	VENDOR'S OFFER (PLEASE STATE)
1	Tenderer must be registered with the Ministry of Health.	
2	<b>TENDER FORM should be filled</b> completely including the <b>USER REQUIREMENT FORM</b> (if available). Submission of incomplete form <b><u>MAY</u></b> cause <b>DISQUALIFICATION OF TENDER.</b>	
3	Each tenderer is allowed to quote <b>ONE BRAND WITH ONE PRICE ONLY</b> for each item. Submission of more than one brand and price will cause <b>DISQUALIFICATION OF TENDER.</b>	
4	All consumables supplied throughout this tender <u>shall</u> have a minimum expiry date of <b>twelve (12) months / on delivery</b> (if applicable). Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made (if applicable).	
5	<b>Brochures / catalogues should be submitted / attached</b> with tender document.	
6	Any <b>room renovation</b> which may be required, <b>it is mandatory to conduct site visit</b> (if applicable)	
7	<b>Samples should be submitted together with tender or within fourteen (14 days)</b> of the tender closing dates (if applicable).	
8	<b>DELIVERY PERIOD:</b> (Please state) Not More Than <b>90 days</b> upon confirmation	
9	<b>PRICE VALIDITY:</b> The quotation shall remain valid for <b>12 MONTHS</b> from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).	
10	The equipment supplied must be <b>newly manufactured, unused, and in its original, sealed packaging.</b> The equipment must not be previously owned, refurbished, or reconditioned in any form. During delivery, the vendor is required to provide proof of manufacture date confirming the equipment is new.	

1. We offer and undertake on your acceptance of our Tender to provide the above mentioned services in accordance with your Invitation To Tender.
2. Our Tender is fully consistent with and does not contradict or derogate from anything in your Invitation To Tender. We have not qualified or changed any of the provisions of your Invitation To Tender.
3. OUR OFFER IS VALID FOR **TWELVE (12)** CALENDAR MONTHS FROM THE TENDER CLOSING DATE.
4. When requested by you, we shall extend the validity of this offer.
5. We further undertake to give you any further information which you may require.

Dated this \_\_\_\_\_ day of \_\_\_\_\_, \_\_\_\_\_

\_\_\_\_\_  
**Signature of authorised officer of Tenderer**

Name:

Designation:

Tenderer's official stamp