

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, ONE HUNDRE SIXTY-NINE (169) UNITS OF CARDIAC/OVERBED TABLE FOR DEPARTMENT OF RENAL SERVICES, MINISTRY OF HEALTH

No.	DISTRIBUTION	UNIT
1.	RENAL UNIT RIPAS HOSPITAL	24
2.	RIMBA DIALYSIS CENTRE	55
3.	KIARONG DIALYSIS CENTRE	10
4.	TUTONG DIALYSIS CENTRE	20
5.	KUALA BELAIT DIALYSIS CENTRE & WARD 1 (SURI SERI BEGAWAN HOSPITAL)	40
6.	RENAL UNIT PIHM TEMBURONG HOSPITAL	20

SECTION I - USER REQUIREMENT	
STANDARD FEATURES:	<ul style="list-style-type: none"> • Shall be designed for use by dialysis patients during treatment • Provide stable, height adjustment surface for supporting trays, food or light personal equipment. • Must be durable, easily maneuverable and constructed using high-quality hospital-grade materials • The table shall be of manual operation type with a smooth, effortless height adjustment mechanism. • Construction materials shall be non-corrosive, rust-proof, and resistant to hospital-grade disinfectants. • The design shall allow easy positioning around dialysis chairs or beds without obstructing patient access. • The table shall be stable, free from wobbling, and suitable for continuous daily use in a clinical environment. • The product must comply with relevant international medical furniture standards.
FUNCTIONAL REQUIREMENTS:	<ul style="list-style-type: none"> • The table shall allow comfortable use for patients seated or reclining during dialysis treatment. • The adjustment mechanism shall operate quietly and smoothly. • All moving parts shall be durable and easily serviceable. • The design shall prevent tipping or accidental collapse during use. • The supplier/contractor shall provide complete inventory registration and asset tagging for all supplied units during commissioning and handover.
ACCESSORIES / OPTIONAL FEATURES:	<ul style="list-style-type: none"> • Optional integrated cup holder, tablet/book stand, or tilt mechanism (up to 30°) may be provided if available. • Optional side handle or grip for easier movement may be included.

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, ONE HUNDRE SIXTY-NINE
(169) UNITS OF CARDIAC/OVERBED TABLE FOR DEPARTMENT OF RENAL SERVICES,
MINISTRY OF HEALTH**

	SECTION II - PROCUREMENT REQUIREMENT
MODEL AND BRAND	<ul style="list-style-type: none"> • State Model and Brand of Equipment.
COUNTRY OF ORIGIN	<ul style="list-style-type: none"> • State Country of Origin.
PRICE (CIF) B\$	<ul style="list-style-type: none"> • State Unit and Total Price Offered.
WHERE MARKETED	<ul style="list-style-type: none"> • Specify in which other countries the equipment marketed.
YEAR MANUFACTURED	<ul style="list-style-type: none"> • State year of equipment manufactured.
WARRANTY	<ul style="list-style-type: none"> • State number of years.
DELIVERY TIME AND INSTALLATION	<ul style="list-style-type: none"> • State number of weeks.
PRICE VALIDITY	<ul style="list-style-type: none"> • State number of weeks.

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, ONE HUNDRE SIXTY-NINE
(169) UNITS OF CARDIAC/OVERBED TABLE FOR DEPARTMENT OF RENAL SERVICES,
MINISTRY OF HEALTH**

	SECTION III - TECHNICAL REQUIREMENT
TYPE:	<ul style="list-style-type: none"> • Manual cardiac/overbed table.
HEIGHT ADJUSTMENT:	<ul style="list-style-type: none"> • Manual adjustment with gas spring or mechanical locking system.
HEIGHT RANGE:	<ul style="list-style-type: none"> • 700 mm to 1100 mm ($\pm 10\%$).
TABLE TOP SIZE:	<ul style="list-style-type: none"> • Approximately 900 mm (L) x 405 mm (W) or equivalent.
TABLE TOP MATERIAL:	<ul style="list-style-type: none"> • ABS plastic or laminated wood with rounded edges, smooth finish.
FRAME MATERIAL:	<ul style="list-style-type: none"> • Powder-coated mild steel or stainless steel.
COLUMN SUPPORT:	<ul style="list-style-type: none"> • Chrome-plated or stainless steel.
CASTORS:	<ul style="list-style-type: none"> • Four (4) swivel castors, minimum 50 mm diameter, with two (2) lockable wheels.
LOAD CAPACITY:	<ul style="list-style-type: none"> • Minimum 20 kg evenly distributed.
SURFACE FINISH:	<ul style="list-style-type: none"> • Smooth, easy to clean, resistant to disinfectants and body fluids.
MOBILITY:	<ul style="list-style-type: none"> • Smooth and stable movement without wobbling.
LOCKING MECHANISM:	<ul style="list-style-type: none"> • Secure system to fix height and position.
INTERNATIONAL SAFETY STANDARD:	<ul style="list-style-type: none"> • The equipment must comply to either one of the five international safety standards, namely; US FDA Standard, European Union CE MARK, Australian TGA Standard, Canadian CSA Standard or Japanese Standard.
BROUCHER:	<ul style="list-style-type: none"> • Tenderers must submit with their offer, a detailed broucher which supports the information the tenderer has entered in the "form-to-be-used (FORM B)" for reference.
USER MANUALS:	<ul style="list-style-type: none"> • Two hard copies and one soft copy [PDF FORMAT] to be submitted during commissioning.

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, ONE HUNDRE SIXTY-NINE
(169) UNITS OF CARDIAC/OVERBED TABLE FOR DEPARTMENT OF RENAL SERVICES,
MINISTRY OF HEALTH**

No.	DISTRIBUTION	UNIT
1.	RENAL UNIT RIPAS HOSPITAL	24
2.	RIMBA DIALYSIS CENTRE	55
3.	KIARONG DIALYSIS CENTRE	10
4.	TUTONG DIALYSIS CENTRE	20
5.	KUALA BELAIT DIALYSIS CENTRE & WARD 1 (SURI SERI BEGAWAN HOSPITAL)	40
6.	RENAL UNIT PIHM TEMBURONG HOSPITAL	20

	SECTION I - USER REQUIREMENT	YES	NO	PLEASE SPECIFY THE SPECIFICATION/ SCOPE OFFERED FAILURE TO FILL THE OFFER DETAIL SHALL BE IDENTIFIED AS NON- COMPLIED TO THE OFFER
STANDARD FEATURES:	• Shall be designed for use by dialysis patients during treatment			
	• Provide stable, height adjustment surface for supporting trays, food or light personal equipment.			
	• Must be durable, easily maneuverable and constructed using high-quality hospital-grade materials			
	• The table shall be of manual operation type with a smooth, effortless height adjustment mechanism.			
	• Construction materials shall be non-corrosive, rust-proof, and resistant to hospital-grade disinfectants.			
	• The design shall allow easy positioning around dialysis chairs or beds without obstructing patient access.			
	• The table shall be stable, free from wobbling , and suitable for continuous daily use in a clinical environment.			
	• The product must comply with relevant international medical furniture standards.			
FUNCTIONAL REQUIREMENTS:	• The table shall allow comfortable use for patients seated or reclining during dialysis treatment.			
	• The adjustment mechanism shall operate quietly and smoothly.			
	• All moving parts shall be durable and easily serviceable.			
	• The design shall prevent tipping or accidental collapse during use.			
	• The supplier/contractor shall provide complete inventory registration and asset tagging for all supplied units during commissioning and handover.			
ACCESSORIES / OPTIONAL FEATURES:	• Optional integrated cup holder, tablet/book stand, or tilt mechanism (up to 30°) may be provided if available.			
	• Optional side handle or grip for easier movement may be included.			

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, ONE HUNDRE SIXTY-NINE (169) UNITS OF CARDIAC/OVERBED
TABLE FOR DEPARTMENT OF RENAL SERVICES,
MINISTRY OF HEALTH**

	SECTION II - PROCUREMENT REQUIREMENT	
MODEL AND BRAND		
COUNTRY OF ORIGIN:		
PRICE (CIF) B\$	UNIT PRICE (B\$ CIF):	TOTAL PRICE (B\$ CIF):
WHERE MARKETED:		
YEAR MANUFACTURED:		
WARRANTY:		
DELIVERY TIME:		
PRICE VALIDITY:		

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING, ONE HUNDRE SIXTY-NINE
(169) UNITS OF CARDIAC/OVERBED TABLE FOR DEPARTMENT OF RENAL SERVICES,
MINISTRY OF HEALTH**

	SECTION III - TECHNICAL REQUIREMENTS
TYPE:	
HEIGHT ADJUSTMENT:	
HEIGHT RANGE:	
TABLE TOP SIZE:	
TABLE TOP MATERIAL:	
FRAME MATERIAL:	
COLUMN SUPPORT:	
CASTORS:	
LOAD CAPACITY:	
SURFACE FINISH:	
MOBILITY:	
LOCKING MECHANISM:	
INTERNATIONAL SAFETY STANDARD:	
BROUCHER:	
USER MANUALS:	