

ITEM(S) SPECIFICATIONS FOR ADVERTISEMENT

TENDER/QUOTATION NAME	TO SUPPLY AND DELIVER SINGLE CHANNEL PIPETTES WITH VARIABLE VOLUMES FOR FORENSIC BIOLOGY LABORATORY , DEPARTMENT OF SCIENTIFIC SERVICES, MINISTRY OF HEALTH (CLUSTERING)
------------------------------	---

USER REQUIREMENTS						VENDOR'S OFFER				
NO.	ITEM NAME	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/CAT NO.	PACKAGING SIZE	QUANTITY	ITEM OFFERED	PACKAGING SIZE	QUANTITY	PRICE PER UNIT (\$)	TOTAL (\$)
1	Single channel pipette with variable volumes (0.2 – 2 µl)	• Single channel pipette • Type: Manual Air Displacement • Adjustable volume • Volume range: 0.2 – 2 µl • Nominal Volume: 2 µl • Ejector Type: Metal • Autoclavable • Compatible tips: PIPETMAN DIAMOND Tips D10, DL10, DF10ST, DFL10ST • Warranty: 3 years • Must be calibrated: a. BioCal TM compliance and ISO 17025 accredited. b. Calibration certificate must be provided. c. 'As found' and 'As left/calibrated' Report (3 volume settings (nominal volume, 50% of nominal volume and minimum volume or 10% of nominal volume) using original manufacturer's tips) d. Piston sealing	GIL / FA10001M	unit	2					

USER REQUIREMENTS						VENDOR'S OFFER				
NO.	ITEM NAME	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/CAT NO.	PACKAGING SIZE	QUANTITY	ITEM OFFERED	PACKAGING SIZE	QUANTITY	PRICE PER UNIT (\$)	TOTAL (\$)
2	Single channel pipette with variable volumes (0.5 – 10 µl)	• Single channel pipette • Type: Manual Air Displacement • Adjustable volume • Volume range: 0.5 – 10 µl • Nominal Volume: 10 µl • Ejector Type: Metal • Autoclavable • Compatible tips: PIPETMAN DIAMOND Tips D10, DL10, DF10ST, DFL10ST • Warranty: 3 years • Must be calibrated: a. BioCal TM compliance and ISO 17025 accredited. b. Calibration certificate must be provided. c. 'As found' and 'As left/calibrated' Report (3 volume settings (nominal volume, 50% of nominal volume and minimum volume or 10% of nominal volume) using original manufacturer's tips) d. Piston sealing	GIL / FA10002M	unit	2					

USER REQUIREMENTS						VENDOR'S OFFER				
NO.	ITEM NAME	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/CAT NO.	PACKAGING SIZE	QUANTITY	ITEM OFFERED	PACKAGING SIZE	QUANTITY	PRICE PER UNIT (\$)	TOTAL (\$)
3	Single channel pipette with variable volumes (2 – 20 µl)	• Single channel pipette • Type: Manual Air Displacement • Adjustable volume • Volume range: 2 – 20 µl • Nominal Volume: 20 µl • Ejector Type: Metal • Autoclavable • Compatible tips: PIPETMAN DIAMOND Tips D200, DF30ST • Warranty: 3 years • Must be calibrated: a. BioCal TM compliance and ISO 17025 accredited. b. Calibration certificate must be provided. c. 'As found' and 'As left/calibrated' Report (3 volume settings (nominal volume, 50% of nominal volume and minimum volume or 10% of nominal volume) using original manufacturer's tips) d. Piston sealing	GIL / FA10003M	unit	2					

USER REQUIREMENTS						VENDOR'S OFFER				
NO.	ITEM NAME	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/CAT NO.	PACKAGING SIZE	QUANTITY	ITEM OFFERED	PACKAGING SIZE	QUANTITY	PRICE PER UNIT (\$)	TOTAL (\$)
4	Single channel pipette with variable volumes (10 – 100 µl)	• Single channel pipette • Type: Manual Air Displacement • Adjustable volume • Volume range: 10 – 100 µl • Nominal Volume: 100 µl • Ejector Type: Metal • Autoclavable • Compatible tips: PIPETMAN DIAMOND Tips D200, DF100ST • Warranty: 3 years • Must be calibrated: a. BioCal TM compliance and ISO 17025 accredited. b. Calibration certificate must be provided. c. 'As found' and 'As left/calibrated' Report (3 volume settings (nominal volume, 50% of nominal volume and minimum volume or 10% of nominal volume) using original manufacturer's tips) d. Piston sealing	GIL / FA10004M	unit	2					

USER REQUIREMENTS						VENDOR'S OFFER				
NO.	ITEM NAME	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/CAT NO.	PACKAGING SIZE	QUANTITY	ITEM OFFERED	PACKAGING SIZE	QUANTITY	PRICE PER UNIT (\$)	TOTAL (\$)
5	Single channel pipette with variable volumes (20 – 200 µl)	• Single channel pipette • Type: Manual Air Displacement • Adjustable volume • Volume range: 20 – 200 µl • Nominal Volume: 200 µl • Ejector Type: Metal • Autoclavable • Compatible tips: PIPETMAN DIAMOND Tips D200, D300, DF200ST, DF300ST • Warranty: 3 years • Must be calibrated: a. BioCal TM compliance and ISO 17025 accredited. b. Calibration certificate must be provided. c. 'As found' and 'As left/calibrated' Report (3 volume settings (nominal volume, 50% of nominal volume and minimum volume or 10% of nominal volume) using original manufacturer's tips) d. Piston sealing	GIL / FA10005M	unit	2					

USER REQUIREMENTS						VENDOR'S OFFER				
NO.	ITEM NAME	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/CAT NO.	PACKAGING SIZE	QUANTITY	ITEM OFFERED	PACKAGING SIZE	QUANTITY	PRICE PER UNIT (\$)	TOTAL (\$)
6	Single channel pipette with variable volumes (100 – 1000 µl)	• Single channel pipette • Type: Manual Air Displacement • Adjustable volume • Volume range: 100 – 1000 µl • Nominal Volume: 1000 µl • Ejector Type: Metal • Autoclavable • Compatible tips: PIPETMAN DIAMOND Tips D200, DF100ST • Warranty: 3 years • Must be calibrated: a. BioCal TM compliance and ISO 17025 accredited. b. Calibration certificate must be provided. c. 'As found' and 'As left/calibrated' Report (3 volume settings (nominal volume, 50% of nominal volume and minimum volume or 10% of nominal volume) using original manufacturer's tips) d. Piston sealing	GIL / FA10006M	unit	2					

USER REQUIREMENTS						VENDOR'S OFFER				
NO.	ITEM NAME	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/CAT NO.	PACKAGING SIZE	QUANTITY	ITEM OFFERED	PACKAGING SIZE	QUANTITY	PRICE PER UNIT (\$)	TOTAL (\$)
7	Single channel pipette with variable volumes (1 – 10 mL)	• Single channel pipette • Type: Manual Air Displacement • Adjustable volume • Volume range: 1 – 10 mL • Nominal Volume: 10 mL • Ejector Type: Plastic • Autoclavable • Compatible tips: PIPETMAN DIAMOND Tips D10mL • Warranty: 3 years • Must be calibrated: a. BioCal TM compliance and ISO 17025 accredited. b. Calibration certificate must be provided. c. 'As found' and 'As left/calibrated' Report (3 volume settings (nominal volume, 50% of nominal volume and minimum volume or 10% of nominal volume) using original manufacturer's tips) d. Piston sealing	GIL / FA10008M	unit	2					

NOTE	VENDOR'S OFFER
<i>Delivery Period Upon the Issuance of PO</i>	
<i>Price Validity must not be less than 90 days</i>	

Bahagian/Unit/Makmal :	FORENSIC BIOLOGY LABORATORY		
Bil. Rujukan Bahagian/Unit/Makmal :	DSS/FOR/PR/2026/04		
Pegawai dirujuk :	Nama : Vianaliza Rajak / Nuramalina Sawal		
	E-mail : Vianaliza.rajak@moh.gov.bn / Nuramalina.sawal@moh.gov.bn		
	Tel.No. : 2382424 ext 6079/6068	Fax No. : -	
UNTUK KEGUNAAN BAHAGIAN PENTADBIRAN SAHAJA			
No. Rujukan PROCUREMENT :			
Tarikh Menghantar Iklan :			