

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY AND DELIVERY DENTAL SLOW SPEED HANDPIECES, DEPARTMENT OF
DENTAL SERVICES**

NO.	TERMS AND CONDITIONS	VENDOR'S OFFER (PLEASE STATE)
1	Tenderer must be registered with the Ministry of Health.	
2	TENDER FORM should be filled completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form MAY cause DISQUALIFICATION OF TENDER .	
3	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF TENDER .	
4	All consumables supplied throughout this tender <u>shall</u> have a minimum expiry date of twelve (12) months / on delivery . Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made.	
5	Brochures / catalogues should be submitted / attached with tender document.	
6	Any room renovation which may be required, it is mandatory to conduct site visit (if applicable)	
7	DELIVERY PERIOD: (Please state) Not More Than 90 days upon confirmation	(Yes / No) (If No, please specify)
8	PRICE VALIDITY: The quotation shall remain valid for 12 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).	

**UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM**

**SUPPLY AND DELIVERY DENTAL SLOW SPEED HANDPIECES, DEPARTMENT OF
DENTAL SERVICES**

SECTION I - USER REQUIREMENT		YES	NO	PLEASE SPECIFY THE SPECIFICATION/ SCOPE OFFERED FAILURE TO FILL THE OFFER DETAIL SHALL BE IDENTIFIED AS NON-COMPLIANCE TO THE OFFER
Item 1: Slow speed handpiece (Contra angle)				
Standard features	Quantity: Forty (40) units			
	Dental Contra-angle handpiece			
	For contra angle burs			
	Transmission 1:1			
	Push-button chuck			
	Cellular optic			
	Anti-Heat system			
	Equipped with mechanism that prevents the aerosol particles to enter the turbine head from the oral cavity.			
Head size (Diameter in mm) :	No larger than 10mm (<10mm)			
Maximum speed:	Approximately 40,000 rpm			
Water spray:	1-port			
Connection:	Universal "E" type connection			
Compatibility:	Compatible with the current system of dental chair			

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY AND DELIVERY DENTAL SLOW SPEED HANDPIECES, DEPARTMENT OF
DENTAL SERVICES**

Item 2: Slow speed handpiece (Straight)				
Standard features	Quantity: Ten (10) units			
	Dental straight handpiece			
	For straight handpiece burs			
	Transmission 1:1			
	Equipped with mechanism that prevents the aerosol particles to enter the turbine head from the oral cavity.			
Maximum speed:	Approximately 40,000 rpm			
Water spray:	1-port			
Connection:	Universal "E" type connection			
Compatibility:	Compatible with the current system of dental chair			

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY AND DELIVERY DENTAL SLOW SPEED HANDPIECES, DEPARTMENT OF
DENTAL SERVICES**

SECTION II - PROCUREMENT REQUIREMENT	
Model and Brand of Equipment.	
Country of Origin.	
Unit (B\$)	
Total Price Offered (CIF)	
Country of Origin	
Year manufactured.	
Warranty period (at least 1 year)	_____ years
Delivery period	_____ weeks
Price validity (not less than 12 months)	_____ years

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY AND DELIVERY DENTAL SLOW SPEED HANDPIECES, DEPARTMENT OF
DENTAL SERVICES**

TECHNICAL SPECIFICATIONS/REQUIREMENTS	
Mains Supply 240VAC 50 Hz.	
State equipment ambient operating temperature range.	
The equipment must comply to either one of the five international safety standards, namely; US FDA Standard, European Union CE MARK, Australian TGA Standard, Canadian CSA Standard or Japanese JIS Standard.	
State number of locally stationed engineers/technicians and the nearest overseas support.	
State overall dimensions of equipment.	
State net weight of the equipment (kg)	
Equipment whole life time support Supplier must guarantee spare parts are available at least for 5 years after the installation of the equipment.	_____years
Brochure Tenderers must submit with their offer, a detailed brochure which supports the information the tenderer has entered in the "form-to-be-used (FORM B)" for reference.	
User and Service Manuals Two hard copies and one soft copy [PDF FORMAT] to be submitted to BME office during commissioning.	
Spare-parts and consumables Comprehensive spare-parts and consumables listings in the form of softcopy must be provided. Listing must have the following details; Part Number and Part Description.	
Warranty planned preventive maintenance. Minimum of two times PPM per year during the equipment warranty, six months after the date of commissioning and at the end of warranty period.	
Training Tenderers are required to conduct technical training to Biomedical Engineers and Technicians. Training to be conducted at the manufacturer's training premises / facilities. Tenderers are required to;	

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY AND DELIVERY DENTAL SLOW SPEED HANDPIECES, DEPARTMENT OF
DENTAL SERVICES**

- Provide air tickets, accommodation, transportation, etc. for 2 persons.
- Provide training for 2 persons.
- Minimum no of days for training – 3 days.

OR

- Training to be conducted locally, tenderers are required to:
- Provide training materials, test equipment, demo equipments, etc.
- Provide training to two groups of technical staffs.
- Provide 2 days (minimum) of training for each group.
- Training to be conducted at BME office/ Hospital.