

LAMPIRAN - A1

RUJ: BME/A2K/(1699)/2026/HTD (B8)

9/6/2026

TO REPAIR TWO (2) UNITS OF ICU BED, BRAND: HILLROM, MODEL: TC155 AT ICU 3 RIPAS HOSPITAL.

SCOPE OF WORK

To repair and supply spare parts for two (2) units of ICU Bed, Brand: Hillrom, Model: TC155 at ICU 3 Ripas Hospital. Detail are as follow:

1. BME18789, Brand: Hillrom, Model: TC155, SN: Q028AM0015, at ICU 3 W/order 2025-4716
2. BME18791, Brand: Hillrom, Model: TC155, SN: Q029AM0022, at ICU 3 W/order 2025-6136

- Scope of work to follow equipment manual/recommendation specific for this system which include the following where if necessary:

- (i) installation of items listed in spareparts list (Lampiran A2)
- (ii) alignment
- (iii) calibration
- (iv) optimisation
- (vi) any other related maintenance works required after installation of items listed in spareparts list (Lampiran A2)

- Propose to change all 12 unit of solenoid valve for each ICU Bed following observation and monitoring the system.

- Perform functional test, verification and performance check post-maintenance.

- Disposal of any used spare parts, unless otherwise instructed.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.

LAMPIRAN - A2

RUJ: BME/A2K/(1699)/2026/HTD (B8)

TO REPAIR TWO (2) UNITS OF ICU BED, BRAND: HILLROM, MODEL: TC155 AT ICU 3 RIPAS HOSPITAL.

	SPAREPART	PART NUMBER	QUANTITY
1	VLV SOLENOID 2W/2P DELTROL	6543035	24
2	SERVICE CHARGES		2