

LAMPIRAN - A1

RUJ: BME/A2K/(1700)/2026/HTD (B8)

11/6/2026

ONE-OFF PREVENTIVE MAINTENANCE SERVICE FOR FIVE (5) UNIT RESUSCITATION SYSTEM (AUTOMATED CPR), BRAND: ZOLL, MODEL: AUTOPULSE, AT EMERGENCY DEPARTMENT AND EMAS JETTY RIPAS HOSPITAL.

SCOPE OF WORK

One off preventive maintenance service for five (5) units of Resuscitation system at Emergency Department and EMAS Jetty Ripas Hospital. The details of the equipment are as follows:

1. BME29510, Brand: Zoll, Model: Autopulse S/N: 43645 at Emergency Department Ripas Hospital, W/order 2025-8111
2. BME29511, Brand: Zoll, Model: Autopulse S/N: 43683 at Emergency Department Ripas Hospital, W/order 2025-8109
3. BME29512, Brand: Zoll, Model: Autopulse S/N: 43646 at Emergency Department Ripas Hospital, W/order 2025-8108
4. BME29513, Brand: Zoll, Model: Autopulse S/N: 43649 at Emergency Department Ripas Hospital, W/order 2025-8110
5. BME29516, Brand: Zoll, Model: Autopulse S/N: 43681 at EMAS Jetty Ripas Hospital, W/order 2025-9717

- Scope of work to follow maintenance manual/recommendation specific for models, which include the following where necessary:

- (i) inspection
- (ii) cleaning
- (iii) installation of PM kits, consumables and spare parts
- (iv) alignment
- (v) calibration
- (vi) any other related preventive maintenance works required

- Functional check post-maintenance.

- Apply preventive maintenance sticker/tag after successful service.

- Disposal of any used spare parts, unless otherwise instructed.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.

- Vendor(s) to submit maintenance checklist/details/scope specific for the equipment along with other necessary documents to join this tender.

- For work which require sending the unit overseas, a loaner unit of the same model is to be provided.

- Extra battery is for Emergency 2 units and EMAS 1 unit.

LAMPIRAN - A2

RUJ: BME/A2K/(1700)/2026/HTD (B8)

ONE-OFF PREVENTIVE MAINTENANCE SERVICE FOR FIVE (5) UNIT RESUSCITATION SYSTEM (AUTOMATED CPR), BRAND: ZOLL, MODEL: AUTOPULSE, AT EMERGENCY DEPARTMENT AND EMAS JETTY RIPAS HOSPITAL.

	SPAREPART	PART NUMBER	QUANTITY
1	PPM AND SYSTEM CALIBRATION		6
2	BATTERY		8