

TENDER REFERENCE NO.: KK/125/2026/HTD

**MINISTRY OF HEALTH
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND
COMMISSIONING OF ONE (1) UNIT OF MICROSCOPE
SYSTEM FOR BERAKAS DENTAL CLINIC, DEPARTMENT
OF DENTAL SERVICES, MINISTRY OF HEALTH**

TENDER FEES : \$10.00

RECEIPT NO. :

CLOSING DATE : ON TUESDAY, 28th July 2026

TIME : 2.00 PM

FOA :

**THE CHAIRMAN
MINI TENDER BOARD, TENDER BOX
GROUND FLOOR, MINISTRY OF HEALTH
COMMONWEALTH DRIVE
BANDAR SERI BEGAWAN BB3910
NEGARA BRUNEI DARUSSALAM**

(CLUSTERING)

SECTION 2

SPECIFICATIONS AND REQUIREMENTS

TENDER REFERENCE NO: KK/125/2026/HTD

INVITATION TO TENDER

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF ONE (1) UNIT OF MICROSCOPE SYSTEM FOR BERAKAS DENTAL CLINIC, DEPARTMENT OF DENTAL SERVICES, MINISTRY OF HEALTH

1. STANDARD FEATURES
DESCRIPTION
▪ The purpose of this device is to be used in micro dentistry and dental surgical procedures
▪ Floor stand mounted with lockable wheels
TECHNICAL SPECIFICATIONS:
▪ Optical System magnification: not less than 5 steps magnification, up to 2.5X or better
▪ Objective Lens: expected working distance of 20 to 43 cm with a wide field of view
▪ Eyepieces: 10x-12.5x wide field eyepieces
▪ The optical system must be mounted on a mechanical rotating assembly allowing maximum flexibility of placement relative to the patient while allowing the operator to remain in a comfortable position
▪ Light Source illumination system: integrated LED illumination with ability to focus illumination using motor adjustments, 5500K minimum
▪ Light source illumination system modes: expected but not limited to; ✓ orange color mode, ✓ green color mode
▪ Augmented visualization modes: expected but not limited to fluorescence mode
▪ Integrated HD quality camera with recording and USB, HDMI output
▪ Must be floor stand mounted with lockable wheels
▪ Complete with a suitable length suspension arm and foldable tube for flexible positioning
POWER:
▪ Must be compatible with 220-240 V AC, 50-60 Hz mains power supply
ACCESSORIES:
▪ Any essential accessories required for the day-to-day operation of the unit and not mentioned above such as power supply cables and the like must be supplied for each unit in sufficient numbers
▪ One offline storage to store images or video if required, compatible with output of main machine (1TB formatted capacity)

1. STANDARD FEATURES	
DESCRIPTION	
	▪ At least 20 inch monitor compatible with output of main machine to display live images/video
	▪ Asepsis Starter Kit required for splash protection of equipment

2. WARRANTY	
1	Tenderer to include warranty period of at least two (2) years
2	Tenderers to acknowledge the Warranty Undertaking Form in Section 4 stating the terms of warranty provided for the equipment in the tender for the period of two years.

3. END USER TRAINING	
1	Conduct user training to the all-end users by an application specialist or competent local engineer including but not limited to: ▪ Basic user operation, user troubleshooting and user maintenance ▪ Provide Operating manual (Hardcopy and/or Softcopy)
2	Tenderer must prepare a training attendance or proof of training done to end user during commissioning and the refresher course (6) months after commissioning.

SECTION 2 – PRICE PROPOSAL	
UNIT PRICE	TOTAL PRICE

SECTION 3 - PROCUREMENT AND TECHNICAL SPECIFICATION	
BRAND:	MODEL:
COUNTRY OF ORIGIN:	YEAR INTRODUCED TO MARKET:
WARRANTY PERIOD:	LAST COUNTRY SOLD TO:
YEAR INTRODUCED TO MARKET:	LAST COUNTRY SOLD TO:
PRICE VALIDITY: [AT LEAST <u>ONE (1) YEAR</u> PRICE VALIDTY]	DELIVERY TIME:
AUTHORIZED DISTRIBUTOR: (AUTHORIZED DISTRIBUTOR LETTER ATTACHED)	
DETAILED BROCHURE INCLUDED	
USER AND SERVICE MANUALS:	
MAINS POWER SUPPLY:	
BATTERY	
POWER ADAPTER/CHARGER OUTPUT RATING:	
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:	
NUMBER OF TECHNICAL SUPPORT (ENGINEER/TECHNICIAN) Please provide training or certification for locals who is trained/certified	
DIMENSIONS AND WEIGHT OF MAIN UNIT:	
EQUIPMENT WHOLE LIFE TIME SUPPORT:	

SECTION 4 – WARRANTY UNDERTAKING FORM

Tenderer, on behalf of the manufacturer, acknowledged and agrees that when equipment is under the warranty period, must cover the scope of normal warranty below at no additional cost:

NORMAL WARRANTY

- Warrants the supplied medical equipment and its accessories to be in good condition, in working order and free from defects to the extend such equipment do not comply with specifications, under normal use for the warranty period. The scope of warranty covers to its maximum extent permitted by applicable law.
- During warranty, tenderer must rectify issues arise from any mechanical, technical or software faulty as soon as it is reported.
- **Exchange warranty;** Providing replacement units or OEM parts:
 - A. Warranty against defects – Manufacturing defects or Equipment malfunction resulted from mechanical, electrical or software failure during Commissioning or within the first _____ months of use
 - B. Faulty workmanship or unsatisfactory condition during delivery or commissioning
 - C. If a unit or accessory is deemed used item or refurbished item (not a new unit) by the user and BME Unit.
- _____ **time Planned Preventive Maintenance (PPM) PER YEAR** according to Manufacturer's Preventive Maintenance Guideline and to include one-time replacements of battery at the end of 2 years warranty period or any other relevant parts to prolong equipment lifespan.

EXCLUSION FROM WARRANTY

MOH understand that the following circumstances are not covered in the warranty and Tenderer may quote for repair and subject to MOH approval:

- Unauthorized modifications - an alteration or repair by anyone other than the Manufacturer or Authorized agent during warranty period.
- Accidental damage or problems caused by negligence or mishandling, subject to appropriate justification by both parties.
- Vandalism and Natural disasters
- Normal wear and tear

ANY OTHER EXCLUSION

Tenderer may propose below to include items or terms which is not listed in the exclusion list above for MOH consideration.

NO.	TERMS AND CONDITIONS
1	Tenderer must be registered with the Ministry of Health.
2	TENDER FORM should be filled completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form <u>MAY</u> cause DISQUALIFICATION OF TENDER.
3	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF TENDER.
4	All consumables supplied throughout this tender <u>shall</u> have a minimum expiry date of twelve (12) months / on delivery (if applicable). Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made (if applicable).
5	Brochures / catalogues should be submitted / attached with tender document.
6	Any room renovation which may be required, it is mandatory to conduct site visit (if applicable)
7	Samples should be submitted together with tender or within fourteen (14 days) of the tender closing dates (if applicable).
8	DELIVERY PERIOD: (Please state) Not More Than 90 days upon confirmation
9	PRICE VALIDITY: The quotation shall remain valid for 12 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).
10	The equipment supplied must be newly manufactured, unused, and in its original, sealed packaging. The equipment must not be previously owned, refurbished, or reconditioned in any form. The vendor is required to provide proof of manufacture date confirming the equipment is new.

SECTION 3
TENDER FORM

To:

TENDER REFERENCE NO: KK/125/2026/HTD

INVITATION TO TENDER
SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF ONE (1) UNIT OF
MICROSCOPE SYSTEM FOR BERAKAS DENTAL CLINIC, DEPARTMENT OF DENTAL
SERVICES, MINISTRY OF HEALTH

TENDER OF (name of tenderer) : _____

Company/Business Registration No. : _____

Tender Closing Date : _____

1. STANDARD FEATURES			
DESCRIPTION	Tick (✓)		STATE OR SPECIFY OR REMARKS OR BROCHURE PAGE
	Y	N	
▪ The purpose of this device is to be used in micro dentistry and dental surgical procedures			
▪ Floor stand mounted with lockable wheels			
TECHNICAL SPECIFICATIONS:			
▪ Optical System magnification: not less than 5 steps magnification, up to 2.5X or better			
▪ Objective Lens: expected working distance of 20 to 43 cm with a wide field of view			
▪ Eyepieces: 10x-12.5x wide field eyepieces			
▪ The optical system must be mounted on a mechanical rotating assembly allowing maximum flexibility of placement relative to the patient while allowing the operator to remain in a comfortable position			
▪ Light Source illumination system: integrated LED illumination with ability to focus illumination using motor adjustments, 5500K minimum			State the illumination:
▪ Light source illumination system modes: expected but not limited to; ✓ orange color mode, ✓ green color mode			
▪ Augmented visualization modes: expected but not limited to fluorescence mode			
▪ Integrated HD quality camera with recording and USB, HDMI output			
▪ Must be floor stand mounted with lockable wheels			
▪ Complete with a suitable length suspension arm and foldable tube for flexible positioning			

1. STANDARD FEATURES				
DESCRIPTION		Tick (✓)		STATE OR SPECIFY OR REMARKS OR BROCHURE PAGE
		Y	N	
POWER:				
Must be compatible with 220-240 V AC, 50-60 Hz mains power supply				
ACCESSORIES:				
Any essential accessories required for the day-to-day operation of the unit and not mentioned above such as power supply cables and the like must be supplied for each unit in sufficient numbers				
One offline storage to store images or video if required, compatible with output of main machine (1TB formatted capacity)				
At least 20 inch monitor compatible with output of main machine to display live images/video				
Asepsis Starter Kit required for splash protection of equipment				
2. WARRANTY				
1	Tenderer to include warranty period of at least two (2) years			
2	Tenderers to acknowledge the Warranty Undertaking Form in Section 4 stating the terms of warranty provided for the equipment in the tender for the period of two years.			
3. END USER TRAINING				
1	Conduct user training to the all-end users by an application specialist or competent local engineer including but not limited to: - Basic user operation, user troubleshooting and user maintenance - Provide Operating manual (Hardcopy and/or Softcopy)			
2	Tenderer must prepare a training attendance or proof of training done to end user during commissioning and the refresher course (6) months after commissioning.			

SECTION 2 – PRICE PROPOSAL	
UNIT PRICE: BND\$	TOTAL PRICE: BND\$

SECTION 3 - PROCUREMENT AND TECHNICAL SPECIFICATION											
BRAND:				MODEL:							
COUNTRY OF ORIGIN:				UNIT PRICE (B\$):							
WARRANTY PERIOD:				TOTAL PRICE (B\$):							
YEAR INTRODUCED TO MARKET:				LAST COUNTRY SOLD TO:							
PRICE VALIDITY: [AT LEAST ONE (1) YEAR PRICE VALIDITY]				DELIVERY TIME:							
AUTHORIZED DISTRIBUTOR: (AUTHORIZED DISTRIBUTOR LETTER ATTACHED)		APPOINTED BRUNEI DISTRIBUTOR									
		PROCURE FROM OVERSEA AUTHORIZED DISTRIBUTOR		COMPANY NAME:							
				COMPANY ORIGIN:							
DETAILED BROCHURE INCLUDED				YES			NO	☒ or specify where appropriate			
USER AND SERVICE MANUALS:				YES			NO	Tenderers to acknowledge that they must provide at least TWO sets of USER AND SERVICE manuals when applying commissioning form. One Set for End User, One Set for BME. (Please provide hardcopy or softcopy)			
MAINS POWER SUPPLY:				220V-240V		OTHERS:					
				50-60HZ		OTHERS:					
BATTERY				RECHARGEABLE				SINGLE-USE			
				OTHERS:							
				TYPE OF BATTERY:							
				RATING:							
POWER ADAPTER/CHARGER OUTPUT RATING:											
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:											
NUMBER OF TECHNICAL SUPPORT (ENGINEER/TECHNICIAN) Please provide training or certification for locals who is trained/certified		LOCAL				☐ Trained / Certified ☐ Not yet trained on the product					
		OVERSEA (SPECIFY LOCATION)				NEAREST LOCATION:					
DIMENSIONS AND WEIGHT OF MAIN UNIT:		☐ mm ☐ cm ☐ inch				☐ Kilogram (Kg) ☐ Gram(g) ☐ Pound (lbs)					
EQUIPMENT WHOLE LIFE TIME SUPPORT:		The supplier shall ensure that spare parts for the equipment are available for a minimum of 8 years after installation, with the support period extending beyond the expected lifecycle of the equipment. No of years: _____ (Please specify)									

SECTION 4 – WARRANTY UNDERTAKING FORM

Tenderer, on behalf of the manufacturer, acknowledged and agrees that when equipment is under the warranty period, must cover the scope of normal warranty below at no additional cost:

NORMAL WARRANTY

- Warrants the supplied medical equipment and its accessories to be in good condition, in working order and free from defects to the extent such equipment do not comply with specifications, under normal use for the warranty period. The scope of warranty covers to its maximum extent permitted by applicable law.
- During warranty, tenderer must rectify issues arise from any mechanical, technical or software faulty as soon as it is reported.
- **Exchange warranty;** Providing replacement units:
 - A. Warranty against defects – Manufacturing defects or Equipment malfunction resulted from mechanical, electrical or software failure during Commissioning or within the first _____ months of use
 - B. Faulty workmanship or unsatisfactory condition during delivery or commissioning
 - C. If a unit or accessory is deemed used item or refurbished item (not a new unit) by the user and BME Unit.
- **Two time Planned Preventive Maintenance (PPM) PER YEAR** according to Manufacturer's Preventive Maintenance Guideline and to include one-time replacements of battery at the end of 2 years warranty period or any other relevant parts to prolong equipment lifespan.

EXCLUSION FROM WARRANTY

MOH understand that the following circumstances are not covered in the warranty and Tenderer may quote for repair and subject to MOH approval:

- Unauthorized modifications - an alteration or repair by anyone other than the Manufacturer or Authorized agent during warranty period.
- Accidental damage or problems caused by negligence or mishandling, subject to appropriate justification by both parties.
- Vandalism and Natural disasters
- Normal wear and tear

ANY OTHER EXCLUSION

Tenderer may propose below to include items or terms which is not listed in the exclusion list above for MOH consideration.

TENDERER ACKNOWLEDGMENT

COMPANY CHOP AND SIGNATURE

NO.	TERMS AND CONDITIONS	VENDOR'S OFFER (PLEASE STATE)
1	Tenderer must be registered with the Ministry of Health.	
2	TENDER FORM should be filled completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form <u>MAY</u> cause DISQUALIFICATION OF TENDER.	
3	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF TENDER.	
4	All consumables supplied throughout this tender <u>shall</u> have a minimum expiry date of twelve (12) months / on delivery (if applicable). Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made (if applicable).	
5	Brochures / catalogues should be submitted / attached with tender document.	
6	Any room renovation which may be required, it is mandatory to conduct site visit (if applicable)	
7	Samples should be submitted together with tender or within fourteen (14 days) of the tender closing dates (if applicable).	
8	DELIVERY PERIOD: (Please state) Not More Than 90 days upon confirmation	
9	PRICE VALIDITY: The quotation shall remain valid for 12 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).	
10	The equipment supplied must be newly manufactured, unused, and in its original, sealed packaging. The equipment must not be previously owned, refurbished, or reconditioned in any form. The vendor is required to provide proof of manufacture date confirming the equipment is new.	

1. We offer and undertake on your acceptance of our Tender to provide the above mentioned services in accordance with your Invitation To Tender.
2. Our Tender is fully consistent with and does not contradict or derogate from anything in your Invitation To Tender. We have not qualified or changed any of the provisions of your Invitation To Tender.
3. OUR OFFER IS VALID FOR **TWELVE (12)** CALENDAR MONTHS FROM THE TENDER CLOSING DATE.
4. When requested by you, we shall extend the validity of this offer.
5. We further undertake to give you any further information which you may require.

Dated this _____ day of _____, _____

Signature of authorised officer of Tenderer

Name:

Designation:

Tenderer's official stamp