

TENDER REF. NO.: KK/149/2026/SSBH

**MINISTRY OF HEALTH
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND
COMMISSIONING VIDEO LARYNGOSCOPE FOR SURI SERI
BEGAWAN HOSPITAL, KUALA BELAIT**

TENDER FEE : \$10.00

RECEIPT NO. :

CLOSING DATE : ON TUESDAY, 04th AUGUST 2026

TIME : 2.00 PM

FOA :

**THE CHAIRMAN
MINI TENDER BOARD, TENDER BOX
GROUND FLOOR, MINISTRY OF HEALTH
COMMONWEALTH DRIVE
BANDAR SERI BEGAWAN BB3910
NEGARA BRUNEI DARUSSALAM**

[CLUSTERING]

CONTENTS

SECTION 1 – USER REQUIREMENTS	2
1. TWO (2) UNITS OF VIDEO LARYNGOSCOPE	Error! Bookmark not defined.
2. WARRANTY	Error! Bookmark not defined.
3. END USER TRAINNING	Error! Bookmark not defined.
4. TECHNICAL TRAINNING	Error! Bookmark not defined.
SECTION 2 - PRICE PROPOSAL	4
SECTION 3 - PROCUMENT AND TECHNICAL SPECIFICATION	5
SECTION 4 – WARRANTY UNDERTAKING FORM.....	6

SECTION 2

SPECIFICATIONS AND REQUIREMENTS

TENDER REFERENCE NO.: KK/149/2026/SSBH

INVITATION TO TENDER

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING VIDEO
LARYNGOSCOPE FOR SURI SERI BEGAWAN HOSPITAL, KUALA BELAIT**

SECTION 1 – USER REQUIREMENTS	
REF.	DESCRIPTION
1	TWO (2) UNITS OF VIDEO LARYNGOSCOPE
2	WARRANTY
3	END USER TRAINING
4	TECHNICAL TRAINING

	SECTION 1 – USER REQUIREMENTS
REF.	DESCRIPTION
1	TWO (2) UNITS OF VIDEO LARYNGOSCOPE
1.1	Two (2) units of Video Laryngoscope for: 1. Emergency Department and 2. Operating Theatre
1.2	Airway management system designed for both routine or difficult endotracheal intubations and can perform both direct and video laryngoscopy.
1.3	The system must be a rechargeable system on wheels.
1.4	The system display screen size must be at least 8" with two camera inputs.
1.5	The display monitor user interface is via touch screen or buttons
1.6	The system is capable to record and playback video clips and images into or from SD card
1.7	The display monitor is IP54 Splash-proof or better for easy disinfection cleaning
1.8	The display monitor is in ABS plastic housing with shock resistant capabilities
1.9	The display monitor must be mounted on cart of five lockable antistatic castors with sturdy and balance design for stability.
1.10	Cart has storage basket to store laryngoscopes
1.11	Each unit comes with one (1) unit of compatible CMOS Standard Macintosh laryngoscope blade size 2 with working length not more than 92mm or effective angle of view of up to 80°. (Suitable for paediatrics and older children)
1.12	Each unit comes with one (1) unit of compatible CMOS Standard Macintosh laryngoscope blade size 3 with working length not more than 104mm or effective angle of view of up to 70°. (Suitable for standard adults)
1.13	Each unit comes with one (1) unit of compatible CMOS Standard Macintosh laryngoscope blade size 4 with working length not more than 124mm or effective angle of view of up to 60°. (Suitable for large or obese adult)
1.14	Each unit comes with one (1) unit of compatible CMOS Hyper angulated laryngoscope blade for adult with working length not more than 109mm or effective angle of view of up to 80°. (Suitable for difficult airways)
1.15	Each unit comes with one (1) unit of connecting cable to connect all supplied laryngoscope to the display monitor.
1.16	Each unit comes with one (1) unit of Sterilization wire tray
1.17	Each unit comes with twenty (20) units of guide rod with atraumatic tip or equivalent
1.18	Each unit comes with one (1) unit of water-resistant and washable protective bag
2	WARRANTY
2.1	Tenderer to include comprehensive warranty for a period of at least one (1) year
2.2	Tenderers to ACKNOWLEDGE the Warranty Undertaking Form in Section 4 stating the terms of warranty provided for the equipment in the tender for the period of two years. This includes but not limited to:
2.2.1	Scope of Warranty
2.2.2	One time Planned Preventive Maintenance per year during warranty in accordance to Manufacturer's Standard
2.2.3	Comprehensive Breakdown and Corrective maintenance repair during warranty

SECTION 1 – USER REQUIREMENTS	
REF.	DESCRIPTION
3	END USER TRAINING
3.1	Conduct user training to the all-end users by an application specialist or competent local engineer including but not limited to:
3.1.1	Basic user operation, user troubleshooting and user maintenance
3.1.2	Provide Operating manual (Hardcopy and/or Softcopy)
3.1.3	Tenderer must prepare a training attendance or proof of training done to end user during commissioning and the refresher course (6) months after commissioning.
4	TECHNICAL TRAINING
4.1	Introductory Technical Training to Biomedical Engineers and Technicians at BME Office by competent Tenderer's Engineer/Technicians that includes but not limited to: ▪ Troubleshooting and basic corrective maintenance ▪ Handling and basic inspection maintenance *(Two sessions/groups if required)

SECTION 2 - PRICE PROPOSAL	
UNIT PRICE	TOTAL PRICE

SECTION 3 - PROCUMENT AND TECHNICAL SPECIFICATION	
BRAND:	MODEL:
COUNTRY OF ORIGIN:	UNIT PRICE (B\$):
WARRANTY PERIOD:	TOTAL PRICE (B\$):
YEAR INTRODUCED TO MARKET:	LAST COUNTRY SOLD TO:
PRICE VALIDITY: [AT LEAST ONE (1) YEAR PRICE VALIDTY]	DELIVERY TIME:
AUTHORIZED DISTRIBUTOR: (AUTHORIZED DISTRIBUTOR LETTER ATTACHED)	
DETAILED BROCHURE INCLUDED	
USER AND SERVICE MANUALS:	
MAINS POWER SUPPLY:	
BATTERY	
POWER ADAPTER/CHARGER OUTPUT RATING:	
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:	
NUMBER OF TECHNICAL SUPPORT (ENGINEER/TECHNICIAN) Please provide training or certification for locals who is trained/certified	
DIMENSIONS AND WEIGHT OF MAIN UNIT:	
EQUIPMENT WHOLE LIFE TIME SUPPORT:	

SECTION 4 – WARRANTY UNDERTAKING FORM

Tenderer, on behalf of the manufacturer, acknowledged and agrees that when equipment is under the warranty period, must cover the scope of normal warranty below at no additional cost:

NORMAL WARRANTY

- Warrants the supplied medical equipment and its accessories to be in good condition, in working order and free from defects to the extend such equipment do not comply with specifications, under normal use for the warranty period. The scope of warranty covers to its maximum extent permitted by applicable law.
- During warranty, tenderer must rectify issues arise from any mechanical, technical or software faulty as soon as it is reported.
- **Exchange warranty**; Providing replacement units or OEM parts:
 - A. Warranty against defects – Manufacturing defects or Equipment malfunction resulted from mechanical, electrical or software failure during Commissioning or within the first _____ months of use
 - B. Faulty workmanship or unsatisfactory condition during delivery or commissioning
 - C. If a unit or accessory is deemed used item or refurbished item (not a new unit) by the user and BME Unit.
- **One time Planned Preventive Maintenance (PPM) PER YEAR** according to Manufacturer's Preventive Maintenance Guideline including PM kits and any other relevant parts to prolong equipment lifespan.

EXCLUSION FROM WARRANTY

MOH understand that the following circumstances are not covered in the warranty and Tenderer may quote for repair and subject to MOH approval:

- Unauthorized modifications - an alteration or repair by anyone other than the Manufacturer or Authorized agent during warranty period.
- Accidental damage or problems caused by negligence or mishandling, subject to appropriate justification by both parties.
- Vandalism and Natural disasters
- Normal wear and tear

ANY OTHER EXCLUSION

Tenderer may propose below to include items or terms which is not listed in the exclusion list above for MOH consideration.

SECTION 3
TENDER FORM

To:

TENDER REFERENCE NO.: KK/149/2026/SSBH

**INVITATION TO TENDER
SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING VIDEO
LARYNGOSCOPE FOR SURI SERI BEGAWAN HOSPITAL, KUALA BELAIT**

TENDER OF (*name of tenderer*) : _____

Company/Business Registration No. : _____

Tender Closing Date : _____

SECTION 1 – USER REQUIREMENTS				
REF. NO.	DESCRIPTION	Tick (✓)		STATE OR SPECIFY OR REMARKS OR BROCHURE PAGE
		YES	NO	
1	TWO (2) UNITS OF VIDEO LARYNGOSCOPE			
2	WARRANTY			
3	END USER TRAINING			
4	TECHNICAL TRAINING			

SECTION 1 – USER REQUIREMENTS				
REF.	DESCRIPTION	Tick (✓)		STATE OR SPECIFY OR REMARKS OR BROCHURE PAGE
		YES	NO	
1	TWO (2) UNITS OF VIDEO LARYNGOSCOPE			
1.1	Two (2) units of Video Laryngoscope for: 1. Emergency Department and 2. Operating Theatre			
1.2	Airway management system designed for both routine or difficult endotracheal intubations and can perform both direct and video laryngoscopy.			
1.3	The system must be a rechargeable system on wheels.			
1.4	The system display screen size must be at least 8" with two camera inputs.			
1.5	The display monitor user interface is via touch screen or buttons			
1.6	The system is capable to record and playback video clips and images into or from SD card			
1.7	The display monitor is IP54 Splash-proof or better for easy disinfection cleaning			
1.8	The display monitor is in ABS plastic housing with shock resistant capabilities			
1.9	The display monitor must be mounted on cart of five lockable antistatic castors with sturdy and balance design for stability.			
1.10	Cart has storage basket to store laryngoscopes			
1.11	Each unit comes with one (1) unit of compatible CMOS Standard Macintosh laryngoscope blade size 2 with working length not more than 92mm or effective angle of view of up to 80°. (Suitable for paediatrics and older children)			
1.12	Each unit comes with one (1) unit of compatible CMOS Standard Macintosh laryngoscope blade size 3 with working length not more than 104mm or effective angle of view of up to 70°. (Suitable for standard adults)			
1.13	Each unit comes with one (1) unit of compatible CMOS Standard Macintosh laryngoscope blade size 4 with working length not more than 124mm or effective angle of view of up to 60°. (Suitable for large or obese adult)			
1.14	Each unit comes with one (1) unit of compatible CMOS Hyper angulated laryngoscope blade for adult with working length not more than 109mm or effective angle of view of up to 80°. (Suitable for difficult airways)			
1.15	Each unit comes with one (1) unit of connecting cable to connect all supplied laryngoscope to the display monitor.			
1.16	Each unit comes with one (1) unit of Sterilization wire tray			
1.17	Each unit comes with twenty (20) units of guide rod with atraumatic tip or equivalent			

SECTION 1 – USER REQUIREMENTS				
REF.	DESCRIPTION	Tick (✓)		STATE OR SPECIFY OR REMARKS OR BROCHURE PAGE
		YES	NO	
1.18	Each unit comes with one (1) unit of water-resistant and washable protective bag			
2	WARRANTY			
2.1	Tenderer to include comprehensive warranty for a period of at least one (1) year			
2.2	Tenderers to ACKNOWLEDGE the Warranty Undertaking Form in Section 4 stating the terms of warranty provided for the equipment in the tender for the period of two years. This includes but not limited to:			
2.2.1	Scope of Warranty			
2.2.2	One time Planned Preventive Maintenance per year during warranty in accordance to Manufacturer's Standard			
2.2.3	Comprehensive Breakdown and Corrective maintenance repair during warranty			
3	END USER TRAINING			
3.1	Conduct user training to the all-end users by an application specialist or competent local engineer including but not limited to:			
3.1.1	Basic user operation, user troubleshooting and user maintenance			
3.1.2	Provide Operating manual (Hardcopy and/or Softcopy)			
3.1.3	Tenderer must prepare a training attendance or proof of training done to end user during commissioning and the refresher course (6) months after commissioning.			
4	TECHNICAL TRAINING			
4.1	Introductory Technical Training to Biomedical Engineers and Technicians at BME Office by competent Tenderer's Engineer/Technicians that includes but not limited to: ▪ Troubleshooting and basic corrective maintenance ▪ Handling and basic inspection maintenance *(Two sessions/groups if required)			

SECTION 2 - PRICE PROPOSAL	
UNIT PRICE: BND\$	TOTAL PRICE: BND\$

SECTION 3 - PROCUREMENT AND TECHNICAL SPECIFICATION					
BRAND:		MODEL:			
COUNTRY OF ORIGIN:		UNIT PRICE (B\$):			
WARRANTY PERIOD:		TOTAL PRICE (B\$):			
YEAR INTRODUCED TO MARKET:		LAST COUNTRY SOLD TO:			
PRICE VALIDITY: [AT LEAST ONE (1) YEAR PRICE VALIDITY]		DELIVERY TIME:			
AUTHORIZED DISTRIBUTOR: (AUTHORIZED DISTRIBUTOR LETTER ATTACHED)	APPOINTED BRUNEI DISTRIBUTOR				
		PROCURE FROM OVERSEA AUTHORIZED DISTRIBUTOR	COMPANY NAME:		
			COMPANY ORIGIN:		
DETAILED BROCHURE INCLUDED		YES		NO	☒ or specify where appropriate
USER AND SERVICE MANUALS:		YES		NO	Tenderers to acknowledge that they must provide at least TWO sets of USER AND SERVICE manuals when applying commissioning form. One Set for End User, One Set for BME. (Please provide hardcopy or softcopy)
MAINS POWER SUPPLY:		220V-240V		OTHERS:	
		50-60HZ		OTHERS:	
BATTERY		RECHARGEABLE		SINGLE-USE	REPLACEABLE
		OTHERS:			
	TYPE OF BATTERY:				
	RATING:				
POWER ADAPTER/CHARGER OUTPUT RATING:					
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:					
NUMBER OF TECHNICAL SUPPORT (ENGINEER/TECHNICIAN) Please provide training or certification for locals who is trained/certified	LOCAL			☐ Trained / Certified ☐ Not yet trained on the product	
	OVERSEA (SPECIFY LOCATION)			NEAREST LOCATION:	
DIMENSIONS AND WEIGHT OF MAIN UNIT:		☐ mm ☐ cm ☐ inch		☐ Kilogram (Kg) ☐ Gram(g) ☐ Pound (lbs)	
EQUIPMENT WHOLE LIFE TIME SUPPORT:	The supplier shall ensure that spare parts for the equipment are available for a minimum of 8 years after installation, with the support period extending beyond the expected lifecycle of the equipment. No of years: _____ (Please specify)				

SECTION 4 – WARRANTY UNDERTAKING FORM

Tenderer, on behalf of the manufacturer, acknowledged and agrees that when equipment is under the warranty period, must cover the scope of normal warranty below at no additional cost:

NORMAL WARRANTY

- Warrants the supplied medical equipment and its accessories to be in good condition, in working order and free from defects to the extend such equipment do not comply with specifications, under normal use for the warranty period. The scope of warranty covers to its maximum extent permitted by applicable law.
- During warranty, tenderer must rectify issues arise from any mechanical, technical or software faulty as soon as it is reported.
- **Exchange warranty**; Providing replacement units or OEM parts:
 - A. Warranty against defects – Manufacturing defects or Equipment malfunction resulted from mechanical, electrical or software failure during Commissioning or within the first _____ months of use
 - B. Faulty workmanship or unsatisfactory condition during delivery or commissioning
 - C. If a unit or accessory is deemed used item or refurbished item (not a new unit) by the user and BME Unit.
- **One time Planned Preventive Maintenance (PPM) PER YEAR** according to Manufacturer's Preventive Maintenance Guideline including PM kits and any other relevant parts to prolong equipment lifespan.

EXCLUSION FROM WARRANTY

MOH understand that the following circumstances are not covered in the warranty and Tenderer may quote for repair and subject to MOH approval:

- Unauthorized modifications - an alteration or repair by anyone other than the Manufacturer or Authorized agent during warranty period.
- Accidental damage or problems caused by negligence or mishandling, subject to appropriate justification by both parties.
- Vandalism and Natural disasters
- Normal wear and tear

ANY OTHER EXCLUSION

Tenderer may propose below to include items or terms which is not listed in the exclusion list above for MOH consideration.

TENDERER ACKNOWLEDGMENT

COMPANY CHOP AND SIGNATURE

1. We offer and undertake on your acceptance of our Tender to provide the above-mentioned services in accordance with your Invitation To Tender.
2. Our Tender is fully consistent with and does not contradict or derogate from anything in your Invitation To Tender. We have not qualified or changed any of the provisions of your Invitation To Tender.
3. OUR OFFER IS VALID FOR **TWELVE (12)** CALENDAR MONTHS FROM THE TENDER CLOSING DATE.
4. When requested by you, we shall extend the validity of this offer.
5. We further undertake to give you any further information which you may require.

Dated this _____ day of _____, _____

Signature of authorised officer of Tenderer

Name:

Designation:

Tenderer's official stamp