


SPECIFICATIONS AND REQUIREMENTS

USER DESCRIPTIONS AND REQUIREMENTS	YES	NO
1. Skill Guide with extension cable (10 units)		
• <i>CPR device that is compatible with Little Anne QCPR, Little Junior QCPR and Little Baby QCPR</i>		
• <i>Provide real-time feedback</i>		
○ <i>Compression depth</i>		
○ <i>Compression rate</i>		
○ <i>Incomplete release</i>		
○ <i>Incorrect hand position</i>		
○ <i>Ventilation volume</i>		
○ <i>Compression and ventilation counter</i>		
• <i>Provide real-time summative feedback</i>		
○ <i>Compression score %</i>		
○ <i>Ventilation score %</i>		
○ <i>CPR duration mm.ss</i>		
○ <i>Flow fraction %</i>		
• <i>Has extension cable</i>		
• <i>Warranty: at least 12 months</i>		
		
2. AED trainer pad – Adult (30 units)		
• <i>Can be used with Laerdal AED Trainer</i>		
• <i>Connected with the cable</i>		
• <i>Adult pad</i>		
• <i>Warranty: at least 12 months</i>		
		

fay

USER DESCRIPTIONS AND REQUIREMENTS		YES	NO
3. AED trainer pad – Paediatric (15 units)			
• Can be used with Laerdal AED Trainer			
• Connected with the cable			
• Paediatric pad			
• Warranty: at least 12 months			
			
4. Manikin Face Shield 36 pieces per box (15 boxes)			
• Disposable			
• For training purposes only			
• Provide a hygiene barrier during CPR training on a manikin			
• One size fits all			
• Latex-free			
• Per roll: At least 36 pieces			
			
5. Intraosseous Access Needle – Size 14 (15 units)			
• Cook Medical's Intraosseous Infusion Needle			
• Standard Hub Design			
• T-shape Handle			
• Side Ports: 2 laterally opposed side ports near the distal tip			
• Needle Gauge: 14			
• Needle Length: 3 cm			
• Needle Sideports: 2			
			

fy

USER DESCRIPTIONS AND REQUIREMENTS	YES	NO
6. Intraosseous Access Needle – Size 16 (15 units)		
• Cook Medical's Intraosseous Infusion Needle		
• Standard Hub Design		
• T-shape Handle		
• Side Ports: 2 laterally opposed side ports near the distal tip		
• Needle Gauge: 16		
• Needle Length: 2.5 cm		
• Needle Sideports: 2		
		
7. First Aid Bag (4 units)		
• Material: 1,000 denier nylon		
• Waterproof bottom and plenty of room		
• At least 3 adjustable inner compartments		
• 2 front pockets		
• 2 side pockets		
• Dimensions: at least L38 cm x W18 cm x H23 cm		
• Shoulder strap		
		

QUOTATION REFERENCE:	
QUOTATION NAME:	PEMBELIAN PERALATAN LATIHAN DAN BARANGAN GUNA HABIS BAGI KURSUS (NON-CLUSTERING)

NO	ITEM	QUANTITY	PRICE (B\$)	TOTAL PRICE (B\$)
1.	<i>Skill Guide with extension cable</i>	<i>10 units</i>		
2.	<i>AED Trainer Pad (Adult)</i>	<i>30 units</i>		
3.	<i>AED Trainer Pad (Paediatric)</i>	<i>15 units</i>		
4.	<i>Face Shield</i>	<i>36 pieces per box (15 boxes)</i>		
5.	<i>Intraosseous Access Needle Size 14</i>	<i>15 units</i>		
6.	<i>Intraosseous Access Needle Size 16</i>	<i>15 units</i>		
7.	<i>First Aid Bag</i>	<i>4 units</i>		
	TOTAL PRICE (B\$)			

QUOTATION REFERENCE:**PEMBELIAN PERALATAN LATIHAN DAN BARANGAN GUNA HABIS BAGI KURSUS (*NON-CLUSTERING*)**

DELIVERY OF PERIOD: Not more than 60 days upon confirmation	(Yes/No) (If No, please specify)	
PRICE VALIDITY: <i>The quotation shall remain valid for 6 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation with that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have the written consent of the suppliers.</i>		

TERMS AND CONDITIONS			
a.	Tenderer must be registered with the Ministry of Health	Acknowledgement: Company Ref. No.: I hereby certify the above quote to be correct. Signature: Name: Designation: Date : 	Company's Official Stamp
b.	Please fill in the QUOTATION FORM completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form <u>may</u> cause DISQUALIFICATION OF QUOTATION		
c.	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF QUOTATION		
d.	Delivery Period: Not more than 60 days upon confirmation		
e.	Please do not use correction pen for amendment		