

LAMPIRAN - A1

RUJ: BME/A2K/(1618)/2026/HTD (A17)

3/8/2026

ONE-OFF PPM FOR SEVENTEEN (17) UNITS OF WEIGHING BALANCES IN DEPARTMENT OF LAB SERVICES,
MINISTRY OF HEALTH

SCOPE OF WORK

1. PREVENTIVE MAINTENANCE SHALL FOLLOW EQUIPMENT MANUAL AND BRAND/MODEL SPECIFIC RECOMMENDATIONS AND MAY INCLUDE BUT NOT LIMITED TO:

INSPECT THE EXTERIOR OF THE INSTRUMENT
INSPECT AND CLEAN THE TRAY
ENSURE POWER INLET IS WORKING
CHECK MAIN CABLES FOR DAMAGE
CHECK THE INSTRUMENT FOR ANY CRACKS IN THE CASEWORK
CHECK THE DISPLAY AND KEYBOARD OPERATION
PERFORM WEIGHT CALIBRATION USING CALIBRATED STANDARD WEIGHT

2. POST-MAINTENANCE REQUIREMENTS

Functional and performance testing
Provide a PPM sticker for machines stating company name, time and date of ppm

3. WORK MUST BE CARRIED OUT BY AUTHORISED ENGINEERS TRAINED IN MAINTAINING THE EQUIPMENT

4. VENDORS SHALL SUBMIT A MAINTENANCE CHECKLIST AND SUPPORTING DOCUMENTS SPECIFIC TO THE EQUIPMENT AS PART OF THE TENDER.

5. PREVENTIVE MAINTENANCE MAY BE PERFORMED UP TO ONE MONTH BEFORE THE NEXT SCHEDULED DATE BUT NOT EARLIER.

List of weighing scale equipment for PPM:

1	BRUNEI MUARA BDC, CPL	Centron Technologies	CB220	3236	11722	PPM/BME/BDC/2009/003
2	BRUNEI MUARA BDC, CPL	Centron Technologies	CB220	3679	12889	PPM/BDC/BME/2017/069
3	BRUNEI MUARA BDC, CPL	TERUMO	COMPOSCALE	CS300	812087	7036
PPM/BDC/BME/2009/015						
4	BRUNEI MUARA NCMRL, Andrology Laboratory	METTLER TOLEDO			JP303G B232184857	16288
	PPM/AND/BME/2014/BAL01					
5	BRUNEI MUARA NCMRL, Andrology Laboratory	METTLER TOLEDO			JP303G B232184858	16289
	PPM/AND/BME/2014/BAL02					
6	BRUNEI MUARA NCMRL, CYTOLOGY LABORATORY	METTLER TOLEDO			Toledo Gmbh (CH-8606)	
	1117410703 2436 PPM/CYT/BME/2000/BAL01					
7	BRUNEI MUARA NCMRL, HISTOLOGY LABORATORY	GX GFSeries/GX200			GX GFSeries/GX200	
	54042 12056 PPM/HIS/BME/2011/BAL02					
8	BRUNEI MUARA NCMRL, HISTOLOGY LABORATORY	OHAUS Pioneer PA2102			8329290443	12058
	PPM/HIS/BME/2011/BAL03					
9	BRUNEI MUARA NCMRL, HISTOLOGY LABORATORY	UWE HS - 15K			UWE HS - 15K	HS0011617
	10763 PPM/HIS/BME/2009/BAL01					
10	BRUNEI MUARA NCCRL DENVER	ACA 100		41255	N/A	PPM/CC002/09
11	BRUNEI MUARA NCMRL, Media room	A&D	EJ-300	5A2800176	11937	PPM/MIC/BME/2010/29
12	BRUNEI MUARA NHRL, Morphology Section	METTLER TOLEDO			1117410704	2435
13	BRUNEI MUARA NMRL, SUMBILING	OHAUS	PA214	8330070032	11793	PPM/NMRL/BME/027
14	BELAIT Blood Transfusion	TERUMO	COMPOSCALE	CS300	812086	7035 NONE
15	TEMBURONG Lab analysis	A&D	EJ-300	5A2809004	11939	PPM/PHM/2010/25
16	TEMBURONG Lab analysis	MEDITECH	TCS-200LP	415973	20677	PPM/PHM/BME/2018/53
17	TEMBURONG Lab analysis	TERUMO	CS 300	1405523	16784	PPM/PHM/BME/2018/48

LAMPIRAN - A2

RUJ: BME/A2K/(1618)/2026/HTD (A17)

**ONE-OFF PPM FOR SEVENTEEN (17) UNITS OF WEIGHING BALANCES IN DEPARTMENT OF LAB SERVICES,
MINISTRY OF HEALTH**

	SPAREPART	PART NUMBER	QUANTITY
1	ONE OFF PPM - weighing scale		17

RUJ: BME/A2K/(1618)/2026/HTD(A17)

ONE-OFF PPM FOR SEVENTEEN (17) UNITS OF WEIGHING BALANCES IN DEPARTMENT OF LAB SERVICES, MINISTRY OF HEALTH

No	District	Location	Brand	Model	Serial No.	BME No.	Lab Reference No.
1	BRUNEI MUARA	BDC, CPL	Centron Technologies	CB220	3236	11722	PPM/BME/BDC/2009/003
2	BRUNEI MUARA	BDC, CPL	Centron Technologies	CB220	3679	12889	PPM/BDC/BME/2017/069
3	BRUNEI MUARA	BDC, CPL	TERUMO	COMPOSCALE CS300	812087	7036	PPM/BDC/BME/2009/015
4	BRUNEI MUARA	NCMRL, Andrology Laboratory	METTLER TOLEDO	JP303G	B232184857	16288	PPM/AND/BME/2014/BAL01
5	BRUNEI MUARA	NCMRL, Andrology Laboratory	METTLER TOLEDO	JP303G	B232184858	16289	PPM/AND/BME/2014/BAL02
6	BRUNEI MUARA	NCMRL, CYTOLOGY LABORATORY	METTLER TOLEDO	Toledo Gmbh (CH-8606)	1117410703	2436	PPM/CYT/BME/2000/BAL01
7	BRUNEI MUARA	NCMRL, HISTOLOGY LABORATORY	GX GFSeries/GX200	GX GFSeries/GX200	54042	12056	PPM/HIS/BME/2011/BAL02
8	BRUNEI MUARA	NCMRL, HISTOLOGY LABORATORY	OHAUS	Pioneer PA2102	8329290443	12058	PPM/HIS/BME/2011/BAL03
9	BRUNEI MUARA	NCMRL, HISTOLOGY LABORATORY	UWE HS - 15K	UWE HS - 15K	HS0011617	10763	PPM/HIS/BME/2009/BAL01
10	BRUNEI MUARA	NCCRL	DENVER	ACA 100	41255	N/A	PPM/CC002/09
11	BRUNEI MUARA	NCMRL, Media room	A&D	EJ-300	5A2800176	11937	PPM/MIC/BME/2010/29
12	BRUNEI MUARA	NHRL, Morphology Section	METTLER TOLEDO		1117410704	2435	
13	BRUNEI MUARA	NMRL, SUMBILING	OHAUS	PA214	8330070032	11793	PPM/NMRL/BME/027
14	BELAIT	Blood Transfusion	TERUMO	COMPOSCALE CS300	812086	7035	NONE
15	TEMBURONG	Lab analysis	A&D	EJ-300	5A2809004	11939	PPM/PHM/2010/25
16	TEMBURONG	Lab analysis	MEDITECH	TCS-200LP	415973	20677	PPM/PHM/BME/2018/53
17	TEMBURONG	Lab analysis	TERUMO	CS 300	1405523	16784	PPM/PHM/BME/2018/48