

LAMPIRAN - A1

RUJ: BME/A2K/(1711)/2026/HTD (B8)

1/7/2026

TO REPAIR ONE (1) UNIT OF RHINO-LARYNGO VIDEOSCOPE, BRAND : OLYMPUS, MODEL : ENF-V2 AT OTORHINOLYANGOLOGY CLINIC, RIPAS HOSPITAL

SCOPE OF WORK

To repair one (1) unit of Rhino-Laryngo Videoscope System, Brand: Olympus, Model: ENF-V2 at Otorhinolaryngology Clinic Ripas Hospital. Detail are as follow:

1. BME8619, Rhino-Laryngo Videoscope, Brand: Olympus, Model: ENF-V2, SN: 2150529, at Otorhinolaryngology Clinic, W/order 2026-12536

- Scope of work to follow equipment manual/recommendation specific for this system which include the following where if necessary:

- (i) installation of items listed in spareparts list (Lampiran A2)
- (ii) alignment
- (iii) calibration
- (iv) optimisation
- (vi) any other related maintenance works required after installation of items listed in spareparts list (Lampiran A2)

- Perform functional test, verification and performance check post-maintenance.

- Disposal of any used spare parts, unless otherwise instructed.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.

LAMPIRAN - A2

RUJ: BME/A2K/(1711)/2026/HTD (B8)

TO REPAIR ONE (1) UNIT OF RHINO-LARYNGO VIDEOSCOPE, BRAND : OLYMPUS, MODEL : ENF-V2 AT OTORHINOLYANGOLOGY CLINIC, RIPAS HOSPITAL

	SPAREPART	PART NUMBER	QUANTITY
1	COST OF REPAIR AND SERVICE, ANGULATION TEST		1
2	D/E INSERTION UNIT FOR ENF-V2		1