

LAMPIRAN - A1

RUJ: BME/A2K/(1717)/2026/HTD (B8)

18/7/2026

TO REPAIR ONE (1) UNIT OF FRACTIONAL NITRIC OXIDE (FENO), BRAND: NIOX, MODEL: NIOX VERO AT D.O. R.M. RIPAS HOSPITAL.

SCOPE OF WORK

To repair one (1) unit of Fractional Nitric Oxide (FENO), Brand: NIOX, Model: NIOX VERO at D.O.R.M. Ripas Hospital. Detail are as follow:

1. BME31930, Brand: NIOX, Model: NIOX VERO, SN: 350501968, at D.O.R.M. W/order 2026-11314

- Scope of work to follow equipment manual/recommendation specific for this system which include the following where if necessary:

(i) installation of items listed in spareparts list (Lampiran A2)

(ii) alignment

(iii) calibration

(iv) optimisation

(vi) any other related maintenance works required after installation of items listed in spareparts list (Lampiran A2)

- Perform functional test, verification and performance check post-maintenance.

- Disposal of any used spare parts, unless otherwise instructed.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.

LAMPIRAN - A2

RUJ: BME/A2K/(1717)/2026/HTD (B8)

TO REPAIR ONE (1) UNIT OF FRACTIONAL NITRIC OXIDE (FENO), BRAND: NIOX, MODEL: NIOX VERO AT D.O. R.M. RIPAS HOSPITAL.

SPAREPART		PART NUMBER	QUANTITY
1	NIOX® SENSOR (1 BOX PACK OF 3)	TK-100-NH	1