

LAMPIRAN - A1

RUJ: BME/A2K/(1728)/2025/HTD (B8)

5/8/2026

TO REPAIR FIVE (5) UNITS OF PATIENT TROLLEY, BRAND: MEDICAL MASTER, MODEL: MET-6600, AT THE EMERGENCY DEPARTMENT, RIPAS HOSPITAL.

SCOPE OF WORK

To repair five (5) units of Patient Trolley at Emergency Department Ripas Hospital. The details of the equipment are as follows:

1. Brand: Medical Master, Model: MET-6600 at Emergency Department Ripas Hospital, W/order 2026-12956
2. BME32101, Brand: Medical Master, Model: MET-6600 at Emergency Department Ripas Hospital, W/order 2026-12949
3. Brand: Medical Master, Model: MET-6600 at Emergency Department Ripas Hospital, W/order 2026-13089
4. BME16122, Brand: Medical Master, Model: MET-6600 at Emergency Department Ripas Hospital, W/order 2026-12947
5. BME16112, Brand: Medical Master, Model: MET-6600 at Emergency Department Ripas Hospital, W/order 2026-12951

- Scope of work to follow maintenance manual/recommendation specific for models, which include the following where necessary:

- (i) inspection
- (ii) cleaning
- (iii) installation of PM kits, consumables and spare parts
- (iv) alignment
- (v) calibration
- (vi) any other related preventive maintenance works required

- Functional check post-maintenance.

- Disposal of any used spare parts, unless otherwise instructed.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.

- Vendor(s) to submit maintenance checklist/details/scope specific for the equipment along with other necessary documents to join this tender.

LAMPIRAN - A2

RUJ: BME/A2K/(1728)/2025/HTD (B8)

TO REPAIR FIVE (5) UNITS OF PATIENT TROLLEY, BRAND: MEDICAL MASTER, MODEL: MET-6600, AT THE EMERGENCY DEPARTMENT, RIPAS HOSPITAL.

	SPAREPART	PART NUMBER	QUANTITY
1	TO TROUBLESHOOT AND REPAIR		5
2	GAS SPRING (SMALL)		3
3	CENTRAL LOCK PEDAL		3
4	FITTING GAS SPRING		1
5	200 mm TOTAL BRAKE CASTER		6
6	GAS SPRING (BIG)		1
7			