

ITEM(S) SPECIFICATIONS FOR ADVERTISEMENT

TENDER/QUOTATION NAME	TO SUPPLY AND DELIVER CALIBRATED CLASS A VOLUMETRIC FLASKS FOR PHARMACY LABORATORY, DEPARTMENT OF SCIENTIFIC SERVICES, MINISTRY OF HEALTH
------------------------------	--

USER REQUIREMENTS						VENDOR'S OFFER				
NO.	ITEM NAME	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/CAT NO.	PACKAGING SIZE	QUANTITY	ITEM OFFERED	PACKAGING SIZE	QUANTITY	PRICE PER UNIT (\$)	TOTAL (\$)
1	50 ml Clear Volumetric Flask with PE Stopper	• Class A volumetric flask compliant with EN ISO 1042 / DIN 12664 • Individual calibration certificate issued by an ISO/IEC 17025-accredited laboratory • Stated measurement uncertainty • Demonstrable traceability to SI units • Clear borosilicate glass	-	Unit	120 units					
2	50 ml Amber Volumetric Flask with PE Stopper	• Class A volumetric flask compliant with EN ISO 1042 / DIN 12664 • Individual calibration certificate issued by an ISO/IEC 17025-accredited laboratory • Stated measurement uncertainty • Demonstrable traceability to SI units • Borosilicate glass	-	Unit	50 units					

Procurement Ref. No. : DSS/HS/PROC/B50K/2026-2027/036(PCY)
Section/Laboratory Ref. No.: PCY/PR/2026-2027/07

USER REQUIREMENTS						VENDOR'S OFFER				
NO.	ITEM NAME	ITEM DESCRIPTIONS AND SPECIFICATIONS	PART/CAT NO.	PACKAGING SIZE	QUANTITY	ITEM OFFERED	PACKAGING SIZE	QUANTITY	PRICE PER UNIT (\$)	TOTAL (\$)
3	100 ml Clear Volumetric Flask with PE Stopper	- Class A volumetric flask compliant with EN ISO 1042 / DIN 12664 - Individual calibration certificate issued by an ISO/IEC 17025-accredited laboratory - Stated measurement uncertainty - Demonstrable traceability to SI units - Clear borosilicate glass	-	Unit	10 units					
4	250 ml Amber Volumetric Flask with PE Stopper	- Class A volumetric flask compliant with EN ISO 1042 / DIN 12664 - Individual calibration certificate issued by an ISO/IEC 17025-accredited laboratory - Stated measurement uncertainty - Demonstrable traceability to SI units - Borosilicate glass	-	Unit	6 units					

NOTE	VENDOR'S OFFER
Delivery Period Upon the Issuance of PO	
Price Validity must not be less than 90 days	

Bahagian/Unit/ Makmal :	Makmal Farmasi		
Bil. Rujukan Bahagian/Unit/ Makmal :	PCY/PR/2026-2027/07		
Pegawai dirujuk :	Nama : Adeebah Batrisya binti Adrus		
	E-mail : batrisya.adrus@moh.gov.bn		
	Tel.No. : 2393301	Fax No. : -	
UNTUK KEGUNAAN BAHAGIAN PENTADBIRAN SAHAJA			
No. Rujukan PROCUREMENT :	DSS/HS/PROC/B50K/2026-2027/036 (PCY)		
Tarikh Menghantar Iklan :			