

LAMPIRAN - A1

RUJ: BME/A2K/(1737)/2026/HTD (B8)

19/8/2026

TO REPLACE AND INSTALL SPARE PARTS ONE (1) UNIT OF C-MAC INTUBATION FIBERSCOPE FOR CMOS ENDOSCOPES, BRAND: KARL STORZ, MODEL: 8403ZX, S/N: NN3293, BME29652 AT OPERATING THEATRE RIPAS HOSPITAL.

SCOPE OF WORK

To replace and install spare parts one (1) unit of C-Mac intubation Fibrescope for C-Mos endoscopes, Brand: Karl Storz, Model: 8403ZX at OT Ripas Hospital. Detail are as follow:

1. BME29652, Brand: Karl Storz, Model: 8403ZX, SN: NN2393, at Operating Theatre, W/order 2026-12844

- Scope of work to follow equipment manual/recommendation specific for this system which include the following where if necessary:
 - (i) installation of items listed in spareparts list (Lampiran A2)
 - (ii) alignment
 - (iii) calibration
 - (iv) optimisation
 - (vi) any other related maintenance works required after installation of items listed in spareparts list (Lampiran A2)
- Perform functional test, verification and performance check post-maintenance.
- Disposal of any used spare parts, unless otherwise instructed.
- The quotation submitted shall remain valid for a period of not less than twelve (12) months from the date of submission.
- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.

LAMPIRAN - A2

RUJ: BME/A2K/(1737)/2026/HTD (B8)

TO REPLACE AND INSTALL SPARE PARTS ONE (1) UNIT OF C-MAC INTUBATION FIBERSCOPE FOR CMOS ENDOSCOPES, BRAND: KARL STORZ, MODEL: 8403ZX, S/N: NN3293, BME29652 AT OPERATING THEATRE RIPAS HOSPITAL.

	SPAREPART	PART NUMBER	QUANTITY
1	C-MAC VIDEO LARYNGOSCOPE MAC #3, CMOS TECHNOLOGY	8404AX	1