

LAMPIRAN - A1

RUJ: BME/A2K/(1743)/2026/HTD (B8)

7/9/2026

TO REPAIR ONE (1) UNIT M4 MICRODEBRIDER HANDPIECE, AT OPERATING THEATER , HOSPITAL RIPAS.

SCOPE OF WORK

To repair one (1) unit of M4 microdebrider handpiece, at Operating Theatre Ripas Hospital.
Detail are as follow:

1. Brand: Medtronic, Model: M4, SN: 2509, at Operating Theatre W/order 2026-13316

- Scope of work to follow equipment manual/recommendation specific for this system which include the following where if necessary:

(i) installation of items listed in spareparts list (Lampiran A2)

(ii) alignment

(iii) calibration

(iv) optimisation

(vi) any other related maintenance works required after installation of items listed in spareparts list (Lampiran A2)

- Perform functional test, verification and performance check post-maintenance.

- Disposal of any used spare parts, unless otherwise instructed.

- The quotation submitted shall remain valid for a period of not less than twelve (12) months from the date of submission.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.